

NON PARTICIPANT COPY

LABORATORY NAME:

LABORATORY NUMBER:

Amino Acid Racemization Proficiency Study

Report IV: OSTRICH EGG SHELL (B)

June 2012

Acknowledgements.

Firstly, thanks go to all the laboratories who agreed to take part in this study. Also to Matthew Collins, Kirsty Penkman, James Cussens at the University of York, UK and to Norman MacLeod at the Natural History Museum, London, UK for their support, to Richard Allen and Bea Demarchi for analytical technical assistance and to Ken Mathieson, FAPAS, Ferra, Sand Hutton, York for initial spreadsheet design ideas. This work was carried out at the NERC recognised North East Amino Acid Racemization Laboratory at the University of York and was funded by the Arts and Humanities Research Council (AHRC), UK with assistance from NHM in London.



Contents

ACKNOWLEDGEMENTS.	1
1 INTRODUCTION	10
1.1 Amino Acid Racemisation	10
1.2 Proficiency Testing	11
1.2.1 <i>Organisation</i>	12
2 TEST MATERIALS	13
2.1 Preparation	13
2.2 Homogeneity	13
2.3 Distribution	13
2.4 Result Submission	14
3 HOMOGENEITY	15
3.1 General Procedure	15
3.1.1 <i>Statistical analysis.</i>	15
3.2 Evaluation of Ostrich Egg Shell (B) Test Material Homogeneity Data	16
4 STATISTICAL EVALUATION; SUMMARY STATISTICS	26
4.1 Precision Analysis	26
4.2 Summary Statistics	28
4.2.1 <i>Experimental Standard Uncertainty of the Mean $u(x)$</i>	28
4.2.2 <i>Setting the correct coverage factor for Expanded Uncertainty determination.</i>	29
4.3 t-Distribution vs Normal Distribution	30
5 STATISTICAL EVALUATION; ACCURACY & PERFORMANCE ANALYSIS	83
5.1 Background to understanding Performance Evaluation	83
5.1.1 <i>z-Scores</i>	83
5.1.2 <i>The Target Standard Deviation; σ_p</i>	84
5.2 In the absence of Fitness-for-Purpose Criteria	84
5.2.1 <i>Relative percentage bias</i>	85
5.3 The Assigned Value, X	86
5.3.1 <i>The uncertainty of the Assigned value $u(X)$.</i>	86
5.4 Derivation of X for Amino Acids in Ostrich Egg Shell (B) Test Material	86

5.5	Interpreting Results - a word of caution.	87
6	MEASUREMENT UNCERTAINTY	113
6.1	Estimation of Measurement Uncertainty from Inter-laboratory comparisons.	113
6.2	Standard uncertainty due to Bias ($ubias$).	114
6.2.1	<i>For a result from a single proficiency test.</i>	114
6.2.2	<i>For results from multiple proficiency tests</i>	115
6.3	Combined uncertainty (uC).	115
6.4	Expanded Uncertainty (U).	116
6.5	Calculating Measurement Uncertainty for Amino Acids in Ostrich Egg Shell (B) Test Material	116
6.5.1	<i>Measurement Uncertainty Evaluation for a series of results using RMS_{bias}.</i>	117
6.5.2	<i>Measurement Uncertainty Evaluation for a single result.</i>	117
Appendix 1: Analytical Methods Used by Participants		145
	<i>Reverse Phase HPLC/ HPLC-Ion Exchange</i>	145
	<i>Gas Chromatography</i>	149
	<i>Internal Quality Control</i>	153
Appendix 2: Glossary of Abbreviations, Symbols, Terms & Definitions		155
	<i>Abbreviations</i>	155
	<i>Symbols</i>	155
	<i>Terms and Definitions</i>	156
Appendix 3: Tables of Critical Values		165
	<i>Student t-distribution</i>	165
	<i>Factors F_1 and F_2 (95% significance level)</i>	165
	<i>Cochran's Critical values (95% significance level)</i>	166
Appendix 4: References		167
	<i>Contact Details;</i>	170

Tables

Table 3.1: Homogeneity D/L Values for Ostrich Egg Shell (B) Test Material	18
Table 3.1: Precision Estimates derived from Participants' submitted results	26
Table 3.2: Summary Statistics for L and D Aspartic Acid / Asparagine Peak Area Data	31
Table 3.3: Summary Statistics for L and D Aspartic Acid / Asparagine Concentration Data (pM)	32
Table 3.4: Summary Statistics for L and D Aspartic Acid / Asparagine D/L Ratio Value	33
Table 3.5: Summary Statistics for L and D Glutamic Acid / Glutamine Peak Area Data	36
Table 3.6: Summary Statistics for L and D Glutamic Acid / Glutamine Concentration Data (pM)	37

Table 3.7: Summary Statistics for L and D Glutamic Acid / Glutamine D/L Ratio Value	38
Table 3.8: Summary Statistics for L and D Serine Peak Area Data	41
Table 3.9: Summary Statistics for L and D Serine Concentration Data (pM)	42
Table 3.10: Summary Statistics for L and D Serine D/L Ratio Value	43
Table 3.11: Summary Statistics for L and D Arginine Peak Area Data	46
Table 3.12: Summary Statistics for L and D Arginine Concentration Data (pM)	47
Table 3.13: Summary Statistics for L and D Arginine D/L Ratio Value	48
Table 3.14: Summary Statistics for L and D Alanine Peak Area Data	51
Table 3.15: Summary Statistics for L and D Alanine Concentration Data (pM)	52
Table 3.16: Summary Statistics for L and D Alanine D/L Ratio Value	53
Table 3.17: Summary Statistics for L and D Valine Peak Area / Height Data	56
Table 3.18: Summary Statistics for L and D Valine Concentration Data (pM)	57
Table 3.19: Summary Statistics for L and D Valine D/L Ratio Value	58
Table 3.20: Summary Statistics for L and D Phenylalanine Peak Area Data	61
Table 3.21: Summary Statistics for L and D Phenylalanine Concentration Data (pM)	62
Table 3.22: Summary Statistics for L and D Phenylalanine D/L Ratio Value	63
Table 3.23: Summary Statistics for D-Alloisoleucine/L-Isoleucine Peak Area Data	66
Table 3.24: Summary Statistics for D-Alloisoleucine/L-Isoleucine Concentration Data (pM)	67
Table 3.25: Summary Statistics for D-Alloisoleucine/L-Isoleucine D/L Ratio Value	68
Table 3.26: Summary Statistics for L and D Leucine Peak Area Data	71
Table 3.27: Summary Statistics for L and D Leucine Concentration Data (pM)	72
Table 3.28: Summary Statistics for L and D Leucine D/L Ratio Value	73
Table 3.29: Summary Statistics for L and D Tyrosine Peak Area Data	76
Table 3.30: Summary Statistics for L and D Tyrosine Concentration Data (pM)	77
Table 3.31: Summary Statistics for L and D Tyrosine D/L Ratio Value	78
Table 3.32: Summary Statistics for L and D Methionine Peak Area Data	81
Table 3.33: Summary Statistics for HPLC Internal Standards ; Peak Area/Height Data	82
Table 5.1: Results and Relative Percentage Bias for Total Hydrolysed Amino Acids in OES (B) Test Material	88
Table 5.2: Assigned Values, Standard Deviations and Standard Uncertainties	93
Table 5.3: Satisfactory Performance(Percentage within 95% Confidence Interval)	94
Table 6.1: Estimation of Relative Standard Uncertainty, Combined and Expanded Uncertainty for Amino Acids (using $RMS_{bias}\%$ to access bias contributions) across ALL Laboratories.	118
Table 6.2: Estimation of Relative Standard Uncertainty, Combined and Expanded Uncertainty Estimations for Individual Laboratories	119

Figures

Figure 3.1: Homogeneity Amino Acid D/L Values in Analytical Sequence Order.	20
Figure 3.2: Homogeneity Amino Acid D/L Values; Paired Sub-samples showing Outliers.	23
Figure 3.1: Relationship between the t-distribution and the Normal distribution at a 95% Confidence Level, for low values of n (degrees of freedom ($n-1$) between 1-35).	30
Figure 3.2: Distribution of D/L Values submitted for Aspartic Acid / Asparagine	34
Figure 3.3: Experimental Expanded Uncertainty ($k=2$) of the Mean D/L value for Aspartic Acid / Asparagine (value of n displayed).	35
Figure 3.4: Experimental Expanded Uncertainty ($k=t_{(0.05,df)}$) of the Mean D/L value for Aspartic Acid / Asparagine (value of n displayed).	35
Figure 3.5: Distribution of D/L Values submitted for Glutamic Acid / Glutamine	39
Figure 3.6: Experimental Expanded Uncertainty ($k=2$) of the Mean D/L value for Glutamic Acid / Glutamine (value of n displayed).	40
Figure 3.7: Experimental Expanded Uncertainty ($k=t_{(0.05,df)}$) of the Mean D/L value for Glutamic Acid / Glutamine (value of n displayed).	40
Figure 3.8: Distribution of D/L Values submitted for Serine	44
Figure 3.9: Experimental Expanded Uncertainty ($k=2$) of the Mean D/L value for Serine (value of n displayed).	45
Figure 3.10: Experimental Expanded Uncertainty ($k=t_{(0.05,df)}$) of the Mean D/L value for Serine (value of n displayed).	45
Figure 3.11: Distribution of D/L Values submitted for Arginine	49
Figure 3.12: Experimental Expanded Uncertainty ($k=2$) of the Mean D/L value for Arginine (value of n displayed).	50
Figure 3.13: Experimental Expanded Uncertainty ($k=t_{(0.05,df)}$) of the Mean D/L value for Arginine (value of n displayed).	50
Figure 3.14: Distribution of D/L Values submitted for Alanine	54
Figure 3.15: Experimental Expanded Uncertainty ($k=2$) of the Mean D/L value for Alanine (value of n displayed).	55
Figure 3.16: Experimental Expanded Uncertainty ($k=t_{(0.05,df)}$) of the Mean D/L value for Alanine (value of n displayed).	55
Figure 3.17: Distribution of D/L Values submitted for Valine	59
Figure 3.18: Experimental Expanded Uncertainty ($k=2$) of the Mean D/L value for Valine (value of n displayed).	60
Figure 3.19: Experimental Expanded Uncertainty ($k=t_{(0.05,df)}$) of the Mean D/L value for Valine (value of n displayed).	60
Figure 3.20: Distribution of D/L Values submitted for Phenylalanine	64
Figure 3.21: Experimental Expanded Uncertainty ($k=2$) of the Mean D/L value for Phenylalanine (value of n displayed).	65
Figure 3.22: Experimental Expanded Uncertainty ($k=t_{(0.05,df)}$) of the Mean D/L value for Phenylalanine (value of n displayed).	65

Figure 3.23: Distribution of D/L Values submitted for D-Alloisoleucine/L-Isoleucine	69
Figure 3.24: Experimental Expanded Uncertainty (k=2) of the Mean D/L value for D-Alloisoleucine/L-Isoleucine (value of n displayed).	70
Figure 3.25: Experimental Expanded Uncertainty (k=t_(0.05,df)) of the Mean D/L value for D-Alloisoleucine/L-Isoleucine (value of n displayed).	70
Figure 3.26: Distribution of D/L Values submitted for Leucine	74
Figure 3.27: Experimental Expanded Uncertainty (k=2) of the Mean D/L value for Leucine (value of n displayed).	75
Figure 3.28: Experimental Expanded Uncertainty (k=t_(0.05,df)) of the Mean D/L value for Leucine (value of n displayed).	75
Figure 3.29: Distribution of D/L Values submitted for Tyrosine	79
Figure 3.30: Experimental Expanded Uncertainty (k=2) of the Mean D/L value for Tyrosine (value of n displayed).	80
Figure 3.31: Experimental Expanded Uncertainty (k=t_(0.05,df)) of the Mean D/L value for Tyrosine (value of n displayed).	80
Figure 5.1: Distribution of Participants' Average Measurement Values	95
Figure 4.2: Relative Percentage Bias for Aspartic Acid / Asparagine D/L Results (all data) in Ostrich Egg Shell (B) Test Material	96
Figure 4.3: Relative Percentage Bias for Aspartic Acid / Asparagine D/L Results (rpHPLC data only) in Ostrich Egg Shell (B) Test Material	97
Figure 4.4: Relative Percentage Bias for Glutamic Acid / Glutamate D/L Results (all data) in Ostrich Egg Shell (B) Test Material	98
Figure 4.5: Relative Percentage Bias for Glutamic Acid / Glutamate D/L Results (rpHPLC data only) in Ostrich Egg Shell (B) Test Material	99
Figure 4.6: Relative Percentage Bias for Serine D/L Results (all / rpHPLC data) in Ostrich Egg Shell (B) Test Material	100
Figure 4.7: Relative Percentage Bias for Arginine D/L Results (rpHPLC data only) in Ostrich Egg Shell (B) Test Material	101
Figure 4.8: Relative Percentage Bias for Alanine D/L Results (all data) in Ostrich Egg Shell (B) Test Material	102
Figure 4.9: Relative Percentage Bias for Alanine D/L Results (rpHPLC data only) in Ostrich Egg Shell (B) Test Material	103
Figure 4.10: Relative Percentage Bias for Valine D/L Results (all data) in Ostrich Egg Shell (B) Test Material	104
Figure 4.11: Relative Percentage Bias for Valine D/L Results (rpHPLC data only) in Ostrich Egg Shell (B) Test Material	105
Figure 4.12: Relative Percentage Bias for Phenylalanine D/L Results (all data) in Ostrich Egg Shell (B) Test Material	106
Figure 4.13: Relative Percentage Bias for Phenylalanine D/L Results (rpHPLC data only) in Ostrich Egg Shell (B) Test Material	107
Figure 4.14: Relative Percentage Bias for D-Alloisoleucine/L-Isoleucine Results (all data) in Ostrich Egg Shell (B) Test Material	108

Figure 4.15: Relative Percentage Bias for D-Alloisoleucine/L-Isoleucine Results (rpHPLC data only) in Ostrich Egg Shell (B) Test Material	109
Figure 4.16: Relative Percentage Bias for Leucine D/L Results (all data) in Ostrich Egg Shell (B) Test Material	110
Figure 4.17: Relative Percentage Bias for Leucine D/L Results (rpHPLC data only) in Ostrich Egg Shell (B) Test Material	111
Figure 4.18: Relative Percentage Bias for Tyrosine D/L Results (rpHPLC data only) in Ostrich Egg Shell (B) Test Material	112
Figure 6.1: Bias and Precision Components to Measurement Uncertainty Estimation.	114
Figure 6.2: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Aspartic acid / Asparagine D/L Values in Ostrich Egg Shell (B) Test Material	128
Figure 6.3: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Aspartic acid / Asparagine D/L Values in Ostrich Egg Shell (B) Test Material	128
Figure 6.4: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Aspartic acid / Asparagine rpHPLC D/L Values in Ostrich Egg Shell (B) Test Material	129
Figure 6.5: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Aspartic acid / Asparagine rpHPLC D/L Values in Ostrich Egg Shell (B) Test Material	129
Figure 6.6: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Glutamic acid / Glutamine D/L Values in Ostrich Egg Shell (B) Test Material	130
Figure 6.7: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Glutamic acid / Glutamine D/L Values in Ostrich Egg Shell (B) Test Material	130
Figure 6.8: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Glutamic acid /Glutamine rpHPLC D/L Values in Ostrich Egg Shell (B) Test Material	131
Figure 6.9: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Glutamic acid / Glutamine rpHPLC D/L Values in Ostrich Egg Shell (B) Test Material	131
Figure 6.10: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Serine D/L Values in Ostrich Egg Shell (B) Test Material	132
Figure 6.11: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Serine D/L Values in Ostrich Egg Shell (B) Test Material	132
Figure 6.12: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Arginine D/L Values in Ostrich Egg Shell (B) Test Material	133
Figure 6.13: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Arginine D/L Values in Ostrich Egg Shell (B) Test Material	133
Figure 6.14: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Alanine D/L Values in Ostrich Egg Shell (B) Test Material	134

Figure 6.15: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Alanine D/L Values in Ostrich Egg Shell (B) Test Material	134
Figure 6.16: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Alanine (rpHPLC) D/L Values in Ostrich Egg Shell (B) Test Material	135
Figure 6.17: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Alanine (rpHPLC) D/L Values in Ostrich Egg Shell (B) Test Material	135
Figure 6.18: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Valine D/L Values in Ostrich Egg Shell (B) Test Material	136
Figure 6.19: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Valine D/L Values in Ostrich Egg Shell (B) Test Material	136
Figure 6.20: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Valine (rpHPLC) D/L Values in Ostrich Egg Shell (B) Test Material	137
Figure 6.21: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Valine (rpHPLC) D/L Values in Ostrich Egg Shell (B) Test Material	137
Figure 6.22: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Phenylalanine D/L Values in Ostrich Egg Shell (B) Test Material	138
Figure 6.23: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Phenylalanine D/L Values in Ostrich Egg Shell (B) Test Material	138
Figure 6.24: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Phenylalanine (rpHPLC) D/L Values in Ostrich Egg Shell (B) Test Material	139
Figure 6.25: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Phenylalanine (rpHPLC) D/L Values in Ostrich Egg Shell (B) Test Material	139
Figure 6.26: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for D-Alloisoleucine/L-Isoleucine Values in Ostrich Egg Shell (B) Test Material	140
Figure 6.27: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on D-Alloisoleucine/L-Isoleucine Values in Ostrich Egg Shell (B) Test Material	140
Figure 6.28: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for D-Alloisoleucine/ L-Isoleucine rpHPLC Values in Ostrich Egg Shell (B) Test Material	141
Figure 6.29: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on D-Alloisoleucine/ L-Isoleucine rpHPLC Values in Ostrich Egg Shell (B) Test Material	141
Figure 6.30: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Leucine D/L Values in Ostrich Egg Shell (B) Test Material	142
Figure 6.31: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Leucine D/L Values in Ostrich Egg Shell (B) Test Material	142

Figure 6.32: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Leucine rpHPLC D/L Values in Ostrich Egg Shell (B) Test Material	143
Figure 6.33: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Leucine rpHPLC D/L Values in Ostrich Egg Shell (B) Test Material	143
Figure 6.34: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Tyrosine D/L Values in Ostrich Egg Shell (B) Test Material	144
Figure 6.35: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Tyrosine D/L Values in Ostrich Egg Shell (B) Test Material	144

1 INTRODUCTION

1.1 Amino Acid Racemisation

Amino Acid racemization (or epimerizationⁱ for molecules with two carbon centres) is a diagenetic process that occurs naturally following protein synthesis. The process involves the slow inter-conversion between the two chiral forms of amino acids; the building blocks of proteins, from the Laevo (L-form) in life to the Dextro (D-form). Conversion of the L to D form continues until equilibrium is reached, for most amino acids this is usually equal to 1. This process can take many thousands of years, thus the D/L ratio value can be used as an indicator of time. This technique has been particularly successful in dating quaternary sediments using protein decomposition in fossil biominerals such as shell. The unique mineral crystalline structure of shells trap original proteins, with minimal loss and free from contamination.

The rates of racemization for the 20 or so different amino acids vary, are highly temperature dependent, matrix and species specific. Because the thermal history of a site is rarely known, it becomes difficult to determine precise age estimates. For this reason, most research tends to apply the technique as a relative stratigraphic tool within a defined locality using independently calibrated material; the assumption being that if all sites share the same temperature history, any observed D/L differences can be interpreted as relative age differences. Similarly, it becomes possible to use D/L values as indicators of relative temperature differences between same age sites, if independently dated using other appropriate techniques.

The last 30 years has seen significant changes in the analysis of amino acid racemization. Early research based on ion-exchange liquid chromatography (IE-LC) focused on the ratio between the D and L form of isoleucine but as methods developed, it became possible to detect and measure increasing numbers of amino acids, from six or seven using gas chromatography (GC) to ten or more routinely determined today using reverse-phase HPLC (rp-HPLC). These advances have continued to improve the precision in routine analysis and its acceptability as a valid dating method within the geochronology community. AAR now requires mg sample sizes, is relatively fast and with inexpensive preparation and analytical costs, is a useful screening method with the potential to provide age estimates that go far beyond current radiocarbon timescales, covering the entire quaternary period.

Nonetheless, AAR data is still often viewed dismissively. Important unaccounted differences between AAR age estimates and other dating methods have been previously reported (Wehmiller, 1992) with wide precision estimates for numerical ages up to 40-50% where the age equation was not calibrated locally, improving to 15% when it is (McCoy, 1987). More recently a value of 30% representing 53-142 years in Holocene shells has been reported following the removal of outliers (Kosnik et al., 2008).

ⁱ Note; The more general term 'racemization' will be used throughout this report to refer to both racemization and epimerization.

Clearly, the accuracy of numerical age estimates relies heavily on the accuracy of analytical data. Wehmiller and Miller (2000) in their review of aminostratigraphic dating methods, report intra-laboratory precision estimates for repeated instrumental determinations of the same hydrolysate of 2%, for multiple analyses of different fragments of the same material, between 3-5%, whilst for multiple samples from the same sample location, between 5-10%. Previous inter-laboratory studies have focused on comparing individual laboratory precision estimates derived from replicate instrumental measurements (Wehmiller, 1984). These studies have demonstrated the variability in precision between different amino acids and methods. Whilst most laboratories report CV% values between 2-5%, there are often significant differences between laboratories that would result in substantial numerical age differences of 25% or greater, and call for the need for a common working standard with D/L reference values.

In spite of these efforts, there remains inconsistency in the use and expression of precision estimates, ambiguity in the reporting of uncertainty, and an absence of any assessment of method or laboratory bias, not least due to the absence of a suitable reference material. It is with regard to these issues that the current study has been undertaken and attempts to address.

Many laboratories continue to report uncertainty estimates as the CV of replicate instrumental measurements. Although analytical precision (i.e.; instrumental repeatability) is an important component of the overall uncertainty budget, it is usually amongst one of the smallest contributions and is often negligible compared to method and laboratory precision estimates. However, determination of method/laboratory precision through method validation or inter-laboratory collaborative trial, are outside the scope of this report.

Experience within other industry sectors has demonstrated, through regular participation in proficiency tests, that analytical performance improves over time. It is now nearly thirty years since the last inter-laboratory study was carried out using powdered fossil material (Wehmiller, 1984), and it is timely to coordinate a new inter-laboratory study in support of current methodologies.

1.2 Proficiency Testing

It has long been widely appreciated that participation in inter-laboratory studies is a valuable tool enabling method comparisons and development. Proficiency testing (PT) is a specific type of inter-laboratory evaluation providing an objective and formalized evaluation of accuracy against a consensus value enabling an objective comparison with other laboratories' data and is an important indicator of bias. Accuracy and by inference, performance, is characterized by elements of both precision and trueness. A laboratory may be inaccurate due to systematic bias effects, random error influencing poor repeatability, or both. In the absence of Certified Reference Materials (CRMs) for bias determination, participation in a proficiency test can provide a valuable alternative for laboratories.

Proficiency testing is commonly encountered in sectors that rely heavily on regulation and compliance such as medicine and public health, forensic science, chemical and geochemical analytical services, manufacturing industries, calibration and engineering, food and feed industries. Today more than 1,300 PT schemes worldwide are listed on the EPTISⁱⁱ website. Participation in such a scheme is also a requirement of analytical laboratories seeking accreditation to ISO 17025 (2005).

The regular analysis of an independent quality control material forms a valuable part of external quality control (EQC) enabling comparability on a much wider scale with other laboratories, analysts

ⁱⁱ European Proficiency Testing Information Service; http://www.eptis.bam.de/en/about/what_is_eptis/index.htm

and methods. As such, it is an essential element of any laboratory's Quality Assurance (QA) programme, together with the use of validated methods and internal quality control (IQC) procedures.

Whilst performance in individual rounds can identify unexpected error influences needing investigation, long term trends are probably of greater value and can be observed using control charts (Thompson et al., 2006). The spread of results from a laboratory over a period of time should be compatible with that laboratory's own evaluation of uncertainty. The standard deviation of the differences between the laboratory values and the assigned values providing a means of evaluating the standard uncertainty (Eurachem 2000), see Section 6.2.2.

Test materials left over after the end of a proficiency test can also act as suitable matrix specific reference materials in the absence of CRMs. Because the value of the analyte has been determined by a consensus, it has minimal bias associated with it and a known uncertainty.

1.2.1 *Organisation*

This report is organized in to a number of sections. The next section, Section 2, details how test materials were prepared and distributed, and Section 3 presents the homogeneity data and discusses some of the issues encountered with the assessment of homogeneity for this test material. A summary evaluation of submitted results is presented in Section 4. Values for peak area and peak height together with concentrations and D/L values are tabulated with individual laboratory standard deviations, percentage relative standard deviations (RSD%) otherwise referred to as the coefficient of variation (CV%), instrumental replicate standard uncertainty estimates (u) representing precision from repeated measurements, (i.e.; instrumental repeatability) and the percentage relative standard uncertainty (RSU%). Section 5 assesses the accuracy of the results compared to the assigned value and calculates the relative percentage bias as an indication of performance. The last section, Section 6 then turns to the subject of measurement uncertainty and discusses the requirement for bias estimation in addition to precision estimates for uncertainty determination. The section demonstrates how proficiency test data can be used to derive indicative standard uncertainty contributions and values for combined and expanded uncertainty estimates. Finally method details as provided by the participants have been collated and together with the glossary of terms and symbols used in this report, relevant statistical tables and references, make up the Appendices at the end of the report.

2 TEST MATERIALS

Ostrich Egg Shell (B)

2.1 Preparation

The calcitic ostrich egg shell test material was prepared from a blown modern ostrich egg supplied by Oslinc Ostrich Farm, Boston in Lincolnshire, UK, in 2010. A section of the egg shell was broken into pieces and approximately 50 g was cleaned with repeated washing in ultrapure water using a sonicator. Rehydrated shell membrane lining was removed by peeling and scraping and further washed until the water remained clear. The cleaned ostrich egg shell was then lightly covered and left to air dry for 48 hours. The broken shell pieces were placed on a flat heat-proof dish and heated in the oven for 8 hours at 140 °C. After cooling, pieces of the heated shell were lightly milled using short bursts of an electric coffee mill to avoid heating of the motor and blade. The reduced fragments and coarse powder were further ground using a sterile pestle and mortar and sieved, to ≤ 250 µm before finally being tumble-blended overnight on a roller mixer.

Half the cleaned, heated, powdered ostrich egg shell was measured and individual 20 mg sub-samples were weighed into sterile glass vials and labeled as Ostrich Egg Shell (B) (OES (B)). The remaining half of the powdered material was bleached for 48 hours using 50µl of 12% NaOCl per mg of powder. After washing and drying this material was also weighed (20 mg sub-samples) into sterile glass vials and labelled as Ostrich Egg Shell (A) (OES (A)). Both sets of test material were stored at room temperature prior to distribution.

2.2 Homogeneity

Ten randomly selected test materials were sub-sampled to give 10 duplicate samples (10 x a and b), which were then analysed for total hydrolysable amino acids (THAA) using reverse phase HPLC (rpHPLC) according to the standard method (Kaufman and Manley W.F., 1998). The results, together with their statistical evaluation, are given in Section 3.

2.3 Distribution

Participants were previously asked to notify the organizer with details of their proposed analytical method and were sent the appropriate number of individual test materials necessary to give sufficient bulk material required by the different methods. Those using rpHPLC were sent a single individually numbered 20mg test material, those using ion-exchange HPLC (HPLC-IE) were sent three individual test materials (60mg total) and those using gas chromatography (GC) were sent ten individual test materials (200mg total). Participants receiving multiple test materials were asked to pool the contents to get the required quantity rather than simply having a larger sample sent because of the risk of heterogeneity in larger sub-samples. This way, a defined minimum measure of homogeneity could be assured between individual sub-samples of a specified weight, which would not be lost when pooled.

Test materials were dispatched to eight laboratories located around the world on 15 July 2010.

Due to the small number of participants in the study, additional sets of test materials were provided to those laboratories who had more than one instrument, those using more than one method and those who had more than one member of staff available to carry out the analysis. As a result this increased the possible number of sets of results up to twenty three.

2.4 Result Submission

Participants were asked to submit results and method information on electronic documents sent following dispatch and no later than October 2010. The final set of results was submitted mid-December but three participants were unable to return any results on this occasion due to instrumental difficulties or other commitments. A total of fifteen sets of results were submitted.

Whilst the original intention of this study was to determine performance for only D/L amino acid values, a number of laboratories also asked to submit raw chromatogram data. Consequently, a results proforma was prepared enabling the submission of peak area and height data, together with concentrations and D/L values. Participants were asked to indicate their primary means of determination, i.e.; using peak areas, heights or concentrations. Due to the delay in results being submitted and the time required in assessing the data, the additional information has been summarized and tabulated in Section 4 but not evaluated. Where more than one replicate value was submitted, **instrumental repeatability** standard uncertainty estimates have been determined and plotted to demonstrate the effect of the expanded uncertainty at a 95% confidence level (2 std deviations approximately) on the mean value. Where results were submitted as the mean and standard deviation, these values have been used for the calculation of the standard uncertainty directly.

One laboratory provided free amino acid data (FAA) but these have not been assessed or tabulated on this occasion. In this report only data given for the total hydrolysable amino acid fraction (THAA), have been evaluated. Instrumental replicate measurements provided by individual laboratories have been averaged as necessary to give a single value for each amino acid in the test material supplied. These are tabulated in Section 5, together with an evaluation of performance, assessed as the relative percentage bias, which are also presented as histograms at the end of the section.

Each set of results was given a unique laboratory number. The analytical methods used by each participant are summarised in Appendix I.

3 HOMOGENEITY

Ostrich Egg Shell (B) Test Material

3.1 General Procedure

The purpose of carrying out homogeneity testing, is to prove that any variation in composition between individual test materials, characterized by the sampling standard deviation (s_{sam}) is negligible compared to the variation in measurement determinations carried out by participants of the proficiency test. Due to the time and expense of preparing homogeneous test materials and carrying out the analysis, it is reasonable to start with the assumption that test materials are homogeneous and by carrying out homogeneity testing we are looking for evidence of heterogeneity, rather than vice versa. The following procedure for the assessment of homogeneity follows that given in the standard ISO 13528:2005, and the 2006 IUPAC International Harmonized Protocol (Thompson et al).

It is recommended that ten (and no fewer than seven) randomly selected prepared and packaged test materials are selected at random using a random number generator. Each sample is then individually homogenized and two separate portions are removed and labeled 1a and 1b; 2a & 2b;....10a & 10b etc. Each individual sub-sample is then prepared according to the appropriate method and analysed in a random order under repeatability conditions, (i.e.; at the same time or in as short a time as possible, as a single batch on the same day by the same analyst on the same instrument etc).

Resulting data should be scrutinized first for obviously anomalous values eg values greater or less than 10 times the average. It is helpful to plot data in run order to identify trends, stability issues or measurement problems. However, assuming no problems are identified the data should be sorted and sub-samples re-paired to undergo the following statistical evaluation.

3.1.1 Statistical analysis.

- a) Data are initially subjected to a Cochran's outlier test.

The Cochran's test statistic is determined by the ratio of the maximum squared difference to the sum of squared differences;

$$C = D_{max}^2 / \sum D_i^2$$

Where; C is the Cochran's statistic,
 D_{max} is the largest difference between duplicates, and
 D_i is the difference between each pair of duplicates.

The C-value is then compared against tabulated critical values based on the required confidence level and the degrees of freedom, $m-1$, where m is the number of duplicate pairs. If $C > C_{crit}$, the pair is identified as a Cochran's outlier and removed from the data set.

b) Evaluation of Analytical Variance

Occasionally, genuine inhomogeneity between samples is missed due to large within-sample analytical variances, i.e.; between the two sub-sample values (eg; 1a & 1b). This can mask significant between-sample differences (eg; 1 - 10). It is therefore recommended to evaluate the analytical precision first to ensure that the method is sufficiently precise to detect inhomogeneity.

Data are assessed using a one-way ANOVA to estimate the analytical variance.

The analytical variance $s_{an}^2 = MS_w$ where MS_w = within groups mean square.

Note how s_{an} is analogous to the repeatability standard deviation, s_r in Section 4.1

Satisfactory analytical precision is assumed if the analytical deviation is less than half the target value for standard deviation (σ_p) for the proficiency test (Fearn and Thompson, 2001);

i.e.; $s_{an}/\sigma_p < 0.5$

Note; due to the absence of an external target value for standard deviation (σ_p), a target value for homogeneity (σ_h) has been determined such that $s_{an}/0.5 = \sigma_h$

c) Evaluation of Sampling Variance.

The sampling variance $s_{sam}^2 = \frac{MS_b - MS_w}{2}$ where MS_b = between groups mean square.

Or as $s_{sam} = 0$, if the above estimate is negative (Fearn & Thompson, 2001)

Note how s_{sam} is analogous to the between-sample standard deviation, s_L in Section 4.1.

Calculate the permissible sampling variance $s_{all}^2 = (0.3 \times \sigma_p)^2$

Calculate the critical value (c) for the test using tabulated values for F_1 and F_2 (ISO 13528:2005, Thompson et al; 2006, Fearn and Thompson; 2001).

$$c = F_1 s_{all}^2 + F_2 s_{an}^2$$

If $s_{sam}^2 < c$, the sampling variance has not exceeded the allowable fraction of the target standard deviation. There is no evidence of inhomogeneity and the test has been passed.

3.2 Evaluation of Ostrich Egg Shell (B) Test Material Homogeneity Data

Ten test materials were selected at random from the bulk of previously prepared individual test materials. Each test material was divided into two sub-samples and prepared according to the standard procedure prior to hydrolysis for total hydrolysed amino acids. The twenty individual sub-samples were then randomized and analysed as a single batch under repeatability conditions using reverse-phase HPLC.

The D/L results for all twenty sub-samples for each amino acid were plotted in run order to identify trends or problems with the data and are shown in Figure 3.1. There were no problems encountered with the analysis and no Cochran's outliers detected subsequently.

The D/L results and statistical evaluation are given in Table 3.1. Figure 3.2 shows the paired D/L values for each amino acid

In all cases, σ_h , the target standard deviation (for sufficient homogeneity), was set as the minimum value necessary to ensure fitness-for-purpose, i.e.; that σ_h was at least twice the analytical precision (repeatability) and that the allowable sampling variance was sufficient to accommodate the observed between-sample differences.

Table 3.1: Homogeneity D/L Values for Ostrich Egg Shell (B) Test Material

sample id	analyte									
	Asx D/L		Glx D/L		Ser D/L		Arg D/L		Ala D/L	
	replicate 1	replicate 2								
1	0.234	0.232	0.070	0.070	0.119	0.115	0.117	0.117	0.075	0.077
2	0.229	0.230	0.070	0.070	0.115	0.113	0.119	0.119	0.076	0.077
3	0.242	0.238	0.071	0.071	0.123	0.122	0.120	0.114	0.074	0.079
4	0.240	0.235	0.071	0.071	0.123	0.119	0.123	0.119	0.077	0.075
5	0.228	0.233	0.070	0.070	0.114	0.115	0.116	0.116	0.075	0.075
6	0.239	0.238	0.071	0.071	0.123	0.125	0.114	0.118	0.077	0.078
7	0.237	0.235	0.070	0.071	0.123	0.119	0.119	0.118	0.082	0.076
8	0.231	0.240	0.070	0.071	0.116	0.123	0.117	0.118	0.075	0.077
9	0.228	0.242	0.070	0.071	0.116	0.124	0.118	0.117	0.074	0.076
10	0.230	0.234	0.069	0.069	0.117	0.120	0.117	0.113	0.078	0.075
mean, N	0.235	20	0.070	20	0.119	20	0.117	20	0.076	20
origin of target sd (σ_h)	perception									
abs. target sd (σ_h) & as RSD%	0.0084	3.6	0.0009	1.3	0.0060	5.0	0.0043	3.7	0.0040	5.3
s_{an}	0.0042		0.0005		0.0030		0.0021		0.0020	
s_{an} / σ_h	0.4944		0.4999		0.4952		0.4879		0.4976	
$s_{an} / \sigma_h < 0.5?$	yes									
s_{sam}^2	3.67E-06		1.79E-07		6.71E-06		4.95E-07		0.00E+00	
σ_{all}^2	6.42E-06		7.54E-08		3.20E-06		1.70E-06		1.47E-06	
critical	2.97E-05		3.53E-07		1.48E-05		7.74E-06		6.85E-06	
$s_{sam}^2 < \text{critical?}$	ACCEPT									

Table 3.1: Homogeneity D/L Values for Ostrich Egg Shell (B) Test Material (continued).

sample id	analyte							
	Val D/L		PheD/L		D-Aile/L-Ile		Leu D/L	
	replicate 1	replicate 2	replicate 1	replicate 2	replicate 1	replicate 2	replicate 1	replicate 2
1	0.024	0.023	0.062	0.063	0.032	0.031	0.057	0.060
2	0.022	0.024	0.062	0.062	0.032	0.032	0.057	0.060
3	0.025	0.023	0.064	0.063	0.034	0.031	0.060	0.060
4	0.027	0.025	0.063	0.063	0.033	0.033	0.062	0.060
5	0.022	0.025	0.062	0.063	0.031	0.033	0.060	0.061
6	0.024	0.026	0.062	0.063	0.032	0.034	0.057	0.063
7	0.026	0.022	0.063	0.063	0.033	0.031	0.056	0.061
8	0.022	0.022	0.062	0.062	0.031	0.035	0.060	0.063
9	0.022	0.025	0.061	0.063	0.031	0.034	0.059	0.057
10	0.028	0.023	0.062	0.062	0.033	0.031	0.061	0.060
mean, N	0.024	20	0.062	20	0.033	20	0.060	20
origin of target sd (σ_h)	perception		perception		perception		perception	
abs. target sd (σ_h) & as RSD%	0.0036	15	0.00137	2.2	0.00313	9.7	0.00447	7.5
s_{an}	0.0018		0.0007		0.0016		0.0022	
s_{an} / σ_h	0.4964		0.4983		0.4999		0.4952	
$s_{an} / \sigma_h < 0.5?$	yes		yes		yes		yes	
s_{sam}^2	0.00E+00		7.37E-10		0.00E+00		0.00E+00	
σ_{all}^2	1.17E-06		1.70E-07		8.82E-07		1.80E-06	
critical	5.44E-06		7.92E-07		4.13E-06		8.35E-06	
$s_{sam}^2 < \text{critical?}$	ACCEPT		ACCEPT		ACCEPT		ACCEPT	

Figure 3.1: Homogeneity Amino Acid D/L Values in Analytical Sequence Order.





Figure 3.1: Homogeneity Amino Acid D/L Values in Analytical Sequence Order (continued).





Figure 3.1: Homogeneity Amino Acid D/L Values in Analytical Sequence Order; (continued)




Figure 3.2: Homogeneity Amino Acid D/L Values; Paired Sub-samples showing Outliers.




Figure 3.2: Homogeneity Amino Acid D/L Values; Paired Sub-samples showing Outliers.





Figure 3.2: Homogeneity Amino Acid D/L Values; Paired Sub-samples showing Outliers.



4 STATISTICAL EVALUATION; *Summary Statistics*

4.1 Precision Analysis

In keeping with the style of previously conducted inter-laboratory comparisons (Wehmiller, 1984, Wehmiller, 2010), participants were invited to submit peak information and concentration data in addition to the D/L value data requested for the proficiency study. Consequently a substantial quantity of information was captured. Due to time constraints it was not possible to evaluate all of this additional data, although a comparison of L and D amino acid concentrations would be enlightening.

Table 4.1 summarises indicative values of repeatability and reproducibility precision estimates for each amino acid derived from all participants' individual D/L values. Estimates were calculated using a one way analysis of variance (ANOVA), allowing for unequal replicate numbers. It should be noted that where **all** data have been used in the evaluation of precision estimates in Table 4.1, this includes GC D/L values derived from both peak area and height data where given, although the laboratory subsequently confirmed that in practice only peak area data would be used for chronology building. Results from the analysis of relative bias presented in Section 5, suggest possible empirical differences between methods. Therefore, all rpHPLC data and HPLC-IE data for D-alloisoleucine/L-isoleucine, have also been evaluated separately. However, because all HPLC-IE data came from the same laboratory, reproducibility (RSD_R) values should more correctly be interpreted as an intra-laboratory reproducibility or intermediate precision estimate. As GC data were submitted as average D/L values, it was not possible to determine comparable GC specific precision estimates.

The repeatability standard deviation s_r (Table 4.1), is a measure of the overall within laboratory precision derived from all participating laboratories. **On this occasion, this represents an inter-laboratory approximation of the instrumental precision only**, due to random error effects. This reflects the variability that a single laboratory might be expected to achieve for replicate measurements of the same sample. Typically, this may be slightly larger than instrumental precision estimates derived from a single laboratory (i.e. the $CV\%$ (or $RSD\%$) given in Tables 4.2 – 4.33) but smaller than method repeatability which includes additional variability arising from the analysis of different samples of the same material by a single laboratory, under repeatability conditions. Often the s_r is more conveniently given as the relative repeatability standard deviation expressed as a percentage, ($RSD_r\%$).

s_L is the overall inter-laboratory between sample standard deviation, and indicates the level of agreement between participants. **s_R is the inter-laboratory reproducibility standard deviation and a measure of the overall precision for any given amino acid** in the specified test material. s_R incorporates both the within and between laboratory variability and is a single measure of the variability or uncertainty of the measurement procedure associated with precision. Such determinations are more commonly used to assess data from method specific collaborative trials (Horwitz, 1995, AOAC, 2000) known as the “top-down” approach to uncertainty estimation (RSC Analytical Methods Committee, 1995). The relative standard deviation of reproducibility ($RSD_R\%$)

obtained from a collaborative trial may then be used for the assessment of proficiency test data as it provides an external value for the target standard deviation, i.e.; it describes how the data is expected to behave under conditions of best practice. However, in the absence of a collaborative trial, precision evaluation of the submitted PT results will help give an **indication** of the agreement between laboratories, albeit being slightly exaggerated due to additional method variation between participants. (Note; in the case of empirical methods, PT data should be assessed against method specific precision estimates).

All submitted results have been included in this evaluation without removal of outliers as would otherwise be the case with collaborative trial data. On this occasion it is the intention to observe the behaviour of all submitted results rather than to define best practice. It should be noted that these values have not been used in the later performance evaluation but are given for information and indicative purposes only. Further details on the calculations of S_R , S_L and S_r can be found in (ISO 5725, 1994, ISO 21748, 2010). Precision estimates are calculated using ANOVA, thus;

$$s_r = \sqrt{\text{within group mean square}}$$

$$s_L = \sqrt{\frac{\text{between group mean square} - \text{within group mean square}}{n}}$$

$$s_R = \sqrt{s_r^2 + s_L^2}$$

Table 4.1: Precision Estimates derived from Participants' submitted results

amino acid	no of sets of results (m)	total no of replicates (N)	mean	S_r	$RSR\%$	S_L	$RSR\%$	S_R	$RSR\%$
Asx D/L-all ^a	15	30	0.216	0.0008	0.39	0.0264	12.21	0.0264	12.21
Asx D/L-rpHPLC	11	26	0.210	0.0008	0.40	0.0172	8.15	0.0172	8.16
Glx D/L-all ^a	13	28	0.057	0.0003	0.47	0.0083	14.71	0.0083	14.72
Glx D/L-rpHPLC	11	26	0.056	0.0003	0.47	0.0087	15.45	0.0087	15.46
Ser D/L-rpHPLC	11	26	0.111	0.0009	0.82	0.0033	2.94	0.0034	3.05
Arg D/L-rpHPLC	9	15	0.101	0.0038	3.75	0.0079	7.83	0.0088	8.68
Ala D/L-all ^a	15	30	0.061	0.0065	10.56	0.0057	9.34	0.0086	14.09
Ala D/L-rpHPLC	11	26	0.063	0.0065	10.21	0.0030	4.77	0.0071	11.27
Val D/L-all ^a	15	31	0.019	0.0009	4.90	0.0027	14.35	0.0028	15.17
Val D/L-rpHPLC	11	27	0.019	0.0009	4.72	0.0019	9.99	0.0021	11.05
Phe D/L-all ^a	12	27	0.053	0.0012	2.35	0.0041	7.78	0.0043	8.13
Phe D/L-rpHPLC	11	26	0.053	0.0012	2.35	0.0042	7.95	0.0044	8.29
D-Aile/L-Ile -all ^b	17	35	0.026	0.0010	3.99	0.0067	25.54	0.0067	25.85
D-Aile/L-Ile -rpHPLC	11	27	0.026	0.0011	4.21	0.0078	29.65	0.0079	29.94
D-Aile/L-Ile -HPLC-IE	2	4	0.024	0.0000	0.00	0.0000	0.00	0.0000	0.00
D-Aile/L-Ile -GC					Not determined				
Leu D/L-all ^a	12	27	0.050	0.0056	11.21	0.0038	7.55	0.0067	13.52
Leu D/L-rpHPLC	8	23	0.050	0.0056	11.16	0.0031	6.26	0.0064	12.80
Tyr D/L-rpHPLC	7	11	0.059	0.0020	3.44	0.0039	6.65	0.0044	7.49

^a = rpHPLC and GC data

^b = rpHPLC, GC and HPLC-IE data

4.2 Summary Statistics

Summary statistics are presented in Tables 4.2-4.33 for rpHPLC peak areas and concentrations, peak-height values for HPLC-IE and D/L values for all participants. Individual laboratory replicate D/L values as submitted, are also shown graphically against the assigned values determined in Section 5, for comparison. It should be noted that GC data was submitted as the mean \bar{x} of n replicates with a stated standard deviation, s , and these have been displayed as the mean value with associated error bars on the charts. Data are presented as submitted on the result proforma for each of the total hydrolysed amino acids, including internal standard data provided by participants. Only one laboratory reported data for the free amino acids and this has not been included in this report. Calculations have been carried out on each laboratory's results to give the instrumental precision estimate as the standard deviation (s) and relative standard deviation, $RSD\%$, also known as the coefficient of variance, $CV\%$, for each amino acid, where;

$$RSD\% \text{ or } CV\% = (s/\bar{x}) \times 100$$

Additionally, the experimental standard deviation (or standard error or standard uncertainty) of the mean ($u(\bar{x})$) and the relative standard uncertainty of the mean ($RSU\%$), have been determined. Each laboratory's expanded uncertainty to 2 std deviations or an approximate 95% confidence level, has been evaluated for each amino acid and data are presented in figures to illustrate the effect of uncertainty on the mean value of submitted replicate data.

4.2.1 Experimental Standard Uncertainty of the Mean $u(\bar{x})$

Depending on information sources, there are various names used to describe ($u(\bar{x})$) as mentioned above. Standard uncertainty is always expressed as a standard deviation, thus either experimental standard deviation or standard uncertainty of the mean would be acceptable. In this report, $u(\bar{x})$ will be referred to as the *experimental standard uncertainty of the mean* and reflects the confidence in the mean of replicate values, i.e.; the larger the value of n , the greater the confidence in the mean \bar{x} as an estimate of the true value μ , and the smaller the uncertainty. **Note;** **The observed standard deviation of replicate instrumental measurements describes the distribution of data and is not the same as the uncertainty estimate for the mean.** (Strictly speaking this should be determined using independent repeated measurements and not replicate measurements of the same sample).

Thus;

Experimental standard uncertainty of the mean is obtained from; $u(\bar{x}) = s/\sqrt{n}$

Which, expressed as a percentage relative to the mean; $RSU\% = \left(u(\bar{x})/\bar{x} \right) \times 100$

It is important to appreciate that $u(\bar{x})$ is the uncertainty associated with the mean of replicate instrumental results only. It **contributes** to the **bias** component of the overall combined uncertainty associated with the measurement system (see Figure 6.1) but is **only one component of the uncertainty that should be reported with the mean of analytical results**. Measurement uncertainty determination is discussed this in more detail in Section 6 later in the report.

As a standard uncertainty, $u(\bar{x})$ represents a confidence level equivalent to 68% or 1 standard deviation. This means that 68 percent of the means of repeated replicate results will fall within these limits either side of the mean determined by $\bar{x} \pm u(\bar{x})$. This gives little confidence as in nearly one out of every three occasions, the mean is likely to fall outside of this range. However, in practice it is often more helpful to consider a confidence interval equivalent to 2 standard deviations or a

95.4% probability level in experimental design (usually rounded to 95% for simplicity). This equates to a 1 in 20 chance of falling outside the range. 3 standard deviations would be equivalent to 99.7% confidence or 1 in 300.

To determine these extended limits of confidence an Expanded Uncertainty (U) is calculate thus;

$$U = u(\bar{x}) \times k \quad \text{where } k \text{ is the coverage factor set according to the required confidence level.}$$

Expanded uncertainty is more usually determined following the combination of all individual standard uncertainty components as demonstrated in Section 6. However, it may also be helpful to observe the effect of uncertainty on individual elements to aid method development or quality improvements.

The coverage factor, k , and its role in determining the Expanded uncertainty is now considered in more detail below.

4.2.2 *Setting the correct coverage factor for Expanded Uncertainty determination.*

Theoretically, if analytical results represented an entire population and the true value μ and standard deviation σ were known, it would be possible to calculate the range of values within which repeated experimental means \bar{x} of n measurements were likely to fall with a certain level of confidence. As discussed above, for most general applications, a 2 standard deviation or approximately 95% confidence level is usually acceptable. Thus in this instance $k = 2$ (actually its 1.96σ) and the relevant confidence interval where (approx) 95% of \bar{x} values would lie would be in the range;

$$\mu - \left[2 \times \frac{\sigma}{\sqrt{n}} \right] \quad \text{to} \quad \mu + \left[2 \times \frac{\sigma}{\sqrt{n}} \right]$$

However, in real terms, the true value of μ and σ cannot be known and the aim of experimental investigations is to get the best estimate of μ from the sample mean, \bar{x} . Where the number of replicate measurements is large, i.e.; $n=30$ or more (Currell and Dowman, 2005) then the distribution of mean values conforms with the expectation of normality. However for decreasing values of n , the characteristic bell shaped curve of the normal distribution flattens and widens reflecting the reduced confidence in the value \bar{x} as the best estimate of μ and our uncertainty estimate increases. To compensate for the use of the sample standard deviation, s , rather than the population standard deviation σ , $k=2$ is replaced by the critical t -value as a correction term. The value of t depends on the value of n and the required level of confidence and can be read from any two-tailed t -table in statistical texts. Thus for $n=5$ (degrees of freedom=4) at 95% confidence level ($\alpha=0.05$), $t=3.18$ compared to the original value of $k=2$, or for a pair of replicates; $n=2$, $df=1$, $t=12.7$ and the expanded uncertainty becomes over six times larger than otherwise predicted if $k=2$! Thus the range in which the true value lies with 95% confidence broadens and becomes;

$$\bar{x} - \left[t_{(2,0.05,df)} \times \frac{\sigma}{\sqrt{n}} \right] \quad \text{to} \quad \bar{x} + \left[t_{(2,0.05,df)} \times \frac{\sigma}{\sqrt{n}} \right]$$

In practice and often for simplicity rather than intent, laboratories can often be found to overlook this t -value correction by quoting expanded uncertainties derived from the more favorable $k=2$.

Relative Expanded uncertainties of the submitted results using both $k=t_{(0.05,df)}$ and the more frequently used $k=2$ have been calculated and values expressed as a percentage. For each amino acid, data are given in tables and presented as two comparative figures. Note that where a single replicate value is reported, no uncertainty estimation can be made.

The differences observed in expanded uncertainties between different amino acids for a single laboratory highlights the ease or difficulty of analysis and instrument repeatability. A comparison of

expanded uncertainties across all laboratories for any individual amino acid also demonstrates the effect of different methods or even using different numbers of replicates for the same method.

Whilst these effects are interesting to observe analytically, the effect of the number of replicates is an important practical consideration. Demands for quality and lower uncertainty estimates must be balanced against the extra cost and time incurred by increasing replicate numbers not to mention material availability and often it is financial and resource constraints that become deciding factors.

4.3 t-Distribution vs Normal Distribution

The relationship between the t-distribution and the Normal or Gaussian distribution at 2 standard deviations (95% confidence) is shown below in Figure 4.1. It illustrates the t-distribution deviation (red line) away from normal (black line) for low sample numbers, (degrees of freedom ($n-1$) between 1 - 35 where n is the sample size). The t-value given on the y-axis is used as the correction term in the calculation of expanded uncertainty. t-values are given in Appendix 3.

It can be clearly seen that for a pair of replicate values; ($df = 1$), there is a significant deviation from normal, introducing a correction factor more than 10x larger (t-value = 12.7) on the standard uncertainty estimate. Increasing the number of replicate values to $n = 3$ ($df = 2$), reduces the t-value correction to 4.3, and for $n = 4$ ($df = 3$), the t-value correction becomes 3.2. Thus the effect of increasing the number of replicate values from 2 to 3 will make a substantial reduction in the expanded uncertainty estimate, whilst increasing the number of replicates from 3 to 4 will still make an improvement, but the difference will not be quite as significant. The level of benefit gained by increasing the numbers of replicates gradually diminishes until normality is achieved at about $n = 25$.

The contribution of a particular standard uncertainty estimate to the overall uncertainty budget, should also be borne in mind. For example; the contribution of instrumental analytical precision is likely to be much smaller than the contribution from method precision between different samples. It therefore makes more sense to put time into increasing the number of individual samples tested than spending the same time increasing the number of instrumental replicates, as there is more to gain in reducing the expanded uncertainty.

Figure 4.1: Relationship between the t-distribution and the Normal distribution at a 95% Confidence Level, for low values of n (degrees of freedom ($n-1$) between 1-35).




Table 4.2: Summary Statistics for L and D Aspartic Acid / Asparagine Peak Area Data

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL					
		L-Asx peak area	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	30650	30729	40407	41620	42489	43184	43319	44541	45353			40255	9	5614.8	13.95	1871.6	4.65	9.30	2.306	10.72
2	RP	20767	20615										20691	2	107.6	0.52	76.1	0.37	0.74	12.710	4.67
3	RP	22708											22708	1							
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP	50180	51126										50653	2	668.9	1.32	473.0	0.93	1.87	12.710	11.87
9	RP	49157	50630										49893	2	1041.8	2.09	736.6	1.48	2.95	12.710	18.77
10	RP	28983	32086										30535	2	2193.9	7.18	1551.3	5.08	10.16	12.710	64.57
11	RP	10104	10233										10168	2	90.9	0.89	64.3	0.63	1.26	12.710	8.04
12	RP	9238	9439										9338	2	142.4	1.53	100.7	1.08	2.16	12.710	13.71
13	RP	18046											18046	1							
14	RP	13534											13534	1							
15	RP	10678	10857										10767	2	126.5	1.17	89.4	0.83	1.66	12.710	10.56
D-Asx peak area	a	b	c	d	e	f	g	h	i	j		mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})	
1	RP	5903	5840	7764	7970	8152	8304	8280	8584	8751	34320		10387	10	8472.1	81.57	2679.1	25.79	51.59	2.262	58.35
2	RP	4105	4072										4088	2	22.9	0.56	16.2	0.40	0.79	12.710	5.04
3	RP	4763											4763	1							
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP	10900	11107										11003	2	147.0	1.34	103.9	0.94	1.89	12.710	12.00
9	RP	11426	11749										11587	2	228.8	1.97	161.8	1.40	2.79	12.710	17.74
10	RP	6547	7251										6899	2	497.8	7.22	352.0	5.10	10.21	12.710	64.85
11	RP	2237	2282										2259	2	31.8	1.41	22.5	1.00	1.99	12.710	12.66
12	RP	2103	2148										2125	2	31.5	1.48	22.3	1.05	2.10	12.710	13.33
13	RP	4029											4029	1							
14	RP	3043											3043	1							
15	RP	2371	2386										2379	2	10.6	0.45	7.5	0.32	0.63	12.710	4.01

Table 4.3: Summary Statistics for L and D Aspartic Acid / Asparagine Concentration Data (pM)

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL					
		L-Asx Conc	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	8328	8295	8662	8835	9063	9013	9026	8989	8883		8788	9	296.8	3.38	98.9	1.13	2.25	2.306	2.60	
2	RP																				
3	RP																				
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP	7185	7223										7204	2	26.6	0.37	18.8	0.26	0.52	12.710	3.32
9	RP	7167	7323										7245	2	110.3	1.52	78.0	1.08	2.15	12.710	13.68
10	RP	8086	8029										8058	2	40.4	0.50	28.5	0.35	0.71	12.710	4.50
11	RP	5408	5150										5279	2	182.7	3.46	129.2	2.45	4.89	12.710	31.10
12	RP	11365	12118										11742	2	532.5	4.54	376.6	3.21	6.41	12.710	40.76
13	RP	12228											12228	1							
14	RP	13866											13866	1							
15	RP	10104	10793										10449	2	487.2	4.66	344.5	3.30	6.59	12.710	41.90
D-Asx Conc	a	b	c	d	e	f	g	h	i	j		mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})	
1	RP	1604	1577	1664	1692	1739	1733	1725	1732	1714	1656	1684	10	57.3	3.40	18.1	1.08	2.15	2.262	2.43	
2	RP																				
3	RP																				
4	IE																				
5	IE																				
6.1	GC												1565	2	6.0	0.38	4.3	0.27	0.54	12.710	3.45
6.2	GC												1683	2	23.7	1.41	16.8	1.00	1.99	12.710	12.66
7.1	GC												1820	2	8.6	0.47	6.0	0.33	0.66	12.710	4.22
7.2	GC												1173	2	34.6	2.95	24.4	2.08	4.17	12.710	26.48
8	RP	1561	1569										2672	2	120.1	4.49	84.9	3.18	6.35	12.710	40.38
9	RP	1666	1699										2730	1							
10	RP	1826	1814										3118	1							
11	RP	1197	1148										2308	2	90.8	3.93	64.2	2.78	5.56	12.710	35.36
12	RP	2587	2757																		
13	RP	2730																			
14	RP	3118																			
15	RP	2244	2372																		

Table 4.4: Summary Statistics for L and D Aspartic Acid / Asparagine D/L Ratio Value

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL				
		D/L Asx	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)
1	RP	0.193	0.190	0.192	0.191	0.192	0.192	0.191	0.193	0.193		0.192	9	0.0009	0.47	0.0003	0.16	0.32	2.306	0.36
2	RP	0.198		0.198								0.198	2	0.0001	0.04	0.0001	0.03	0.06	12.710	0.36
3	RP	0.210										0.210	1							
4	IE																			
5	IE																			
6.1 ¹	GC _A	0.225										0.225	4	0.0200	8.89	0.0100	4.44	8.89	3.182	14.14
6.2 ¹	GC _H	0.299										0.299	2	0.0030	1.00	0.0021	0.71	1.42	12.710	9.02
7.1 ¹	GC _A	0.216										0.216	1							
7.2 ¹	GC _H	0.280										0.280	1							
8	RP	0.217	0.217									0.217	2	0.0000	0.00	0.0000	0.00	0.00	12.710	0.00
9	RP	0.232	0.232									0.232	2	0.0003	0.11	0.0002	0.08	0.16	12.710	1.02
10	RP	0.226	0.226									0.226	2	0.0001	0.03	0.0000	0.02	0.04	12.710	0.28
11	RP	0.221	0.223									0.222	2	0.0011	0.51	0.0008	0.36	0.73	12.710	4.62
12	RP	0.228	0.228									0.228	2	0.0001	0.04	0.0001	0.03	0.06	12.710	0.38
13	RP	0.223										0.223	1							
14	RP	0.225										0.225	1							
15	RP	0.222	0.220									0.221	2	0.0016	0.73	0.0011	0.52	1.03	12.710	6.55

¹= submitted as the mean and standard deviation of n results.

GC_A = derived using peak area

GC_H = derived using peak height

Figure 4.2: Distribution of D/L Values submitted for Aspartic Acid / Asparagine




Figure 4.3: Experimental Expanded Uncertainty ($k=2$) of the Mean D/L value for Aspartic Acid / Asparagine (value of n displayed).




Figure 4.4: Experimental Expanded Uncertainty ($k=t_{(0.05,df)}$) of the Mean D/L value for Aspartic Acid / Asparagine (value of n displayed).




Table 4.5: Summary Statistics for L and D Glutamic Acid / Glutamine Peak Area Data

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL					
		L-Glx peak area	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	39863	40945	53183	53709	54941	55587	55958	57877	58308			52264	9	6934.8	13.27	2311.6	4.42	8.85	2.306	10.20
2	RP	25965	25490										25727	2	335.7	1.30	237.4	0.92	1.85	12.710	11.73
3	RP	29079											29079	1							
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP	64574	65981										65277	2	995.2	1.52	703.7	1.08	2.16	12.710	13.70
9	RP	61208	62517										61862	2	926.0	1.50	654.8	1.06	2.12	12.710	13.45
10	RP	36384	39493										37939	2	2198.5	5.79	1554.6	4.10	8.20	12.710	52.08
11	RP	12746	12923										12834	2	125.4	0.98	88.6	0.69	1.38	12.710	8.78
12	RP	11600	11873										11737	2	193.3	1.65	136.7	1.16	2.33	12.710	14.80
13	RP	23093											23093	1							
14	RP	16708											16708	1							
15	RP	13068	13160										13114	2	65.2	0.50	46.1	0.35	0.70	12.710	4.47
D-Glx peak area	a	b	c	d	e	f	g	h	i	j		mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})	
1	RP	1962	2014	2630	2655	2714	2752	2766	2856	2891	11719		3496	10	2907.8	83.18	919.5	26.30	52.60	2.262	59.50
2	RP	1147	1117										1132	2	21.4	1.89	15.1	1.34	2.67	12.710	16.98
3	RP	1308											1308	1							
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP	3885	3973										3929	2	62.2	1.58	44.0	1.12	2.24	12.710	14.22
9	RP	4072	4168										4120	2	68.1	1.65	48.1	1.17	2.34	12.710	14.85
10	RP	2342	2578										2460	2	167.1	6.79	118.2	4.80	9.61	12.710	61.05
11	RP	825	829										827	2	2.6	0.31	1.8	0.22	0.44	12.710	2.77
12	RP	744	770										757	2	18.3	2.42	13.0	1.71	3.43	12.710	21.78
13	RP	1428											1428	1							
14	RP	1051											1051	1							
15	RP	810	817										813	2	4.4	0.55	3.1	0.39	0.77	12.710	4.90

Table 4.6: Summary Statistics for L and D Glutamic Acid / Glutamine Concentration Data (pM)

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL					
		L-Glx Conc	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	10831	11053	11401	11401	11719	11601	11659	11681	11421		11419	9	301.9	2.64	100.6	0.88	1.76	2.306	2.03	
2	RP																				
3	RP																				
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP	9246	9321										9284	2	53.2	0.57	37.6	0.41	0.81	12.710	5.15
9	RP	9316	9440										9378	2	87.3	0.93	61.8	0.66	1.32	12.710	8.37
10	RP	10598	10318										10458	2	198.1	1.89	140.1	1.34	2.68	12.710	17.02
11	RP	7123	6790										6956	2	235.0	3.38	166.2	2.39	4.78	12.710	30.36
12	RP	14900	15915										15407	2	717.5	4.66	507.3	3.29	6.59	12.710	41.85
13	RP	16336											16336	1							
14	RP	17871											17871	1							
15	RP	12911	13659										13285	2	529.5	3.99	374.4	2.82	5.64	12.710	35.82
D-Glx Conc	a	b	c	d	e	f	g	h	i	j		mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})	
1	RP	533	544	564	564	579	574	576	576	566	565	564	10	14.9	2.65	4.7	0.84	1.67	2.262	1.89	
2	RP																				
3	RP																				
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP	556	561										559	2	3.5	0.63	2.5	0.45	0.89	12.710	5.67
9	RP	620	629										625	2	6.8	1.09	4.8	0.77	1.54	12.710	9.77
10	RP	682	674										678	2	6.1	0.89	4.3	0.63	1.26	12.710	8.03
11	RP	461	435										448	2	18.1	4.05	12.8	2.86	5.72	12.710	36.36
12	RP	956	1032										994	2	54.0	5.43	38.2	3.84	7.68	12.710	48.82
13	RP	1011											1011	1							
14	RP	1124											1124	1							
15	RP	800	848										824	2	33.2	4.03	23.5	2.85	5.70	12.710	36.25

Table 4.7: Summary Statistics for L and D Glutamic Acid / Glutamine D/L Ratio Value

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL				
		D/L Glx	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)
1	RP	0.049	0.049	0.049	0.049	0.049	0.050	0.049	0.049	0.050		0.049	9	0.0001	0.25	0.0000	0.08	0.17	2.306	0.19
2	RP	0.044	0.044									0.044	2	0.0003	0.58	0.0002	0.41	0.83	12.710	5.25
3	RP	0.045										0.045	1							
4	IE																			
5	IE																			
6.1 ¹	GC _A	0.059										0.059	4	0.0080	13.56	0.0040	6.78	13.56	3.182	21.58
6.2	GC _H																			
7.1 ¹	GC _A	0.057										0.057	1							
7.2	GC _H																			
8	RP	0.060	0.060									0.060	2	0.0000	0.00	0.0000	0.00	0.00	12.710	0.00
9	RP	0.067	0.067									0.067	2	0.0001	0.16	0.0001	0.11	0.22	12.710	1.40
10	RP	0.064	0.065									0.065	2	0.0006	1.00	0.0005	0.71	1.41	12.710	8.99
11	RP	0.065	0.064									0.064	2	0.0004	0.67	0.0003	0.47	0.94	12.710	6.00
12	RP	0.064	0.065									0.064	2	0.0005	0.78	0.0004	0.55	1.10	12.710	6.98
13	RP	0.062										0.062	1							
14	RP	0.063										0.063	1							
15	RP	0.062	0.062									0.062	2	0.0000	0.05	0.0000	0.03	0.07	12.710	0.43

¹= submitted as the mean and standard deviation of n results.

GC_A = derived using peak area

GC_H = derived using peak height


Figure 4.5: Distribution of D/L Values submitted for **Glutamic Acid / Glutamine**

Figure 4.6: Experimental Expanded Uncertainty ($k=2$) of the Mean D/L value for Glutamic Acid / Glutamine (value of n displayed).




Figure 4.7: Experimental Expanded Uncertainty ($k=t_{(0.05,df)}$) of the Mean D/L value for Glutamic Acid / Glutamine (value of n displayed).




Table 4.8: Summary Statistics for L and D Serine Peak Area Data

Lab No	method	Submitted Replicate data										Standard Deviation			Uncertainty of Mean & Expanded U at 95% CL								
		L-Ser peak area		a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})	
1	RP	2517	2641	3393	3438	3404	3550	3563	3668	3694				3319	9	433.9	13.08	144.6	4.36	8.72	2.306	10.05	
2	RP	15463	15194											15329	2	190.4	1.24	134.6	0.88	1.76	12.710	11.16	
3	RP	16560												16560	1								
4	IE																						
5	IE																						
6.1	GC																						
6.2	GC																						
7.1	GC																						
7.2	GC																						
8	RP	37245	38118												37682	2	617.2	1.64	436.4	1.16	2.32	12.710	14.72
9	RP	36846	37756												37301	2	643.4	1.72	455.0	1.22	2.44	12.710	15.50
10	RP	21418	23204												22311	2	1262.7	5.66	892.8	4.00	8.00	12.710	50.86
11	RP	7339	7537												7438	2	140.3	1.89	99.2	1.33	2.67	12.710	16.95
12	RP	6704	6863												6784	2	112.4	1.66	79.5	1.17	2.34	12.710	14.89
13	RP	13454													13454	1							
14	RP	9609													9609	1							
15	RP	7696	7781												7739	2	60.1	0.78	42.5	0.55	1.10	12.710	6.98
D-Ser peak area		a	b	c	d	e	f	g	h	i	j				mean	n	std dev	CV%	std u	RSU%			
1	RP	23353	24237	31680	31490	32297	32873	32755	33829	33964					30720	9	4019.9	13.09	1340.0	4.36	8.72	2.306	10.06
2	RP	1721	1696												1708	2	18.0	1.05	12.7	0.74	1.49	12.710	9.45
3	RP	1918													1918	1							
4	IE																						
5	IE																						
6.1	GC																						
6.2	GC																						
7.1	GC																						
7.2	GC																						
8	RP	4365	4469												4417	2	73.6	1.67	52.0	1.18	2.36	12.710	14.97
9	RP	4212	4324												4268	2	79.0	1.85	55.9	1.31	2.62	12.710	16.64
10	RP	2389	2587												2488	2	140.2	5.64	99.2	3.99	7.97	12.710	50.65
11	RP	809	819												814	2	7.2	0.88	5.1	0.62	1.25	12.710	7.93
12	RP	743	761												752	2	12.9	1.71	9.1	1.21	2.42	12.710	15.37
13	RP	1503													1503	1							
14	RP	1105													1105	1							
15	RP	881	891												886	2	7.1	0.80	5.0	0.56	1.13	12.710	7.18

Table 4.9: Summary Statistics for L and D Serine Concentration Data (pM)

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL						
		L-Ser Conc		a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	6345	6543	6791	6684	6889	6861	6825	6827	6652				6713	9	177.9	2.65	59.3	0.88	1.77	2.306	2.04
2	RP																					
3	RP																					
4	IE																					
5	IE																					
6.1	GC																					
6.2	GC																					
7.1	GC																					
7.2	GC																					
8	RP	5333	5385											5359	2	36.8	0.69	26.0	0.49	0.97	12.710	6.17
9	RP	5576	5668											5622	2	65.2	1.16	46.1	0.82	1.64	12.710	10.42
10	RP	6202	6027											6114	2	124.1	2.03	87.7	1.44	2.87	12.710	18.24
11	RP	4077	3937											4007	2	98.9	2.47	69.9	1.75	3.49	12.710	22.19
12	RP	8561	9145											8853	2	413.1	4.67	292.1	3.30	6.60	12.710	41.94
13	RP	9462												9462	1							
14	RP	10218												10218	1							
15	RP	7559	8030											7794	2	332.4	4.26	235.1	3.02	6.03	12.710	38.33
D-Ser Conc		a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})		
1	RP	684	713	727	730	726	741	742	740	724	692	722	10	20.2	2.79	6.4	0.88	1.77	2.262	2.00		
2	RP																					
3	RP																					
4	IE																					
5	IE																					
6.1	GC																					
6.2	GC																					
7.1	GC																					
7.2	GC																					
8	RP	625	631									628	2	4.5	0.71	3.2	0.51	1.01	12.710	6.42		
9	RP	637	649									643	2	8.3	1.29	5.8	0.91	1.82	12.710	11.55		
10	RP	692	672									682	2	14.0	2.05	9.9	1.45	2.90	12.710	18.45		
11	RP	450	428									439	2	15.2	3.47	10.8	2.46	4.91	12.710	31.20		
12	RP	948	1014									981	2	46.3	4.72	32.7	3.34	6.67	12.710	42.42		
13	RP	1057										1057	1									
14	RP	1175										1175	1									
15	RP	865	919									892	2	38.2	4.29	27.0	3.03	6.06	12.710	38.53		

Table 4.10: Summary Statistics for L and D Serine D/L Ratio Value

Lab No	method	Submitted Replicate data										Standard Deviation			Uncertainty of Mean & Expanded U at 95% CL					
		D/L Serine	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)
1	RP	0.108	0.109	0.107	0.109	0.105	0.108	0.109	0.108	0.109		0.108	9	0.0012	1.10	0.0004	0.37	0.73	2.306	0.85
2	RP	0.111	0.112									0.111	2	0.0002	0.19	0.0002	0.13	0.27	12.710	1.71
3	RP	0.116										0.116	1							
4	IE																			
5	IE																			
6.1	GC																			
6.2	GC																			
7.1	GC																			
7.2	GC																			
8	RP	0.117	0.117									0.117	2	0.0000	0.00	0.0000	0.00	0.00	12.710	0.00
9	RP	0.114	0.115									0.114	2	0.0001	0.13	0.0001	0.09	0.18	12.710	1.14
10	RP	0.112	0.112									0.112	2	0.0000	0.02	0.0000	0.02	0.03	12.710	0.21
11	RP	0.110	0.109									0.109	2	0.0011	1.00	0.0008	0.71	1.42	12.710	9.02
12	RP	0.111	0.111									0.111	2	0.0001	0.05	0.0000	0.04	0.08	12.710	0.48
13	RP	0.112										0.112	1							
14	RP	0.115										0.115	1							
15	RP	0.114	0.114									0.114	2	0.0000	0.02	0.0000	0.02	0.03	12.710	0.20

Figure 4.8: Distribution of D/L Values submitted for Serine




Figure 4.9: Experimental Expanded Uncertainty ($k=2$) of the Mean D/L value for Serine (value of n displayed).




Figure 4.10: Experimental Expanded Uncertainty ($k=t_{(0.05,df)}$) of the Mean D/L value for Serine (value of n displayed).




Table 4.11: Summary Statistics for L and D Arginine Peak Area Data

Lab No	method	Submitted Replicate data										Standard Deviation			Uncertainty of Mean & Expanded U at 95% CL						
		L-Arg peak area	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP																				
2	RP	13519	13201										13360	2	225.0	1.68	159.1	1.19	2.38	12.710	15.13
3	RP	14617											14617	1							
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP																				
9	RP	32911	34644										33778	2	1225.3	3.63	866.4	2.57	5.13	12.710	32.60
10	RP	19420	21029										20225	2	1137.9	5.63	804.6	3.98	7.96	12.710	50.56
11	RP	6819	6961										6890	2	100.1	1.45	70.8	1.03	2.05	12.710	13.05
12	RP	6479	6565										6522	2	60.8	0.93	43.0	0.66	1.32	12.710	8.37
13	RP	12550											12550	1							
14	RP	9108											9108	1							
15	RP	7179	6930										7055	2	176.5	2.50	124.8	1.77	3.54	12.710	22.48
D-Arg peak area	a	b	c	d	e	f	g	h	i	j		mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})	
1	RP																				
2	RP	1187	1157										1172	2	20.7	1.77	14.6	1.25	2.50	12.710	15.88
3	RP	1528											1528	1							
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP																				
9	RP	3795	3935										3865	2	99.4	2.57	70.3	1.82	3.64	12.710	23.12
10	RP	2086	2328										2207	2	171.1	7.75	121.0	5.48	10.96	12.710	69.66
11	RP	642	700										671	2	40.9	6.10	28.9	4.31	8.63	12.710	54.83
12	RP	676	626										651	2	35.4	5.44	25.1	3.85	7.70	12.710	48.92
13	RP	1188											1188	1							
14	RP	908											908	1							
15	RP	730	663										696	2	47.3	6.79	33.4	4.80	9.60	12.710	61.01

Table 4.12: Summary Statistics for L and D Arginine Concentration Data (pM)

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL					
		L-Arg Conc	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP																				
2	RP																				
3	RP																				
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP																				
9	RP	4904	5121										5013	2	153.5	3.06	108.5	2.17	4.33	12.710	27.52
10	RP	5538	5378										5458	2	112.6	2.06	79.6	1.46	2.92	12.710	18.54
11	RP	3730	3580										3655	2	106.1	2.90	75.0	2.05	4.10	12.710	26.09
12	RP	8147	8615										8381	2	330.4	3.94	233.6	2.79	5.58	12.710	35.43
13	RP	8691											8691	1							
14	RP	9537											9537	1							
15	RP	6944	7041										6993	2	69.0	0.99	48.8	0.70	1.40	12.710	8.87
D-Arg Conc		a	b	c	d	e	f	g	h	i	j		mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP																				
2	RP																				
3	RP																				
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP																				
9	RP	565	582										574	2	11.5	2.01	8.1	1.42	2.84	12.710	18.04
10	RP	595	595										595	2	0.4	0.07	0.3	0.05	0.09	12.710	0.60
11	RP	351	360										356	2	6.2	1.75	4.4	1.24	2.47	12.710	15.72
12	RP	850	822										836	2	20.3	2.43	14.4	1.72	3.44	12.710	21.88
13	RP	823											823	1							
14	RP	951											951	1							
15	RP	706	674										690	2	22.8	3.30	16.1	2.34	4.67	12.710	29.69

Table 4.13: Summary Statistics for L and D Arginine D/L Ratio Value

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL					
		D/L Arg	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP																				
2	RP	0.088	0.088										0.088	2	0.0001	0.08	0.0001	0.06	0.12	12.710	0.74
3	RP	0.105											0.105	1							
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP																				
9	RP	0.115	0.114										0.114	2	0.0012	1.06	0.0009	0.75	1.49	12.710	9.49
10	RP	0.107	0.111										0.109	2	0.0023	2.13	0.0016	1.51	3.01	12.710	19.14
11	RP	0.094	0.101										0.097	2	0.0045	4.65	0.0032	3.29	6.58	12.710	41.80
12	RP	0.104	0.095										0.100	2	0.0064	6.37	0.0045	4.51	9.01	12.710	57.28
13	RP	0.095											0.095	1							
14	RP	0.100											0.100	1							
15	RP	0.102	0.096										0.099	2	0.0042	4.29	0.0030	3.03	6.07	12.710	38.56

Figure 4.11: Distribution of D/L Values submitted for Arginine




Figure 4.12: Experimental Expanded Uncertainty ($k=2$) of the Mean D/L value for Arginine (value of n displayed).




Figure 4.13: Experimental Expanded Uncertainty ($k=t_{(0.05,df)}$) of the Mean D/L value for Arginine (value of n displayed).




Table 4.14: Summary Statistics for L and D Alanine Peak Area Data

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL				
		L-Ala peak area	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)
1	RP	31343	13057	42749	42429	43434	43783	44162	45820	46076		39206	9	10740.5	27.40	3580.2	9.13	18.26	2.306	21.06
2	RP	21971	21522									21747	2	317.9	1.46	224.8	1.03	2.07	12.710	13.14
3	RP	23838										23838	1							
4	IE																			
5	IE																			
6.1	GC																			
6.2	GC																			
7.1	GC																			
7.2	GC																			
8	RP	56516	57761									57138	2	880.4	1.54	622.6	1.09	2.18	12.710	13.85
9	RP	50258	52453									51355	2	1552.6	3.02	1097.8	2.14	4.28	12.710	27.17
10	RP	30287	32913									31600	2	1856.8	5.88	1313.0	4.15	8.31	12.710	52.81
11	RP	10247	10472									10359	2	159.3	1.54	112.7	1.09	2.18	12.710	13.82
12	RP	9490	9634									9562	2	102.3	1.07	72.4	0.76	1.51	12.710	9.62
13	RP	18110										18110	1							
14	RP	13760										13760	1							
15	RP	11070	10837									10954	2	164.8	1.50	116.5	1.06	2.13	12.710	13.52
D-Ala peak area	a	b	c	d	e	f	g	h	i	j		mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	1809	1086	2359	2409	2411	2516	2548	2578	2672	10160	3055	10	2541.7	83.20	803.7	26.31	52.62	2.262	59.52
2	RP	1268	1240									1254	2	20.0	1.60	14.1	1.13	2.26	12.710	14.34
3	RP	1638										1638	1							
4	IE																			
5	IE																			
6.1	GC																			
6.2	GC																			
7.1	GC																			
7.2	GC																			
8	RP	3502	3572									3537	2	49.7	1.40	35.1	0.99	1.99	12.710	12.62
9	RP	4373	4489									4431	2	81.8	1.85	57.8	1.31	2.61	12.710	16.59
10	RP	2478	2667									2572	2	133.2	5.18	94.2	3.66	7.32	12.710	46.55
11	RP	778	813									795	2	25.4	3.20	18.0	2.26	4.52	12.710	28.74
12	RP	736	761									749	2	17.7	2.36	12.5	1.67	3.34	12.710	21.25
13	RP	1174										1174	1							
14	RP	1025										1025	1							
15	RP	893	880									886	2	9.2	1.04	6.5	0.74	1.47	12.710	9.35

Table 4.15: Summary Statistics for L and D Alanine Concentration Data (pM)

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL					
		L-Ala Conc	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	8516	3525	9164	9007	9264	9138	9202	9248	9025		8454	9	1862.4	22.03	620.8	7.34	14.69	2.306	16.93	
2	RP																				
3	RP																				
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP	8092	8160										8126	2	47.9	0.59	33.9	0.42	0.83	12.710	5.30
9	RP	7070	7320										7195	2	176.8	2.46	125.0	1.74	3.48	12.710	22.09
10	RP	8153	7947										8050	2	145.9	1.81	103.2	1.28	2.56	12.710	16.29
11	RP	5292	5085										5189	2	146.2	2.82	103.3	1.99	3.98	12.710	25.32
12	RP	11265	11934										11600	2	473.3	4.08	334.7	2.89	5.77	12.710	36.67
13	RP	11840											11840	1							
14	RP	13602											13602	1							
15	RP	10108	10396										10252	2	203.5	1.99	143.9	1.40	2.81	12.710	17.84
D-Ala Conc	a	b	c	d	e	f	g	h	i	j		mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})	
1	RP	492	293	506	511	514	525	531	520	523	490	491	10	70.7	14.41	22.4	4.56	9.11	2.262	10.31	
2	RP																				
3	RP																				
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP	501	505										503	2	2.3	0.45	1.6	0.32	0.64	12.710	4.07
9	RP	521	531										526	2	6.7	1.28	4.8	0.91	1.81	12.710	11.50
10	RP	565	546										556	2	13.9	2.51	9.9	1.77	3.55	12.710	22.56
11	RP	340	335										338	2	3.9	1.16	2.8	0.82	1.64	12.710	10.40
12	RP	740	799										770	2	41.4	5.37	29.2	3.80	7.60	12.710	48.30
13	RP	650											650	1							
14	RP	859											859	1							
15	RP	691	715										703	2	17.2	2.45	12.2	1.73	3.46	12.710	22.01

Table 4.16: Summary Statistics for L and D Alanine D/L Ratio Value

Lab No	method	Submitted Replicate data										Standard Deviation			Uncertainty of Mean & Expanded U at 95% CL					
		D/L Ala	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)
1	RP	0.058	0.083	0.055	0.057	0.056	0.057	0.058	0.056	0.058		0.060	9	0.0088	14.80	0.0029	4.93	9.87	2.306	11.38
2	RP	0.058	0.058									0.058	2	0.0001	0.13	0.0001	0.09	0.19	12.710	1.20
3	RP	0.069										0.069	1							
4	IE																			
5	IE																			
6.1 ¹	GC _A	0.047										0.047	8	0.0150	31.91	0.0053	11.28	22.57	2.365	26.68
6.2 ¹	GC _H	0.050										0.050	5	0.0010	2.00	0.0004	0.89	1.79	2.777	2.48
7.1 ¹	GC _A	0.048										0.048	1							
7.2 ¹	GC _H	0.047										0.047	1							
8	RP	0.062	0.062									0.062	2	0.0000	0.00	0.0000	0.00	0.00	12.710	0.00
9	RP	0.074	0.073									0.073	2	0.0009	1.18	0.0006	0.83	1.67	12.710	10.59
10	RP	0.069	0.069									0.069	2	0.0005	0.70	0.0003	0.49	0.99	12.710	6.27
11	RP	0.064	0.066									0.065	2	0.0011	1.66	0.0008	1.17	2.35	12.710	14.92
12	RP	0.066	0.067									0.066	2	0.0009	1.29	0.0006	0.92	1.83	12.710	11.64
13	RP	0.055										0.055	1							
14	RP	0.063										0.063	1							
15	RP	0.068	0.069									0.069	2	0.0003	0.46	0.0002	0.33	0.66	12.710	4.17

¹= submitted as the mean and standard deviation of n results.

GC_A = derived using peak area

GC_H = derived using peak height

Figure 4.14: Distribution of D/L Values submitted for Alanine




Figure 4.15: Experimental Expanded Uncertainty ($k=2$) of the Mean D/L value for Alanine (value of n displayed).




Figure 4.16: Experimental Expanded Uncertainty ($k=t_{0.05,df}$) of the Mean D/L value for Alanine (value of n displayed).




Table 4.17: Summary Statistics for L and D Valine Peak Area / Height Data

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL						
		L-Val peak area		a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	18644	18070	23928	25047	25603	25960	26239	26953	27672	109142		32726	10	27052.1	82.66	8554.6	26.14	52.28	2.262	59.13	
2	RP	11564	11539									11551	2	17.4	0.15	12.3	0.11	0.21	12.710	1.36		
3	RP	12977										12977	1									
4	IE																					
5	IE																					
6.1	GC																					
6.2	GC																					
7.1	GC																					
7.2	GC																					
8	RP	33202	33810											33506	2	429.9	1.28	304.0	0.91	1.81	12.710	11.53
9	RP	33859	34999											34429	2	806.2	2.34	570.1	1.66	3.31	12.710	21.05
10	RP	19158	21095											20126	2	1369.3	6.80	968.3	4.81	9.62	12.710	61.15
11	RP	6796	6722											6759	2	51.6	0.76	36.5	0.54	1.08	12.710	6.87
12	RP	6269	6411											6340	2	100.3	1.58	70.9	1.12	2.24	12.710	14.22
13	RP	12386												12386	1							
14	RP	8687												8687	1							
15	RP	6930	6964											6947	2	24.1	0.35	17.0	0.24	0.49	12.710	3.11
D-Val peak area		a	b	c	d	e	f	g	h	i	j	Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL						
D-Val peak area		a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})		
1	RP	328	310	417	441	450	453	453	464	488	1880		568	10	464.4	81.69	146.8	25.83	51.67	2.262	58.44	
2	RP	244	231									238	2	9.1	3.85	6.5	2.72	5.44	12.710	34.60		
3	RP	268										268	1									
4	IE																					
5	IE																					
6.1	GC																					
6.2	GC																					
7.1	GC																					
7.2	GC																					
8	RP	715	645											680	2	49.2	7.24	34.8	5.12	10.24	12.710	65.08
9	RP	808	852											830	2	31.0	3.74	21.9	2.64	5.29	12.710	33.61
10	RP	412	491											451	2	55.9	12.40	39.5	8.77	17.53	12.710	111.43
11	RP	143	148											146	2	3.3	2.24	2.3	1.59	3.17	12.710	20.16
12	RP	151	133											142	2	12.9	9.05	9.1	6.40	12.79	12.710	81.29
13	RP	215												215	1							
14	RP	225												225	1							
15	RP	164	151											158	2	8.9	5.66	6.3	4.00	8.00	12.710	50.87
D+L Val peak height		a	b	c	d	e	f	g	h	i	j	Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL						
D+L Val peak height		a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})		
004	IE	11.173	11.165											11.169	2	0.0057	0.05	0.0040	0.04	0.07	12.710	0.46
005	IE	11.639	11.520											11.580	2	0.0841	0.73	0.0595	0.51	1.03	12.710	6.53

Table 4.18: Summary Statistics for L and D Valine Concentration Data (pM)

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL						
		L-Val Conc		a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	5066	4878	5129	5317	5461	5418	5467	5440	5420	5265		5286	10	200.9	3.80	63.5	1.20	2.40	2.262	2.72	
2	RP																					
3	RP																					
4	IE																					
5	IE																					
6.1	GC																					
6.2	GC																					
7.1	GC																					
7.2	GC																					
8	RP	4754	4777											4765	2	15.8	0.33	11.2	0.23	0.47	12.710	2.98
9	RP	4391	4502											4446	2	79.0	1.78	55.8	1.26	2.51	12.710	15.96
10	RP	4754	4695											4725	2	41.7	0.88	29.5	0.62	1.25	12.710	7.94
11	RP	3235	3009											3122	2	159.8	5.12	113.0	3.62	7.24	12.710	45.99
12	RP	6860	7321											7091	2	325.6	4.59	230.3	3.25	6.49	12.710	41.27
13	RP	7464												7464	1							
14	RP	7916												7916	1							
15	RP	5833	6158											5996	2	229.9	3.83	162.6	2.71	5.42	12.710	34.46
D-Val Conc		a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})		
1	RP	89	84	89	94	96	94	94	94	96	91	92	10	3.8	4.14	1.2	1.31	2.62	2.262	2.96		
2	RP																					
3	RP																					
4	IE																					
5	IE																					
6.1	GC																					
6.2	GC																					
7.1	GC																					
7.2	GC																					
8	RP	102	91									97	2	7.9	8.19	5.6	5.79	11.58	12.710	73.61		
9	RP	94	99									97	2	3.1	3.17	2.2	2.24	4.49	12.710	28.53		
10	RP	92	98									95	2	4.5	4.74	3.2	3.35	6.70	12.710	42.56		
11	RP	62	60									61	2	1.3	2.11	0.9	1.49	2.99	12.710	18.98		
12	RP	149	137									143	2	8.6	6.04	6.1	4.27	8.54	12.710	54.30		
13	RP	117										117	1									
14	RP	184										184	1									
15	RP	124	121									122	2	2.7	2.17	1.9	1.54	3.07	12.710	19.53		

Table 4.19: Summary Statistics for L and D Valine D/L Ratio Value

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL				
		D/L Valine	a	b	c	d	e	f	g	h	i	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	0.018	0.017	0.017	0.018	0.018	0.017	0.017	0.017	0.018	0.017	0.017	10	0.0002	1.08	0.0001	0.34	0.68	2.262	0.77
2	RP	0.021	0.020									0.021	2	0.0008	3.70	0.0005	2.62	5.23	12.710	33.25
3	RP	0.021										0.021	1							
4	IE																			
5	IE																			
6.1 ¹	GC _A	0.015										0.015	7	0.0010	6.67	0.0004	2.52	5.04	2.447	6.17
6.2 ¹	GC _H	0.014										0.014	5	0.0010	7.14	0.0004	3.19	6.39	2.777	8.87
7.1 ¹	GC _A	0.012										0.012	1							
7.2 ¹	GC _H	0.014										0.014	1							
8	RP	0.022	0.019									0.021	2	0.0021	10.35	0.0015	7.32	14.63	12.710	93.00
9	RP	0.022	0.022									0.022	2	0.0003	1.40	0.0002	0.99	1.98	12.710	12.57
10	RP	0.019	0.021									0.020	2	0.0011	5.62	0.0008	3.97	7.95	12.710	50.49
11	RP	0.019	0.020									0.019	2	0.0006	3.01	0.0004	2.13	4.25	12.710	27.02
12	RP	0.022	0.019									0.020	2	0.0021	10.62	0.0015	7.51	15.02	12.710	95.44
13	RP	0.016										0.016	1							
14	RP	0.023										0.023	1							
15	RP	0.021	0.020									0.020	2	0.0012	6.01	0.0009	4.25	8.49	12.710	53.98

¹= submitted as the mean and standard deviation of n results.

GC_A = derived using peak area

GC_H = derived using peak height

Figure 4.17: Distribution of D/L Values submitted for Valine




Figure 4.18: Experimental Expanded Uncertainty ($k=2$) of the Mean D/L value for Valine (value of n displayed).




Figure 4.19: Experimental Expanded Uncertainty ($k=t_{0.05,df}$) of the Mean D/L value for Valine (value of n displayed).




Table 4.20: Summary Statistics for L and D Phenylalanine Peak Area Data

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL				
		L-Phe peak area	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)
1	RP	15132	15729	20240	20332	20789	21030	21149	21712	22027	90056	26820	10	22347.9	83.33	7067.0	26.35	52.70	2.262	59.61
2	RP	10601	10340									10470	2	184.6	1.76	130.5	1.25	2.49	12.710	15.85
3	RP	11936										11936	1							
4	IE																			
5	IE																			
6.1	GC																			
6.2	GC																			
7.1	GC																			
7.2	GC																			
8	RP	26403	27078									26740	2	477.6	1.79	337.7	1.26	2.53	12.710	16.05
9	RP	25546	26167									25856	2	438.6	1.70	310.1	1.20	2.40	12.710	15.24
10	RP	14869	15834									15351	2	682.6	4.45	482.7	3.14	6.29	12.710	39.96
11	RP	5027	5088									5058	2	42.8	0.85	30.2	0.60	1.20	12.710	7.60
12	RP	4778	4865									4821	2	61.5	1.27	43.5	0.90	1.80	12.710	11.46
13	RP	8987										8987	1							
14	RP	6844										6844	1							
15	RP	5302	5319									5310	2	12.5	0.24	8.9	0.17	0.33	12.710	2.12
D-Phe peak area	a	b	c	d	e	f	g	h	i	j		mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	762	768	980	988	1021	1058	1066	1143	1159	4544	1349	10	1130.9	83.84	357.6	26.51	53.02	2.262	59.97
2	RP	494	472									483	2	15.8	3.26	11.1	2.31	4.61	12.710	29.30
3	RP	587										587	1							
4	IE																			
5	IE																			
6.1	GC																			
6.2	GC																			
7.1	GC																			
7.2	GC																			
8	RP	1351	1381									1366	2	21.3	1.56	15.0	1.10	2.20	12.710	14.00
9	RP	1527	1555									1541	2	20.1	1.31	14.2	0.92	1.85	12.710	11.73
10	RP	852	926									889	2	52.3	5.88	37.0	4.16	8.32	12.710	52.86
11	RP	285	286									285	2	0.2	0.08	0.2	0.05	0.11	12.710	0.69
12	RP	263	275									269	2	8.3	3.10	5.9	2.19	4.38	12.710	27.85
13	RP	516										516	1							
14	RP	368										368	1							
15	RP	297	296									296	2	0.4	0.13	0.3	0.09	0.18	12.710	1.17

Table 4.21: Summary Statistics for L and D Phenylalanine Concentration Data (pM)

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL						
		L-Phe Conc		a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	4112	4246	4339	4316	4434	4389	4407	4382	4314	4345		4328	10	93.3	2.16	29.5	0.68	1.36	2.262	1.54	
2	RP																					
3	RP																					
4	IE																					
5	IE																					
6.1	GC																					
6.2	GC																					
7.1	GC																					
7.2	GC																					
8	RP	3781	3825											3803	2	31.7	0.83	22.4	0.59	1.18	12.710	7.50
9	RP	3616	3674											3645	2	41.2	1.13	29.1	0.80	1.60	12.710	10.16
10	RP	4027	3847											3937	2	127.7	3.24	90.3	2.29	4.59	12.710	29.15
11	RP	2612	2486											2549	2	89.5	3.51	63.3	2.48	4.96	12.710	31.54
12	RP	5707	6063											5885	2	252.2	4.29	178.3	3.03	6.06	12.710	38.51
13	RP	5911												5911	1							
14	RP	6807												6807	1							
15	RP	4870	5134											5002	2	186.3	3.72	131.7	2.63	5.27	12.710	33.47
D-Phe Conc		a	b	c	d	e	f	g	h	i	j			mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	207	207	210	210	218	221	222	231	227	219		217	10	8.4	3.85	2.6	1.22	2.43	2.262	2.75	
2	RP																					
3	RP																					
4	IE																					
5	IE																					
6.1	GC																					
6.2	GC																					
7.1	GC																					
7.2	GC																					
8	RP	193	195											194	2	1.2	0.61	0.8	0.43	0.86	12.710	5.45
9	RP	216	218											217	2	1.6	0.74	1.1	0.52	1.05	12.710	6.65
10	RP	231	225											228	2	4.1	1.81	2.9	1.28	2.55	12.710	16.24
11	RP	148	139											144	2	6.2	4.28	4.4	3.02	6.05	12.710	38.44
12	RP	314	343											329	2	20.1	6.11	14.2	4.32	8.64	12.710	54.89
13	RP	339												339	1							
14	RP	366												366	1							
15	RP	273	286											279	2	9.4	3.36	6.6	2.38	4.75	12.710	30.19

Table 4.22: Summary Statistics for L and D Phenylalanine D/L Ratio Value

Lab No	method	Submitted Replicate data										Standard Deviation			Uncertainty of Mean & Expanded U at 95% CL					
		D/L Phe	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)
1	RP	0.050	0.049	0.048	0.049	0.049	0.050	0.050	0.053	0.053		0.050	9	0.0016	3.21	0.0005	1.07	2.14	2.306	2.47
2	RP	0.047	0.046									0.046	2	0.0007	1.50	0.0005	1.06	2.12	12.710	13.46
3	RP	0.049										0.049	1							
4	IE																			
5	IE																			
6.1 ¹	GC _A	0.051										0.051	8	0.0130	25.49	0.0046	9.01	18.02	2.365	21.31
6.2	GC																			
7.1	GC																			
7.2	GC																			
8	RP	0.051	0.051									0.051	2	0.0000	0.00	0.0000	0.00	0.00	12.710	0.00
9	RP	0.060	0.059									0.060	2	0.0002	0.39	0.0002	0.28	0.55	12.710	3.51
10	RP	0.057	0.058									0.058	2	0.0008	1.44	0.0006	1.02	2.03	12.710	12.92
11	RP	0.057	0.056									0.056	2	0.0004	0.77	0.0003	0.54	1.09	12.710	6.90
12	RP	0.055	0.057									0.056	2	0.0010	1.82	0.0007	1.29	2.58	12.710	16.39
13	RP	0.057										0.057	1							
14	RP	0.054										0.054	1							
15	RP	0.056	0.056									0.056	2	0.0002	0.37	0.0001	0.26	0.52	12.710	3.29

¹= submitted as the mean and standard deviation of n results.

GC_A = derived using peak area

Figure 4.20: Distribution of D/L Values submitted for Phenylalanine




Figure 4.21: Experimental Expanded Uncertainty ($k=2$) of the Mean D/L value for Phenylalanine (value of n displayed).




Figure 4.22: Experimental Expanded Uncertainty ($k=t_{0.05,df}$) of the Mean D/L value for Phenylalanine (value of n displayed).




Table 4.23: Summary Statistics for D-Alloisoleucine/L-Isoleucine Peak Area Data

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL				
		L-Ile peak area*	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)
1	RP	19733	19253	25645	26354	26990	27461	27735	28449	29117	115716	34645	10	28690.8	82.81	9072.8	26.19	52.38	2.262	59.24
2	RP	11914	11806									11860	2	76.6	0.65	54.1	0.46	0.91	12.710	5.80
3	RP	13356										13356	1							
4	IE*	4.729	4.710									4.720	2	0.013	0.28	0.0	0.20	0.40	12.710	2.56
5	IE*	4.827	4.886									4.857	2	0.042	0.86	0.030	0.61	1.21	12.710	7.72
6.1	GC																			
6.2	GC																			
7.1	GC																			
7.2	GC																			
8	RP	35296	36074									35685	2	550.0	1.54	388.9	1.09	2.18	12.710	13.85
9	RP	36131	37255									36693	2	794.7	2.17	561.9	1.53	3.06	12.710	19.46
10	RP	20312	22419									21365	2	1490.1	6.97	1053.6	4.93	9.86	12.710	62.68
11	RP	7062	7153									7108	2	64.4	0.91	45.5	0.64	1.28	12.710	8.14
12	RP	6641	6792									6716	2	106.5	1.59	75.3	1.12	2.24	12.710	14.25
13	RP	13043										13043	1							
14	RP	9263										9263	1							
15	RP	7374	7321									7347	2	37.5	0.51	26.5	0.36	0.72	12.710	4.58
D-Aile peak area*	a	b	c	d	e	f	g	h	i	j		mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	359	351	457	468	488	503	522	539	547	2296	653	10	581.2	88.99	183.8	28.14	56.28	2.262	63.66
2	RP	400	383									391	2	11.6	2.97	8.2	2.10	4.21	12.710	26.73
3	RP	517										517	1							
4	IE*	0.115	0.114									0.115	2	0.0	0.62	0.0	0.44	0.87	12.710	5.55
5	IE*	0.116	0.117									0.117	2	0.0	0.61	0.0	0.43	0.86	12.710	5.45
6.1	GC																			
6.2	GC																			
7.1	GC																			
7.2	GC																			
8	RP	676	687									681	2	7.7	1.13	5.5	0.80	1.60	12.710	10.19
9	RP	1241	1325									1283	2	59.6	4.65	42.1	3.29	6.57	12.710	41.75
10	RP	586	756									671	2	120.3	17.94	85.1	12.68	25.37	12.710	161.22
11	RP	213	214									213	2	0.8	0.37	0.6	0.26	0.53	12.710	3.36
12	RP	213	228									220	2	10.9	4.94	7.7	3.49	6.99	12.710	44.42
13	RP	414										414	1							
14	RP	286										286	1							
15	RP	216	227									221	2	7.9	3.59	5.6	2.54	5.08	12.710	32.28

* = peak height data

Table 4.24: Summary Statistics for D-Alloisoleucine/L-Isoleucine Concentration Data (pM)

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL						
		L-Ile Conc		a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	5362	5197	5497	5594	5757	5731	5779	5742	5703	5583		5595	10	192.7	3.44	60.9	1.09	2.18	2.262	2.46	
2	RP																					
3	RP																					
4	IE																					
5	IE																					
6.1	GC																					
6.2	GC																					
7.1	GC																					
7.2	GC																					
8	RP	5054	5096											5075	2	29.9	0.59	21.2	0.42	0.83	12.710	5.30
9	RP	4862	4973											4918	2	78.7	1.60	55.6	1.13	2.26	12.710	14.38
10	RP	5231	5178											5204	2	37.1	0.71	26.2	0.50	1.01	12.710	6.40
11	RP	3489	3323											3406	2	117.5	3.45	83.1	2.44	4.88	12.710	31.00
12	RP	7541	8048											7795	2	358.2	4.60	253.3	3.25	6.50	12.710	41.30
13	RP	8157												8157	1							
14	RP	8760												8760	1							
15	RP	6440	6718											6579	2	196.0	2.98	138.6	2.11	4.21	12.710	26.78
D-Aile Conc		a	b	c	d	e	f	g	h	i	j			mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	98	95	98	99	104	105	109	109	107	111			103	10	5.6	5.40	1.8	1.71	3.42	2.262	3.87
2	RP																					
3	RP																					
4	IE																					
5	IE																					
6.1	GC																					
6.2	GC																					
7.1	GC																					
7.2	GC																					
8	RP	97	97											97	2	0.2	0.18	0.1	0.13	0.26	12.710	1.64
9	RP	167	177											172	2	7.0	4.08	5.0	2.89	5.77	12.710	36.67
10	RP	151	175											163	2	16.8	10.33	11.9	7.30	14.60	12.710	92.80
11	RP	105	99											102	2	4.1	3.98	2.9	2.81	5.63	12.710	35.77
12	RP	241	270											256	2	20.3	7.95	14.4	5.62	11.24	12.710	71.43
13	RP	259												259	1							
14	RP	271												271	1							
15	RP	188	208											198	2	14.0	7.08	9.9	5.00	10.01	12.710	63.60

Table 4.25: Summary Statistics for D-Alloisoleucine/L-Isoleucine D/L Ratio Value

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL				
		D/L Aile/Ile	a	b	c	d	e	f	g	h	i	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	0.018	0.018	0.018	0.018	0.018	0.018	0.019	0.019	0.019	0.020	0.018	10	0.0006	3.38	0.0002	1.07	2.14	2.262	2.42
2	RP	0.034	0.032									0.033	2	0.0008	2.33	0.0005	1.65	3.29	12.710	20.92
3	RP	0.039										0.039	1							
4	IE	0.024	0.024									0.024	2	0.0000	0.00	0.0000	0.00	0.00	12.710	0.00
5	IE	0.024	0.024									0.024	2	0.0000	0.00	0.0000	0.00	0.00	12.710	0.00
6.1 ¹	GC _A	0.026										0.026	8	0.0010	3.85	0.0004	1.36	2.72	2.365	3.22
6.2 ¹	GC _H	0.030										0.030	5	0.0010	3.33	0.0004	1.49	2.98	2.777	4.14
7.1 ¹	GC _A	0.025										0.025	1							
7.2 ¹	GC _H	0.028										0.028	1							
8	RP	0.019	0.019									0.019	2	0.0000	0.00	0.0000	0.00	0.00	12.710	0.00
9	RP	0.034	0.036									0.035	2	0.0009	2.48	0.0006	1.75	3.51	12.710	22.30
10	RP	0.029	0.034									0.031	2	0.0034	11.03	0.0024	7.80	15.60	12.710	99.16
11	RP	0.030	0.030									0.030	2	0.0002	0.53	0.0001	0.38	0.75	12.710	4.78
12	RP	0.032	0.034									0.033	2	0.0011	3.36	0.0008	2.37	4.75	12.710	30.18
13	RP	0.032										0.032	1							
14	RP	0.031										0.031	1							
15	RP	0.029	0.031									0.030	2	0.0012	4.10	0.0009	2.90	5.80	12.710	36.86

¹= submitted as the mean and standard deviation of n results.

GC_A = derived using peak area

GC_H = derived using peak height

Figure 4.23: Distribution of D/L Values submitted for D-Alloisoleucine/L-Isoleucine




Figure 4.24: Experimental Expanded Uncertainty ($k=2$) of the Mean D/L value for D-Alloisoleucine/L-Isoleucine (value of n displayed).




Figure 4.25: Experimental Expanded Uncertainty ($k=t_{0.05,df}$) of the Mean D/L value for D-Alloisoleucine/L-Isoleucine (value of n displayed).




Table 4.26: Summary Statistics for L and D Leucine Peak Area Data

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL						
		L-Leu peak area		a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	22349	22803	29656	30150	30689	31043	31280	32072	32594	130845		39348	10	32353.2	82.22	10231.0	26.00	52.00	2.262	58.82	
2	RP																					
3	RP																					
4	IE																					
5	IE																					
6.1	GC																					
6.2	GC																					
7.1	GC																					
7.2	GC																					
8	RP	37876	38607											38242	2	516.5	1.35	365.2	0.96	1.91	12.710	12.14
9	RP	37403	38412											37908	2	713.6	1.88	504.6	1.33	2.66	12.710	16.92
10	RP	21435	23341											22388	2	1347.7	6.02	953.0	4.26	8.51	12.710	54.10
11	RP	7348	7449											7399	2	71.1	0.96	50.3	0.68	1.36	12.710	8.63
12	RP	6868	7082											6975	2	151.6	2.17	107.2	1.54	3.07	12.710	19.54
13	RP	13325												13325	1							
14	RP	9775												9775	1							
15	RP	7726	7722											7724	2	2.8	0.04	1.9	0.03	0.05	12.710	0.32
D-Leu peak area		a	b	c	d	e	f	g	h	i	j			mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	986	993	1213	1489	1514	1527	1584	1670	1722	6094		1879	10	1503.9	80.02	475.6	25.30	50.61	2.262	57.24	
2	RP																					
3	RP																					
4	IE																					
5	IE																					
6.1	GC																					
6.2	GC																					
7.1	GC																					
7.2	GC																					
8	RP	1661	1701											1681	2	27.7	1.65	19.6	1.17	2.33	12.710	14.83
9	RP	1923	2033											1978	2	78.2	3.95	55.3	2.80	5.59	12.710	35.54
10	RP	1020	1706											1363	2	485.1	35.59	343.0	25.16	50.33	12.710	319.83
11	RP	393	393											393	2	0.7	0.17	0.5	0.12	0.24	12.710	1.52
12	RP	350	361											355	2	7.6	2.14	5.4	1.51	3.02	12.710	19.20
13	RP																					
14	RP	525												525	1							
15	RP	367	365											366	2	1.7	0.45	1.2	0.32	0.64	12.710	4.09
D+L Leu peak height		a	b	c	d	e	f	g	h	i	j		mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})	
004	IE	5.575	5.545											5.560	2	0.0212	0.38	0.0150	0.27	0.54	12.710	3.43
005	IE	5.802	5.801											5.802	2	0.0007	0.01	0.0005	0.01	0.02	12.710	0.11

Table 4.27: Summary Statistics for L and D Leucine Concentration Data (pM)

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL					
		L-Leu Conc	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	6072	6156	6357	6400	6546	6479	6517	6473	6384	6313	6370	10	154.3	2.42	48.8	0.77	1.53	2.262	1.73	
2	RP																				
3	RP																				
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP	5423	5454										5439	2	21.7	0.40	15.4	0.28	0.56	12.710	3.59
9	RP	6747	6873										6810	2	89.7	1.32	63.4	0.93	1.86	12.710	11.83
10	RP	7399	7226										7313	2	122.0	1.67	86.3	1.18	2.36	12.710	14.99
11	RP	4866	4638										4752	2	161.3	3.39	114.1	2.40	4.80	12.710	30.50
12	RP	10454	11249										10851	2	562.5	5.18	397.7	3.67	7.33	12.710	46.59
13	RP	11170											11170	1							
14	RP	12391											12391	1							
15	RP	9045	9498										9271	2	320.1	3.45	226.4	2.44	4.88	12.710	31.03
D-Leu Conc		a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})	
1	RP	268	268	260	316	323	319	330	337	337	294	305	10	30.2	9.90	9.6	3.13	6.26	2.262	7.08	
2	RP																				
3	RP																				
4	IE																				
5	IE																				
6.1	GC												239	2	1.7	0.70	1.2	0.49	0.99	12.710	6.28
6.2	GC												355	2	12.0	3.39	8.5	2.40	4.79	12.710	30.46
7.1	GC												440	2	124.5	28.29	88.1	20.00	40.01	12.710	254.25
7.2	GC												252	2	11.4	4.52	8.1	3.20	6.40	12.710	40.65
8	RP	238	240										553	2	28.4	5.15	20.1	3.64	7.28	12.710	46.25
9	RP	347	364										665	1							
10	RP	352	528										440	2	13.3	3.03	9.4	2.15	4.29	12.710	27.27
11	RP	261	244																		
12	RP	533	573																		
13	RP																				
14	RP	665																			
15	RP	430	449																		

Table 4.28: Summary Statistics for L and D Leucine D/L Ratio Value

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL				
		D/L Leu	a	b	c	d	e	f	g	h	i	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	0.044	0.044	0.041	0.049	0.049	0.049	0.051	0.052	0.053	0.047	0.048	10	0.0039	8.22	0.0012	2.60	5.20	2.262	5.88
2	RP																			
3	RP																			
4	IE																			
5	IE																			
6.1 ¹	GC _A	0.036										0.036	8	0.0020	5.56	0.0007	1.96	3.93	2.365	4.64
6.2 ¹	GC _H	0.048										0.048	5	0.0010	2.08	0.0004	0.93	1.86	2.777	2.59
7.1 ¹	GC _A	0.058										0.058	1							
7.2 ¹	GC _H	0.052										0.052	1							
8	RP	0.044	0.044									0.044	2	0.0000	0.00	0.0000	0.00	0.00	12.710	0.00
9	RP	0.051	0.053									0.052	2	0.0011	2.07	0.0008	1.47	2.93	12.710	18.63
10	RP	0.048	0.073									0.060	2	0.0180	29.89	0.0128	21.13	42.27	12.710	268.61
11	RP	0.054	0.053									0.053	2	0.0006	1.13	0.0004	0.80	1.60	12.710	10.15
12	RP	0.051	0.051									0.051	2	0.0000	0.04	0.0000	0.03	0.05	12.710	0.34
13	RP																			
14	RP	0.054										0.054	1							
15	RP	0.048	0.047									0.047	2	0.0002	0.42	0.0001	0.30	0.59	12.710	3.77

¹= submitted as the mean and standard deviation of n results.

GC_A = derived using peak area

GC_H = derived using peak height

Figure 4.26: Distribution of D/L Values submitted for Leucine




Figure 4.27: Experimental Expanded Uncertainty ($k=2$) of the Mean D/L value for Leucine (value of n displayed).




Figure 4.28: Experimental Expanded Uncertainty ($k=t_{0.05,df}$) of the Mean D/L value for Leucine (value of n displayed).




Table 4.29: Summary Statistics for L and D Tyrosine Peak Area Data

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL					
		L-Tyr peak area	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP																				
2	RP																				
3	RP																				
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP																				
9	RP	20183	20735										20459	2	389.7	1.90	275.5	1.35	2.69	12.710	17.12
10	RP	12122	13142										12632	2	721.6	5.71	510.2	4.04	8.08	12.710	51.34
11	RP	4217	4262										4240	2	32.0	0.75	22.6	0.53	1.07	12.710	6.77
12	RP	3496	3513										3504	2	11.8	0.34	8.4	0.24	0.48	12.710	3.03
13	RP	7730											7730	1							
14	RP	5529											5529	1							
15	RP	4304	4338										4321	2	24.1	0.56	17.0	0.39	0.79	12.710	5.01
D-Tyr peak area		a	b	c	d	e	f	g	h	i	j		mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP																				
2	RP																				
3	RP																				
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP																				
9	RP	1293	1320										1307	2	18.9	1.44	13.4	1.02	2.04	12.710	12.99
10	RP	728	783										755	2	39.0	5.16	27.6	3.65	7.30	12.710	46.40
11	RP	274											274	1							
12	RP	219	200										210	2	13.4	6.40	9.5	4.53	9.05	12.710	57.54
13	RP	422											422	1							
14	RP	308											308	1							
15	RP	231	235										233	2	2.7	1.15	1.9	0.81	1.63	12.710	10.33

Table 4.30: Summary Statistics for L and D Tyrosine Concentration Data (pM)

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL					
		L-Tyr Conc	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP																				
2	RP																				
3	RP																				
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP																				
9	RP	2348	2393										2371	2	31.7	1.34	22.4	0.95	1.89	12.710	12.03
10	RP	2699	2624										2662	2	52.6	1.98	37.2	1.40	2.80	12.710	17.77
11	RP	1801	1712										1756	2	63.2	3.60	44.7	2.55	5.09	12.710	32.36
12	RP	3432	3599										3516	2	117.7	3.35	83.2	2.37	4.73	12.710	30.09
13	RP	4180											4180	1							
14	RP	4521											4521	1							
15	RP	3250	3442										3346	2	135.4	4.05	95.7	2.86	5.72	12.710	36.37
D-Tyr Conc		a	b	c	d	e	f	g	h	i	j		mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP																				
2	RP																				
3	RP																				
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP																				
9	RP	150	152										151	2	1.3	0.88	0.9	0.62	1.24	12.710	7.90
10	RP	162	156										159	2	4.0	2.53	2.8	1.79	3.57	12.710	22.71
11	RP	117											59	2	82.8	141.42	58.6	100.00	200.00	12.710	1271.00
12	RP	215	205										210	2	7.1	3.39	5.0	2.40	4.80	12.710	30.51
13	RP	228											228	1							
14	RP	252											252	1							
15	RP	175	186										180	2	8.4	4.64	5.9	3.28	6.56	12.710	41.68

Table 4.31: Summary Statistics for L and D Tyrosine D/L Ratio Value

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL					
		D/L Tyr	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP																				
2	RP																				
3	RP																				
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP																				
9	RP	0.064	0.064										0.064	2	0.0003	0.46	0.0002	0.33	0.65	12.710	4.13
10	RP	0.060	0.060										0.060	2	0.0003	0.55	0.0002	0.39	0.78	12.710	4.95
11	RP	0.065											0.065	1							
12	RP	0.063	0.057										0.060	2	0.0040	6.74	0.0029	4.77	9.53	12.710	60.57
13	RP	0.055											0.055	1							
14	RP	0.056											0.056	1							
15	RP	0.054	0.054										0.054	2	0.0003	0.59	0.0002	0.42	0.84	12.710	5.32

Figure 4.29: Distribution of D/L Values submitted for Tyrosine




Figure 4.30: Experimental Expanded Uncertainty ($k=2$) of the Mean D/L value for Tyrosine (value of n displayed).




Figure 4.31: Experimental Expanded Uncertainty ($k=t_{0.05,df}$) of the Mean D/L value for Tyrosine (value of n displayed).




Table 4.32: Summary Statistics for L and D Methionine Peak Area Data

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL					
		L-Met peak area	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP																				
2	RP																				
3	RP																				
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP																				
9	RP	1205	1189										1197	2	11.2	0.94	7.9	0.66	100.9	1205	1189
10	RP	418	411										415	2	5.1	1.23	3.6	0.87	45.7	418	411
11	RP	263											263	1						263	
12	RP	170	158										164	2	8.0	4.86	5.6	3.44	71.7	170	158
13	RP	794											794							794	
14	RP	514											514							514	
15	RP	132	128										130							132	128
D-Met peak area		a	b	c	d	e	f	g	h	i	j		mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP																				
2	RP																				
3	RP																				
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP																				
9	RP	522	449										486	2	51.6	10.62	36.5	7.51	463.4	522	449
10	RP	293	276										284	2	11.7	4.13	8.3	2.92	105.5	293	276
11	RP																				
12	RP																				
13	RP	47											47	1						47	
14	RP	89											89							89	
15	RP	43	39										41							43	39

Table 4.33: Summary Statistics for HPLC Internal Standards; Peak Area/Height Data

Lab No	method	Submitted Replicate data										Standard Deviation				Uncertainty of Mean & Expanded U at 95% CL					
		L-homoArginine peak area	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP	1104	1111	1399	1413	1407	1437	1440	1486	1532	6218	1855	10	1540.1	83.03	487.0	26.26	52.51	2.262	59.40	
2	RP	545	536									540	2	6.0	1.12	4.3	0.79	1.58	12.710	10.04	
3	RP	343										343	1								
4	IE																				
5	IE																				
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP																				
9	RP	2548	2568									2558	2	14.5	0.57	10.2	0.40	0.80	12.710	5.08	
10	RP	1331	1484									1408	2	108.2	7.68	76.5	5.43	10.87	12.710	69.06	
11	RP	867	923									895	2	39.0	4.35	27.6	3.08	6.16	12.710	39.13	
12	RP	1510	1447									1478	2	44.5	3.01	31.5	2.13	4.26	12.710	27.06	
13	RP	1371										1371	1								
14	RP	908										908	1								
15	RP	981	934									958	2	33.4	3.49	23.6	2.47	4.93	12.710	31.35	
Norleucine peak height		a	a	b	c	d	e	f	g	h	i	j	mean	n	std dev	CV%	std u	RSU%	Exp U% (k=2)	t critical (0.05,df)	Exp U% (k=t _{crit})
1	RP																				
2	RP																				
3	RP																				
4	IE	0.088	0.089									0.089	2	0.080	0.80	0.001	0.56	1.13	12.710	7.18	
5	IE	0.078	0.074									0.076	2	0.0	3.72	0.002	2.63	5.26	12.710	33.45	
6.1	GC																				
6.2	GC																				
7.1	GC																				
7.2	GC																				
8	RP																				
9	RP																				
10	RP																				
11	RP																				
12	RP																				
13	RP																				
14	RP																				
15	RP																				

5 STATISTICAL EVALUATION; *Accuracy & Performance Analysis*

5.1 Background to understanding Performance Evaluation

The purpose of this evaluation is to provide a clear and independent statistical evaluation and comparison of participants' results. In routine analysis a laboratory's evaluation of analytical competence is often restricted to intra-laboratory precision evaluation of repeated analyses or the evaluation of bias using certified reference materials (CRM's). However, in the absence of a suitable, matrix matched CRM with a known value and uncertainty, evaluation of method and/or laboratory bias can be impossible without the cooperation of additional laboratories. Estimations of precision may be excellent when taken in isolation, but may give rise to unrealistically small uncertainties.

5.1.1 z-Scores

Participation in a proficiency test provides the opportunity to evaluate analytical bias by comparing an individual laboratory's result against the assigned value for the test material. Performance is traditionally determined by the calculation of a z-score, calculated using the submitted result, a reference or assigned value and the target value for standard deviation, using a procedure recommended in the IUPAC/ISO/AOAC International Harmonised Protocol for the Proficiency Testing of (Chemical) Analytical Laboratories (Thompson et al., 2006), such that;

$$z = \frac{(\bar{x} - \hat{X})}{\sigma_p}$$

where \bar{x} = the mean of participant's reported replicate results (or simply x for a single reported result)

\hat{X} = the assigned value,

and σ_p = the target standard deviation.

Note that; $(x - \hat{X})$ is the calculation for bias.

Satisfactory performance is indicated by achieving a z-score no greater than 2, i.e.; $|z| \leq 2$.

The results of a typical chemical analysis will be normally distributed about the mean with a known standard deviation. Approximately 95% of data will be expected to lie within 2 standard deviations either side of the mean and 99.7% within ± 3 standard deviations. Thus, it is considered 'satisfactory' if a participant's z-score lies within this range. It follows that if a participant's z-score lies outside $|z| > 2$ there is about a 1 in 20 chance that their result is in fact an acceptable result from the extreme of the distribution. If a participant's z-score lies outside $|z| > 3$ the chance that their result is actually acceptable is only about 1 in 300 (Thompson et al., 2006, ISO 13528, 2005).

5.1.2 The Target Standard Deviation; σ_p

The target standard deviation (σ_p) describes how the data is expected to perform for a given analyte and / or test material and determines the limits of satisfactory performance.

These values are often obtained from collaborative trials as the reproducibility standard deviation ($RSD_R\%$), which describes best practice for a specified method for a given matrix/analyte/concentration (Thompson et al., 2006).

$$\sigma_p = \frac{RSD_R}{100} \times c$$

where RSD_R = Relative Standard Deviation of Reproducibility from collaborative trial data, expressed as %

and c = concentration, i.e. the assigned value, \hat{X} , expressed in relevant units.

In the absence of collaborative trial data, the Horwitz equation (Horwitz et al., 1980, Horwitz, 1982, RSC Analytical Methods Committee, 2004) is widely accepted as a suitable predictive measure for the target standard deviation in chemical analysis. However, the Horwitz function is not necessarily suited to every type of chemical analysis and in the absence of a suitable alternative, the use of perception or fitness-for-purpose criteria may need to be employed, taking into consideration any uncertainty in homogeneity of test materials.

The distribution of submitted results and uncertainty of the assigned value ($u(\hat{X})$) (see section 5.3.1) should be small by comparison to the target standard deviation, (σ_p). This ensures that the data are sufficiently tight to give a measure of confidence in the assigned value, (\hat{X}), and that the target value is not overly restrictive.

As a general rule, it can be assumed that participants will be hoping to achieve a satisfactory performance and achieve fitness-for-purpose. It is therefore not an unreasonable expectation that the distribution of submitted results (i.e.; the standard deviation of the assigned value, $\hat{\sigma}$), should be close to the limits of satisfactory performance, σ_p , such that $\hat{\sigma} \approx \sigma_p$. The International Harmonized Protocol (2006) states that if $\hat{\sigma} > 1.2\sigma_p$ then *“laboratories are having difficulty achieving the required reproducibility precision in results from a single population, or that two or more discrepant populations may be represented in the result”*.

A further comment is made in the International Harmonised Protocol concerning the uncertainty of the assigned value to ensure it is sufficiently small so as not to overly influence the calculation of z-scores. It is recommended that $u(\hat{X})^2 \leq 0.1\sigma_p^2$ which approximates to $u(\hat{X}) \leq 0.3\sigma_p$ as also recommended in ISO 13528 (2005). (Note; The exact value chosen represents the appropriate order of magnitude although the exact value is to some extent discretionary).

5.2 In the absence of Fitness-for-Purpose Criteria

To date, there has not been an inter-laboratory collaborative trial carried out according to international guidelines (AOAC, 2000, Horwitz, 1995) to determine single method precision parameters for amino acid racemization analysis on fossil material. The Horwitz equation requires the measurement units to be expressed as a mass fraction, i.e.; mg/Kg = 10^{-6} , which is not appropriate in the current study as D/L results are expressed as a ratio and are thus dimensionless. Therefore, in the absence of an external value for target standard deviation, it was necessary to use perception using fitness-for-purpose criteria.

The target value chosen during homogeneity evaluation, (σ_h) is an excellent indication of the observed variation within test materials and reflects the uncertainty due to matrix plus the analytical method used for their determination. The relative value of σ_h expressed as a percentage; i.e.; the RSD%, is a more useful value and can be used to set the minimum permissible value for σ_p . Whilst an inter-laboratory collaborative trial reproducibility standard deviation (RSD_R%) would also reflect an additional laboratory component of variation, in the absence of such data, it none the less makes a good starting point for evaluating submitted results and provides a minimum fitness-for-purpose target value.

During the statistical evaluation of data, it was observed that for some amino acids in some test materials provided in this series of studies, the homogeneity target value was too wide compared to the submitted data for the test, suggesting that the **precision between different laboratories in some instances was better than that observed between samples analysed by a single laboratory under repeatability conditions for homogeneity!**

5.2.1 *Relative percentage bias*

Whilst these observations were surprising, it posed some difficulties in using objective fitness for purpose criteria for the determination of the target values for standard deviation.

In order to overcome this problem and in order to ensure consistency between test materials, in the absence of independently determined performance criteria it was decided to present the data as an assessment of relative bias (%), such that;

$$\text{Relative bias \%} = \frac{(x - \hat{X})}{\hat{X}} \times 100$$

Satisfactory performance was assessed as plus or minus twice the standard deviation of the assigned value, representing 95% confidence limits, i.e.; $\pm 2\hat{\sigma}$.

In this way it was possible to represent participant's results graphically as histograms in a similar way to z-score charts, with the 2 std deviation satisfactory range being given as percentage values rather than ± 2 .

When calculating z-scores, the use of a standard deviation, σ_p , as the denominator acts to normalize results. This enables performance between different analytes or between different test materials to be compared on a common scale, but requires the target value (σ_p) to be scaled appropriately to the individual analyte or matrix. However, using the assigned value (\hat{X}) as the denominator, and calculating the relative percentage bias, still permits a comparison between analytes and test materials but on a common percentage scale, thus providing perhaps a slightly more intuitive presentation of observed bias for individual results.

Laboratory results were calculated from the mean of submitted replicate data so as not to dominate and unfairly influence the distribution by a single method, analyst or single test material. The distributions of the mean values are presented as dot plots in Figure 5.1. On this occasion, performance has not been determined by the calculation of z-scores but rather an evaluation of bias has been carried out. Laboratory mean values and relative percentage bias for each amino acid are given in Table 5.1. and shown as histograms in Figures 5.2 – 5.18.

5.3 The Assigned Value, \hat{X}

The reference or assigned value, \hat{X} , is the best estimate of the true concentration of each analyte. Depending on the nature of a test material, this can be done in a number of different ways, for example the use of a reference value from a Certified Reference Material, a consensus of expert laboratories, or the consensus of submitted results.

In determining the assigned value for a specific analyte, the robust mean is often used as the best estimate in a large data set as it minimises the effect of outliers and gives a fairer estimate of central tendency. However, for small data sets such as here, whilst the robust mean may still be preferable to the standard mean, the influence of extreme values may still be significant. In such instances, the use of the median may be more suitable or even the mode.

5.3.1 *The uncertainty of the Assigned value $u(\hat{X})$.*

When determining the appropriate measure of central tendency, the effect of the uncertainty of the assigned value ($u(\hat{X})$) on performance assessment also needs to be given consideration. If there is too much uncertainty associated with the assigned value, i.e.; either m is too small or the distribution of results is too large, then this can have an adverse impact by exaggerating observed bias. For the robust mean and median:

$$u(\hat{X}) = \frac{\hat{\sigma}}{\sqrt{m}}$$

Where m = the number of laboratory results used to calculate the robust mean or median

and $\hat{\sigma}$ = the standard deviation of the robust mean or median absolute deviation (sMAD). (Note this is not the same as the target standard deviation used for calculating z-scores (σ_p)).

For the mode, $u(\hat{X})$ is taken to be directly equivalent to the standard error of the mode, (SEM).

5.4 Derivation of \hat{X} for Amino Acids in Ostrich Egg Shell (B) Test Material

In this study all assigned values have been determined as the consensus of submitted data, which due to the low numbers of participants involved, equates to the consensus from expert laboratories!

Whilst assessing the data, in many cases it became clear that the robust mean (Ellison, 2002b, RSC Analytical Methods Committee, 1989, RSC Analytical Methods Committee, 2001) was strongly influenced by extreme values resulting in a skewed distribution with a high or low end tail. This appeared largely influenced by method and on occasions by an individual laboratory where more than one result was submitted using the same method, but carried out using a different instrument or analyst. In addition, when determining the mode (Ellison, 2002a, RSC Analytical Methods Committee, 2006, Lowthian and Thompson, 2002), it became clear that due to the low numbers of results, additional modes were identified due to only a couple of values and in some cases only a single data point. Plots showing the modal distributions derived using the kernel density Excel add-in (Ellison, 2002a) are shown against each histogram for amino acids with eight or more data points.

In cases where there were two evenly matched modes or where a smaller second mode was predominated by data using a specific method such as GC, it would not be appropriate to penalise these laboratories by comparison against an assigned value determined from the primary or first mode. There is no judgment being made as to which set of results is 'correct', therefore, it would not be appropriate to calculate performance for GC results using an assigned value determined from HPLC values if the GC data clustered differently. In situations such as this where the method may be empirical, the mode should not be used. Regrettably submitted results by GC were limited making it difficult to know whether the observed differences are genuine method differences or simply extreme values.

For these reasons, the median has been used as the most appropriate measure of central tendency for all amino acids. The median ignores the effect of outliers and assumes a normal distribution placing data symmetrically placed either side of the mid-point. This allows for any asymmetry arising from bimodality to be seen in the histograms but makes no judgment as to the correct mode.

Proficiency tests in principle tend not to be method prescriptive unless methods are known to be empirical and produce different results. The extent of any such differences between GC and HPLC or even between rpHPLC and HPLC-IE for the analysis of amino acid racemization, have not been fully established to date. Therefore, in this proficiency test, GC data have been included with HPLC values and initially evaluated against the same assigned value.

However, where GC data has been provided, for aspartic acid/asparagine, alanine, and valine, GC data can be seen to contribute to high or low end tails. Whilst in this test material GC results for glutamic acid/ glutamine, phenylalanine, alloisoleucine/isoleucine and leucine appear to fall within the general distribution of the data, for consistency with other test materials in this series, rpHPLC results have also been evaluated separately for comparison. Insufficient data prevented a separate evaluation for GC or HPLC-IE methods individually.

The medians used to set the assigned values for all amino acids, together with the number of laboratory results m , the standard deviation of the assigned value, $\hat{\sigma}$ and the standard uncertainty of the assigned value, $u(\hat{X})$, are given in Table 5.2. Table 5.3 then gives the percentage of laboratories with mean values falling within ± 2 standard deviations of the assigned value.

5.5 Interpreting Results - a word of caution.

Caution should be exercised when evaluating the results from this study. Whilst every effort has been made to provide a statistically sound and informative comparison and assessment of data, results from all statistical evaluations should be treated for information only due to the absence of external reference data and the uncertainty surrounding assessment parameters.

The report indicates a number of issues such as the level of agreement between HPLC and GC or even between reverse phase HPLC and ion-exchange HPLC methods, and whether these approaches should be considered empirical, such that the method defines the output. This is suggested from results of a number of amino acids. A greater number of laboratories submitting GC data may have helped to answer this. Determination of method specific assigned values would therefore provide truer estimates of bias and uncertainty and a more accurate performance evaluation.

Obtaining an independent and externally derived precision estimate for the target standard deviation such as the reproducibility standard deviation obtained from a collaborative trial becomes paramount for the future. As an indicator of best practice this would provide guideline uncertainty estimates (so long as a laboratory's repeatability complied with published values), define reference values for the use of any remaining material in place of CRMs enhancing quality control processes, and permit the objective assessment of participants' PT data in future studies.

Table 5.1: Results and Relative Percentage Bias for Total Hydrolysed Amino Acids in OES (B) Test Material

Lab No.	method	Total Hydrolysed Amino Acid (THAA)							
		Asx D/L (all)		Asx D/L (rpHPLC)		Glx D/L (all)		Glx D/L (rpHPLC)	
		assigned value	0.223	assigned value	0.222	assigned value	0.062	assigned value	0.062
		result D/L	relative bias %	result D/L	relative bias %	result D/L	relative bias %	result D/L	relative bias %
1	RP	0.192	-14.0	0.192	-13.6	0.049	-20.1	0.049	-20.4
2	RP	0.198	-11.5	0.198	-11.1	0.044	-28.9	0.044	-29.0
3	RP	0.210	-6.0	0.210	-5.6	0.045	-27.3	0.045	-27.5
4	IE								
5	IE								
6.1	GC	0.225	0.8			0.059	-4.6		
6.2	GC	0.299	33.9						
7.1	GC	0.216	-3.2			0.057	-7.9		
7.2	GC	0.280	25.4						
8	RP	0.217	-2.8	0.217	-2.3	0.060	-3.0	0.060	-3.3
9	RP	0.232	4.0	0.232	4.5	0.067	7.7	0.067	7.4
10	RP	0.226	1.2	0.226	1.7	0.065	4.8	0.065	4.5
11	RP	0.222	-0.5	0.222	0.0	0.064	4.2	0.064	3.9
12	RP	0.228	1.9	0.228	2.4	0.064	4.2	0.064	4.0
13	RP	0.223	0.0	0.223	0.5	0.062	0.0	0.062	-0.3
14	RP	0.225	0.7	0.225	1.2	0.063	1.7	0.063	1.4
15	RP	0.221	-1.0	0.221	-0.6	0.062	0.3	0.062	0.0

Results shown are the average of replicate values where more than one value was given, or as submitted by participants, where a mean value was provided.

Table 5.1: Results and Relative Percentage Bias for Total Hydrolysed Amino Acids in OES (B) Test Material (continued)

Lab No.	method	Total Hydrolysed Amino Acid (THAA)							
		Ser D/L (rpHPLC)		Arg D/L (rpHPLC)		Ala D/L		Ala D/L (rpHPLC)	
		assigned value	0.112	assigned value	0.100	assigned value	0.062	assigned value	0.065
		result D/L	relative bias %	result D/L	relative bias %	result D/L	relative bias %	result D/L	relative bias %
1	RP	0.108	-3.3			0.060	-3.6	0.060	-8.2
2	RP	0.111	-0.2	0.088	-12.0	0.058	-7.0	0.058	-11.4
3	RP	0.116	3.6	0.105	4.8	0.069	10.9	0.069	5.6
4	IE								
5	IE								
6.1	GC					0.047	-24.2		
6.2	GC					0.050	-19.4		
7.1	GC					0.048	-22.6		
7.2	GC					0.047	-24.2		
8	RP	0.117	4.7			0.062	0.0	0.062	-4.7
9	RP	0.114	2.4	0.114	14.7	0.073	18.0	0.073	12.4
10	RP	0.112	-0.2	0.109	9.3	0.069	11.3	0.069	6.0
11	RP	0.109	-2.0	0.097	-2.4	0.065	4.9	0.065	0.0
12	RP	0.111	-0.8	0.100	0.1	0.066	7.0	0.066	1.9
13	RP	0.112	0.0	0.095	-5.1	0.055	-11.4	0.055	-15.6
14	RP	0.115	2.9	0.100	0.0	0.063	1.8	0.063	-3.0
15	RP	0.114	2.4	0.099	-1.1	0.069	10.6	0.069	5.4

Results shown are the average of replicate values where more than one value was given, or as submitted by participants, where a mean value was provided.

Table 5.1: Results and Relative Percentage Bias for Total Hydrolysed Amino Acids in OES (B) Test Material (continued)

Lab No.	method	Total Hydrolysed Amino Acid (THAA)							
		Val D/L		Val D/L (rpHPLC)		Phe D/L		Phe D/L (rpHPLC)	
		assigned value	0.020	assigned value	0.020	assigned value	0.054	assigned value	0.056
		result D/L	relative bias %	result D/L	relative bias %	result D/L	relative bias %	result D/L	relative bias %
1	RP	0.017	-13.6	0.017	-14.8	0.050	-6.7	0.050	-10.2
2	RP	0.021	2.1	0.021	0.6	0.046	-14.1	0.046	-17.3
3	RP	0.021	2.4	0.021	0.9	0.049	-8.5	0.049	-11.9
4	IE								
5	IE								
6.1	GC	0.015	-25.6			0.051	-5.1		
6.2	GC	0.014	-30.5						
7.1	GC	0.012	-40.5						
7.2	GC	0.014	-30.5						
8	RP	0.021	1.7	0.021	0.3	0.051	-5.1	0.051	-8.6
9	RP	0.022	7.8	0.022	6.2	0.060	10.9	0.060	6.8
10	RP	0.020	0.0	0.020	-1.4	0.058	7.7	0.058	3.7
11	RP	0.019	-3.6	0.019	-5.0	0.056	5.0	0.056	1.1
12	RP	0.020	0.3	0.020	-1.1	0.056	3.9	0.056	0.0
13	RP	0.016	-22.4	0.016	-23.5	0.057	6.7	0.057	2.8
14	RP	0.023	15.6	0.023	13.9	0.054	0.0	0.054	-3.7
15	RP	0.020	1.5	0.020	0.0	0.056	3.8	0.056	0.0

Results shown are the average of replicate values where more than one value was given, or as submitted by participants, where a mean value was provided.

Table 5.1: Results and Relative Percentage Bias for Total Hydrolysed Amino Acids in OES (B) Test Material (continued)

Lab No.	method	Total Hydrolysed Amino Acid (THAA)							
		D-Aile/L-Ile (all)		D-Aile/L-Ile (rpHPLC)		Leu D/L (all)		Leu D/L (rpHPLC)	
		assigned value	0.030	assigned value	0.031	assigned value	0.051	assigned value	0.052
		result D/L	relative bias %	result D/L	relative bias %	result D/L	relative bias %	result D/L	relative bias %
1	RP	0.018	-38.4	0.018	-40.9	0.048	-7.0	0.048	-7.1
2	RP	0.033	10.0	0.033	5.5				
3	RP	0.039	29.1	0.039	23.9				
4	IE	0.024	-20.0						
5	IE	0.024	-20.0						
6.1	GC	0.026	-13.3			0.036	-30.1		
6.2	GC	0.030	0.0			0.048	-6.7		
7.1	GC	0.025	-16.7			0.058	12.7		
7.2	GC	0.028	-6.7			0.052	1.0		
8	RP	0.019	-36.7	0.019	-39.2	0.044	-14.5	0.044	-14.6
9	RP	0.035	16.5	0.035	11.8	0.052	1.4	0.052	1.2
10	RP	0.031	4.2	0.031	0.0	0.060	17.3	0.060	17.1
11	RP	0.030	0.0	0.030	-4.1	0.053	3.2	0.053	3.0
12	RP	0.033	9.3	0.033	4.9	0.051	-1.0	0.051	-1.2
13	RP	0.032	5.7	0.032	1.4				
14	RP	0.031	3.0	0.031	-1.2	0.054	4.3	0.054	4.1
15	RP	0.030	0.4	0.030	-3.7	0.047	-7.9	0.047	-8.0

Results shown are the average of replicate values where more than one value was given, or as submitted by participants, where a mean value was provided.

Table 5.1: Results and Relative Percentage Bias for Total Hydrolysed Amino Acids in OES (B) Test Material (continued)

Lab No.	method	Total Hydrolysed Amino Acid (THAA)	
		Tyr D/L (rpHPLC)	assigned value
		result D/L	relative bias %
1	RP		
2	RP		
3	RP		
4	IE		
5	IE		
6.1	GC		
6.2	GC		
7.1	GC		
7.2	GC		
8	RP		
9	RP	0.064	6.8
10	RP	0.060	0.0
11	RP	0.065	8.7
12	RP	0.060	0.0
13	RP	0.055	-8.7
14	RP	0.056	-6.9
15	RP	0.054	-9.8

Results shown are the average of replicate values where more than one value was given, or as submitted by participants, where a mean value was provided.

Table 5.2: Assigned Values, Standard Deviations and Standard Uncertainties

analyte	assigned value					
	m	Median (\hat{X})	sMAD ($\hat{\sigma}$)	RSD %	Std uncertainty of median ($u(\hat{X})$)	RSU %
Asx D/L (all ^a)	15	0.223	0.009	4.14	0.0024	1.07
Asx D/L (rpHPLC)	11	0.222	0.008	3.47	0.0023	1.05
Glx D/L (all ^a)	13	0.062	0.004	6.85	0.0012	1.90
Glx D/L (rpHPLC)	11	0.062	0.004	5.88	0.0011	1.77
Ser D/L (rpHPLC)	11	0.112	0.004	3.57	0.0012	1.08
Arg D/L (rpHPLC)	9	0.100	0.007	7.10	0.0024	2.37
Ala D/L (all ^a)	15	0.062	0.010	16.09	0.0026	4.15
Ala D/L (rpHPLC)	11	0.065	0.005	8.34	0.0016	2.51
Val D/L (all ^a)	15	0.020	0.002	11.49	0.0006	2.97
Val D/L (rpHPLC)	11	0.020	0.0004	2.14	0.0001	0.64
Phe D/L (all ^a)	12	0.054	0.005	9.49	0.0015	2.74
Phe D/L (rpHPLC)	11	0.056	0.003	5.51	0.0009	1.66
D-Aile/L-Ile (all ^b)	17	0.030	0.004	14.82	0.0011	3.59
D-Aile/L-Ile (rpHPLC)	11	0.031	0.002	7.28	0.0007	2.19
Leu D/L (all ^a)	12	0.051	0.005	10.18	0.0015	2.94
Leu D/L (rpHPLC)	8	0.052	0.004	8.36	0.0015	2.96
Tyr D/L (rpHPLC)	7	0.060	0.006	10.20	0.0023	3.86

^a = rpHPLC and GC data^b = rpHPLC, GC and HPLC-IE data

m = number of replicate mean values

sMAD = median absolute deviation

RSD% = Relative standard deviation expressed as a percentage

RSU% = Relative standard uncertainty expressed as a percentage

Table 5.3: Satisfactory Performance(Percentage within 95% Confidence Interval)

analyte	assigned value			
	Median (\hat{X})	Satisfactory m	Total number of m	Percent satisfactory
Asx D/L (all ^a)	0.223	11	15	73%
Asx D/L (rpHPLC)	0.222	9	11	82%
Glx D/L (all ^a)	0.062	10	13	77%
Glx D/L (rpHPLC)	0.062	8	11	73%
Ser D/L (rpHPLC)	0.112	11	11	100%
Arg D/L (rpHPLC)	0.100	8	9	89%
Ala D/L (all ^a)	0.062	15	15	100%
Ala D/L (rpHPLC)	0.065	11	11	100%
Val D/L (all ^a)	0.020	11	15	73%
Val D/L (rpHPLC)	0.020	6	11	55%
Phe D/L (all ^a)	0.054	12	12	100%
Phe D/L (rpHPLC)	0.056	9	11	82%
D-Aile/L-Ile (all ^b)	0.030	15	17	88%
D-Aile/L-Ile (rpHPLC)	0.031	8	11	73%
Leu D/L (all ^a)	0.051	11	12	92%
Leu D/L (rpHPLC)	0.052	7	8	88%
Tyr D/L (rpHPLC)	0.060	7	7	100%

^a = rpHPLC and GC data^b = rpHPLC, GC and HPLC-IE data

m = number of participants' results

Figure 5.1: Distribution of Participants' Average Measurement Values




Figure 5.2: Relative Percentage Bias for Aspartic Acid / Asparagine D/L Results (all data) in Ostrich Egg Shell (B) Test Material





Figure 5.3: Relative Percentage Bias for Aspartic Acid / Asparagine D/L Results (rpHPLC data only) in Ostrich Egg Shell (B) Test Material



0.238 D/L

0.222 D/L

0.207 D/L

Figure 5.4: Relative Percentage Bias for Glutamic Acid / Glutamate D/L Results (all data) in Ostrich Egg Shell (B) Test Material





Figure 5.5: Relative Percentage Bias for **Glutamic Acid / Glutamate D/L Results (rpHPLC data only)** in Ostrich Egg Shell (B) Test Material

0.069 D/L

0.087 D/L

0.055 D/L

Figure 5.6: Relative Percentage Bias for Serine D/L Results (all / rpHPLC data) in Ostrich Egg Shell (B) Test Material




Figure 5.7: Relative Percentage Bias for Arginine D/L Results (rpHPLC data only) in Ostrich Egg Shell (B) Test Material




Figure 5.8: Relative Percentage Bias for Alanine D/L Results (all data) in Ostrich Egg Shell (B) Test Material




Figure 5.9: Relative Percentage Bias for Alanine D/L Results (rpHPLC data only) in Ostrich Egg Shell (B) Test Material





Figure 5.10: Relative Percentage Bias for Valine D/L Results (all data) in Ostrich Egg Shell (B) Test Material



0.025 D/L

0.020 D/L

0.016 D/L

Figure 5.11: Relative Percentage Bias for Valine D/L Results (rpHPLC data only) in Ostrich Egg Shell (B) Test Material




Figure 5.12: Relative Percentage Bias for Phenylalanine D/L Results (all data) in Ostrich Egg Shell (B) Test Material




Figure 5.13: Relative Percentage Bias for Phenylalanine D/L Results (rpHPLC data only) in Ostrich Egg Shell (B) Test Material





Figure 5.14: Relative Percentage Bias for **D-Alloisoleucine/L-Isoleucine Results (all data)** in Ostrich Egg Shell (B) Test Material

Figure 5.15: Relative Percentage Bias for D-Alloisoleucine/L-Isoleucine Results (rpHPLC data only) in Ostrich Egg Shell (B) Test Material




Figure 5.16: Relative Percentage Bias for Leucine D/L Results (all data) in Ostrich Egg Shell (B) Test Material




Figure 5.17: Relative Percentage Bias for Leucine D/L Results (rpHPLC data only) in Ostrich Egg Shell (B) Test Material





Figure 5.18: Relative Percentage Bias for Tyrosine D/L Results (**rpHPLC data only**) in Ostrich Egg Shell (B) Test Material



6 MEASUREMENT UNCERTAINTY

Ostrich Egg Shell (B) Test Material

6.1 Estimation of Measurement Uncertainty from Inter-laboratory comparisons.

Proficiency test data can provide a valuable indication of method and laboratory bias in routine analysis. Bias (*bias*) and its associated uncertainty ($u(bias)$) is often evaluated as part of a laboratory's method validation process by analysis of a certified reference material (CRM) or from spiking experiments. This, together with the determination of internal precision estimates (intra-laboratory reproducibility standard deviation (S_{RW})) can define the overall combined uncertainty for a measurement system (u_c), and is referred to as the 'top-down' approach to measurement uncertainty determination (Barwick and Ellison, 2000).

Where such validation data is available, performance in a proficiency test can provide verification of a laboratory's own uncertainty estimates, which should be compatible with the spread of their PT results over time. However in the absence of such data the result can be used as a direct indication of bias itself, which together with an estimate of precision such as the intra-laboratory reproducibility standard deviation (S_{RW}), can provide a value for the combined uncertainty.

It should be recognised that due to the uncertainty of the assigned value, bias and the uncertainty due to bias associated with a PT, The uncertainty estimate is likely to be larger than that resulting from the analysis of a CRM. It is recommended that long term bias trends are observed to lessen the impact from a single proficiency test result and at least 6 rounds of testing are used to evaluate bias estimates (Magnusson et al., 2004)

In addition, it is recommended that intra-laboratory precision estimates (S_{RW}) are determined from replicate analyses of samples under reproducibility conditions over an extended period of time to take account of between run and general day to day variability. To simply use the standard deviation from replicate results submitted for the proficiency test is not a realistic representation of the overall method and laboratory precision. Alternatively, an estimation of the between laboratory reproducibility standard deviation (S_R) determined using an analysis of variance (ANOVA) on results from a collaborative trial, can be used directly in place of the combined standard uncertainty.

$$\text{Thus; } u_c = \sqrt{S_{RW}^2 + u(bias)^2} \cong S_R$$


It is widely recognised that evaluation of PT data can be a valuable addition to the determination of measurement uncertainty, however there is very little information provided by the main guidance documents (JCGM 100; 2008, EURACHEM / CITAC, 2000) on exactly how this should be done. The following methodology is therefore derived from two main sources; the Nordtest Report TR 537ⁱⁱⁱ (Magnusson et al., 2004) produced as a handbook for the Nordic environmental testing laboratories and Eurolab's Technical reports^{iv} Nos 1/2006 and 1/2007 (EUROLAB, 2006, EUROLAB, 2007). All documents are freely downloadable and recommended for further reading on the subject.

ⁱⁱⁱ <http://www.nordicinnovation.net/nordtestfiler/tec537.pdf>

^{iv} http://www.eurolab.org/pub/i_pub.html

For those readers unfamiliar with measurement uncertainty estimation, distinguishing the various uncertainty components can be somewhat baffling. Below helps to illustrate the sources and relevance of the different contributions due to precision and particularly those elements due to bias. These will now be expanded on in the remainder of this section, together with the calculation of the combined standard uncertainty and expanded uncertainty estimates.

Figure 6.1: Bias and Precision Components to Measurement Uncertainty Estimation.



6.2 Standard uncertainty due to Bias ($u(\text{bias})$).

6.2.1 For a result from a single proficiency test.

The simplest expression for the bias uncertainty ($u(\text{bias})$) is the experimental uncertainty of the laboratory mean $u(\bar{x})$ **plus** the uncertainty of the assigned value $u(\hat{X})$ where $u = s/\sqrt{n}$. Note; if a CRM was used as the test material, $u(\hat{X})$ can be taken from the specifications directly.

$$u(\text{bias}) = \sqrt{u(\bar{x})^2 + u(\hat{X})^2} = \sqrt{\frac{s_{\bar{x}}^2}{n_{\bar{x}}} + \frac{s_{\hat{X}}^2}{m_{\hat{X}}}}$$

Where $s_{\bar{x}}$ = standard deviation of the laboratory's submitted result,
 $n_{\bar{x}}$ = number of laboratory replicates,
 $s_{\hat{X}}$ = standard deviation of the assigned value, and
 $m_{\hat{X}}$ = number of laboratories' results contributing to the assigned value.

In routine analysis, bias should be accounted for and corrected for significant systematic effects. However in circumstances where this is not done by convention and the method is said to be empirical, any significant uncorrected bias should contribute to the combined uncertainty budget.

Bias is determined as ;

$$bias = (\bar{x} - \hat{X}) \quad \text{or as a relative value} \quad \frac{bias}{\hat{X}} = \left(\frac{\bar{x} - \hat{X}}{\hat{X}} \right)$$

Where \bar{x} = laboratory result (or the mean of replicate values)
and \hat{X} = the assigned value.

To determine whether the observed bias is significant or not, the t statistic is calculated and compared to the 2-tailed critical value for $n-1$ degrees of freedom. If t is greater than or equal to the critical value, t_{crit} , then the bias is significant and an additional term to account for uncorrected bias in the result needs to be included in the combined uncertainty estimate (EURACHEM / CITAC, 2000).

t is calculated as;

$t = \frac{1-Rec}{u(Rec)}$ where ; $Rec = \bar{x}/\hat{X}$ and usually represents the recovery associated with the analysis of a CRM and $u(Rec)$ is the same as $u(bias)$ given above.

If $t \geq t_{crit}$, Rec is significantly different from 1 and the result \bar{x} remains uncorrected, a bias correction term needs to be included in the combined uncertainty estimate.

However, this scenario is to some extent academic as the uncertainty of the assigned value in a proficiency test is likely to be much larger than that of a CRM (if one were available) and it is recommended to include the bias contribution in the uncertainty evaluation at all times regardless of whether $t \geq t_{crit}$ or not (Magnusson et al., 2004).

Thus, the bias uncertainty now becomes;

$$u(bias) = \sqrt{(\bar{x} - \hat{X})^2 + \frac{s_{\bar{x}}^2}{n_{\bar{x}}} + \frac{\hat{\sigma}^2}{m_{\hat{X}}}} \quad \text{or} \quad \sqrt{(bias)^2 + u(\bar{x})^2 + u(\hat{X})^2}$$

6.2.2 For results from multiple proficiency tests

When multiple results have been obtained from several proficiency tests then the contribution due to bias and the uncertainty due to bias (i.e.; the experimental uncertainty of the replicate mean $u(\bar{x})$), can be replaced by the bias root mean square (RMS_{bias}), thus;

$$u(bias) = \sqrt{RMS_{bias}^2 + u(\hat{X})^2} \quad \text{where} \quad RMS_{bias} = \sqrt{\sum(bias_i)^2/m}$$

The average standard deviation for the assigned values and the average number of participants across all the tests can be determined and used to calculate an average uncertainty value for the tests.

"The use of an RMS value is equivalent to an estimated standard deviation around an assumed value of bias equal to zero. This implies that the RMS value takes into account both the bias and the variation of bias". (EUROLAB, 2007).

6.3 Combined uncertainty (u_C).

The combined uncertainty is therefore calculated as;

$$u_C = \sqrt{S_{RW}^2 + u(\bar{x})^2 + u(\hat{X})^2 + (bias)^2}$$

Where S_{RW} is the intra-laboratory reproducibility precision estimate.

Note concerning z-scores; for laboratories performing within the satisfactory range, i.e.; $|z|=2$, where there is a normal distribution of z-scores, that is, some may be positive and others negative, there will be no overall bias associated with the laboratory's performance. In this case the uncertainty associated with a

result will be based on the uncertainty of that result, i.e.; $u(\bar{x})$, plus the uncertainty of the assigned value $u(\hat{X})$, plus the precision contribution S_{RW} , which in this case is equivalent to the target standard deviation, σ_p . Where the uncertainty of the assigned value and /or the uncertainty of the result is considered negligible compared to the target standard deviation used for assessment (σ_p), then the uncertainty associated with the laboratory's result is simply equivalent to σ_p , or it's RSD value expressed as a percentage.

6.4 Expanded Uncertainty (U).

The final step in determining the measurement uncertainty is to calculate the Expanded uncertainty U by multiplying the combined uncertainty with a coverage factor k .

$$U = u_c \times k \quad \text{where } k \text{ is the coverage factor set according to the required confidence level.}$$

For a discussion of the appropriate value of k , see Section 4.2.2. However, for a large, normally distributed data set, at a 95% or 2 standard deviation confidence level, $k=2$. For smaller data sets $k=t_{(0.05, df)}$.

A combined uncertainty brings together uncertainty contributions from different sources, therefore determining k becomes a little more tricky as there is no single value for the degrees of freedom. One approach is to calculate an effective degree of freedom using the Welch-Satterthwaite formula where the effective degree of freedom is less than or equal to the sum of the individual values, i.e.; $(v_{eff} \leq \sum v_i)$. The use of this equation is covered in detail in Annex G of the Guide to Uncertainty Measurement or "GUM"; (JCGM 100:; 2008).

$$v_{eff} = u_c^4(y) / \sum \frac{u_i^4(y)}{v_i}$$

Where v_{eff} = the effective degrees of freedom,
 v_i = degrees of freedom of individual uncertainty components,
 u_c = combined standard uncertainty
 u_i = individual uncertainty components.

However, Eurachem make the following recommendation; "*Where the combined standard uncertainty is dominated by a single contribution with fewer than six degrees of freedom, it is recommended that k be set equal to the two-tailed value of the Student's t for the number of degrees of freedom associated with that contribution and for the level of confidence required...*" (EURACHEM / CITAC, 2000).

6.5 Calculating Measurement Uncertainty for Amino Acids in Ostrich Egg Shell (B) Test Material

To illustrate how precision and bias components can be used to provide an estimate of analytical uncertainty, the following evaluations have been carried. The information thus presented should perhaps be considered more as an information exercise than a definitive measure of uncertainty. This is due to a number of reasons; such as the relatively small data set, the "uncertainty" surrounding the empirical nature of the results and the effect on the confidence in the assigned value. Also because of the absence of true intra-laboratory precision estimates and the fact that not all laboratories supplied analytical replicate values. Nonetheless, the data presented in the following tables demonstrates how it can be possible to determine measurement uncertainty using proficiency test data and provides some interesting indicative values.

In all cases, individual laboratory expanded uncertainties (U) have been determined using a coverage factor $k=2$. This is to simplify the calculations whilst considering uncertainty components from various sources but also in order to enable direct comparability between laboratories and across analytes.

Results should be expressed as; result $(\bar{x}) \pm U$ (at 95% confidence, using $k=2$)

6.5.1 Measurement Uncertainty Evaluation for a series of results using RMS_{bias} .

As already mentioned in Section 6.3, for PT results with no overall bias (*bias*), where the uncertainty of the assigned values, $u(\hat{X})$, were negligible and where the uncertainty of replicate values, $u(\bar{x})$ were small compared to intra-laboratory precision estimates S_{RW} , then the standard uncertainty for laboratories within the satisfactory range would be equivalent to the target standard deviation, σ_p .

However, in this report, no values for target standard deviation, σ_p , have been given. Under these circumstances and assuming the absence of bias described above still holds, the uncertainty of laboratories' mean values would be equivalent to each laboratory's own intra-laboratory reproducibility S_{RW} , if this information where known. In the absence of this, the instrumental repeatability (i.e.; the RSD% or CV%) derived from the replicate values might be used, ideally with an additional term included to take into account the expected variability between samples. In the absence of this and to avoid the risk of undervaluing the precision contribution, the reproducibility value derived from all participant's results, given in Table 4.1 at the beginning of the report, might be used as a compromise. This would assume that all laboratories were performing at the stated level of precision and makes no allowance for those that were performing better or worse than this.

Whilst the above scenario may be ideal, in reality it is probably a little unrealistic. It would be far more appropriate to assess the bias components and include them in the uncertainty budget, even if their overall contribution is small, at least until the analyst is confident that analytical results are free from bias.

Table 6.1 demonstrates how this could be carried out using a series of results. In this example we are using results from a number of laboratories in a single round of testing to obtain an average uncertainty for the amino acid in the test material. In practice it is perhaps more likely that a single laboratory would want to assess their own data from a series of proficiency tests carried out. The data shown uses the $RMS_{bias}\%$ (see 6.2.2) determined from all the submitted results by all the laboratories for any given amino acid. From this the average combined and expanded uncertainties for each amino acid for this test material can be derived.

Here the precision estimates used are the standard deviations for the assigned values, $(\hat{\sigma})$, i.e.; sMAD (see Section 5.3). They represent the distributions of the laboratories' means and were used to set the satisfactory limits (i.e.; ± 2 std dev), but are not as influenced as the reproducibility standard deviations (S_R and $RSD_R\%$) given in Table 4.1, by poor repeatability of the replicate results and extreme values. (Although in practice each laboratory should use their own intra-laboratory reproducibility (S_{RW}) precision estimate for the analyte in question and the different laboratories would be replaced by results from different rounds of testing for any given laboratory). Nonetheless, the average uncertainty for each amino acid calculated across all the laboratories still provides some interesting results which can be compared to the individual values calculated next.

6.5.2 Measurement Uncertainty Evaluation for a single result.

Table 6.2 then looks at individual laboratory uncertainty estimates for each amino acid. Although this approach is not recommended and long term trends (as described above), give more appropriate approximations, it can be helpful to observe unexpected random error effects between rounds of proficiency testing. Here the individual bias components have been assessed separately as discussed in Section 6.2.1 and the CV% or RSD% determined from instrumental replicates have been used where available, in place the laboratory's own estimation of precision for that analyte, S_{RW} . However it should be noted that precision based on instrument repeatability is likely to be small compared to any long term true intra-laboratory reproducibility (intermediate precision) estimate and may contribute to smaller expanded uncertainties than might be otherwise expected.

Individual laboratory standard uncertainty components have been presented as histograms, together with each laboratory's combined uncertainty value and the average combined uncertainty for the test material described in the previous section and given in Table 6.1. In addition, expanded uncertainty confidence intervals have been determined and plotted for each amino acid to illustrate the effect of uncertainty on the mean of submitted results.

Table 6.1: Estimation of Relative Standard Uncertainty, Combined and Expanded Uncertainty for Amino Acids (using $\text{RMS}_{\text{bias}}\%$ to access bias contributions) across ALL Laboratories.

analyte	Std uncertainty contributions			Combined & Expanded uncertainties	
	Precision ¹	Bias components ^{2,3}		combined $u_c\%$	Expanded $U\% (k = 2)$
		1	2		
	$\hat{\sigma}$ as RSD%	$u(\hat{X})$ as RSU%	$\text{RMS}_{\text{bias}}\%$		
Asx D/L (all ^a)	4.14	1.07	12.12	12.85	25.71
Asx D/L (rpHPLC)	3.47	1.05	5.84	6.87	13.74
Glx D/L (all ^a)	6.85	1.90	12.98	14.80	29.60
Glx D/L (rpHPLC)	5.88	1.77	13.91	15.20	30.41
Ser D/L (rpHPLC)	3.57	1.08	2.27	4.36	8.72
Arg D/L (rpHPLC)	7.10	2.37	7.49	10.59	21.18
Ala D/L (all ^a)	16.09	4.15	13.88	21.65	43.30
Ala D/L (rpHPLC)	8.34	2.51	7.9	11.76	23.52
Val D/L (all ^a)	11.49	2.97	18.12	21.66	43.32
Val D/L (rpHPLC)	2.14	0.64	8.74	9.02	18.04
Phe D/L (all ^a)	9.49	2.74	7.26	12.26	24.52
Phe D/L (rpHPLC)	5.51	1.66	7.91	9.78	19.56
D-Aile/L-Ile (all ^b)	14.82	3.59	17.95	23.55	47.11
D-Aile/L-Ile (rpHPLC)	7.28	2.19	19.05	20.51	41.02
Leu D/L (all ^a)	10.18	2.94	11.84	15.89	31.78
Leu D/L (rpHPLC)	8.37	2.96	8.43	12.24	24.48
Tyr D/L (rpHPLC)	10.20	3.86	5.32	12.13	24.27

Notes for Table 6.1:

^a = rpHPLC and GC data ^b = rpHPLC, GC and HPLC-IE data

¹ = $\hat{\sigma}$ is the standard deviation for the assigned value, i.e., the median absolute deviation (sMAD), expressed as a percentage (given in Table 5.2).

² = $u(\hat{X})$ is the uncertainty of the assigned value (\hat{X}) expressed as a percentage, (given in Table 5.2).

³ = RMS_{bias} is the observed uncertainty due to bias of the submitted results

Table 6.2: Estimation of Relative Standard Uncertainty, Combined and Expanded Uncertainty Estimations for Individual Laboratories

laboratory number	mean result	Std uncertainty contributions				Combined & Expanded uncertainties	
		Precision ⁴	$u(\hat{X})$ as RSU% ⁵	$u(\bar{x})$ as RSU% ⁶	Relative bias % ⁷	combined u_c %	Expanded U % ($k = 2$)
Asx D/L	σ std dev as CV% ⁴	$u(\hat{X})$ as RSU% ⁵	$u(\bar{x})$ as RSU% ⁶	Relative bias % ⁷	combined u_c %	Expanded U % ($k = 2$)	
1	0.192	0.47	1.07	0.16	14.03	14.08	28.16
2	0.198	0.04	1.07	0.03	11.49	11.54	23.08
3	0.210	n=1	1.07	n=1	6.05		
4							
5							
6.1	0.225	8.89	1.07	4.44	0.79	10.03	20.05
6.2	0.299	1.00	1.07	0.71	33.94	33.98	67.95
7.1	0.216	n=1	1.07	n=1	3.24		
7.2	0.280	n=1	1.07	n=1	25.43		
8	0.217	0.00	1.07	0.00	2.79	2.99	5.98
9	0.232	0.11	1.07	0.08	4.04	4.18	8.35
10	0.226	0.03	1.07	0.02	1.20	1.61	3.22
11	0.222	0.51	1.07	0.36	0.47	1.33	2.65
12	0.228	0.04	1.07	0.03	1.95	2.22	4.44
13	0.223	n=1	1.07	n=1	0.00		
14	0.225	n=1	1.07	n=1	0.71		
15	0.221	0.73	1.07	0.52	1.03	1.73	3.47
Asx D/L rpHPLC	σ std dev as CV% ⁴	$u(\hat{X})$ as RSU% ⁵	$u(\bar{x})$ as RSU% ⁶	Relative bias % ⁷	combined u_c %	Expanded U % ($k = 2$)	
1	0.192	0.47	1.05	0.16	13.63	13.68	27.35
2	0.198	0.04	1.05	0.03	11.07	11.12	22.25
3	0.210	n=1	1.05	n=1	5.60		
4							
5							
6.1							
6.2							
7.1							
7.2							
8	0.217	0.00	1.05	0.00	2.34	2.56	5.12
9	0.232	0.11	1.05	0.08	4.52	4.65	9.29
10	0.226	0.03	1.05	0.02	1.68	1.98	3.96
11	0.222	0.51	1.05	0.36	0.00	1.22	2.44
12	0.228	0.04	1.05	0.03	2.42	2.64	5.28
13	0.223	n=1	1.05	n=1	0.47		
14	0.225	n=1	1.05	n=1	1.18		
15	0.221	0.73	1.05	0.52	0.57	1.49	2.97

⁴ = σ is the standard deviation of submitted results, expressed as a relative % i.e.; $CV\% = (\sigma/\bar{x}) \times 100$ (see Section 4).⁵ = $u(\hat{X})$ is the uncertainty of the assigned value (\hat{X}) expressed as a relative % i.e.; $RSU_{\hat{X}}\% = (u(\hat{X})/\hat{X}) \times 100$ (see Section 5).⁶ = $u(\bar{x})$ is the bias standard deviation for submitted results (\bar{x}) expressed as a relative % $RSU_{\bar{x}}\% = (u(\bar{x})/\bar{x}) \times 100$ (see Section 4).⁷ = Relative bias expressed as a % i.e.; $Bias\% = (\bar{x} - \hat{X})/\hat{X} \times 100$

Table 6.2: Estimation of Relative Standard Uncertainty, Combined and Expanded Uncertainty Estimations for Individual Laboratories (continued).

laboratory number	mean result	Std uncertainty contributions			Combined & Expanded uncertainties	
		Precision ⁴	$u(\hat{X})$ as RSU% ⁵	Bias components ^{5,6,7}	combined u_c %	Expanded $U\%$ ($k = 2$)
Glx D/L		σ std dev as CV% ⁴	$u(\hat{X})$ as RSU% ⁵	$u(\bar{x})$ as RSU% ⁶	Relative bias % ⁷	combined u_c %
1	0.049	0.25	1.90	0.08	20.14	20.23
2	0.044	0.58	1.90	0.41	28.85	28.92
3	0.045	n=1	1.90	n=1	27.26	
4						
5						
6.1	0.059	13.56	1.90	6.78	4.62	15.96
6.2						
7.1	0.057	n=1	1.90	n=1	7.85	
7.2						
8	0.060	0.00	1.90	0.00	3.00	3.55
9	0.067	0.16	1.90	0.11	7.66	7.89
10	0.065	1.00	1.90	0.71	4.80	5.31
11	0.064	0.67	1.90	0.47	4.16	4.65
12	0.064	0.78	1.90	0.55	4.24	4.75
13	0.062	n=1	1.90	n=1	0.00	
14	0.063	n=1	1.90	n=1	1.66	
15	0.062	0.05	1.90	0.03	0.27	1.92
Glx D/L rpHPLC		σ std dev as CV% ⁴	$u(\hat{X})$ as RSU% ⁵	$u(\bar{x})$ as RSU% ⁶	Relative bias % ⁷	combined u_c %
1	0.049	0.25	1.77	0.08	20.36	20.44
2	0.044	0.58	1.77	0.41	29.05	29.11
3	0.045	n=1	1.77	n=1	27.45	
4						
5						
6.1						
6.2						
7.1						
7.2						
8	0.060	0.00	1.77	0.00	3.26	3.71
9	0.067	0.16	1.77	0.11	7.37	7.58
10	0.065	1.00	1.77	0.71	4.52	5.01
11	0.064	0.67	1.77	0.47	3.88	4.35
12	0.064	0.78	1.77	0.55	3.96	4.44
13	0.062	n=1	1.77	n=1	0.27	
14	0.063	n=1	1.77	n=1	1.38	
15	0.062	0.05	1.77	0.03	0.00	1.77

⁴ = σ is the standard deviation of submitted results, expressed as a relative % i.e.; $CV\% = (\sigma/\bar{x}) \times 100$ (see Section 4).⁵ = $u(\hat{X})$ is the uncertainty of the assigned value (\hat{X}) expressed as a relative % i.e.; $RSU_{\hat{X}}\% = (u(\hat{X})/\hat{X}) \times 100$ (see Section 5).⁶ = $u(\bar{x})$ is the bias standard deviation for submitted results (\bar{x}) expressed as a relative % $RSU_{\bar{x}}\% = (u(\bar{x})/\bar{x}) \times 100$ (see Section 4).⁷ = Relative bias expressed as a % i.e.; $Bias\% = (\bar{x} - \hat{X}/\hat{X}) \times 100$

Table 6.2: Estimation of Relative Standard Uncertainty, Combined and Expanded Uncertainty Estimations for Individual Laboratories (continued).

laboratory number	mean result	Std uncertainty contributions			Combined & Expanded uncertainties		
		Precision ⁴	Bias components ^{5,6,7}		combined $u_c\%$	Expanded $U\% (k = 2)$	
Ser D/L		σ std dev as CV% ⁴	$u(\hat{X})$ as RSU% ⁵	$u(\bar{x})$ as RSU% ⁶	Relative bias % ⁷	combined $u_c\%$	Expanded $U\% (k = 2)$
1	0.108	1.10	1.08	0.37	3.31	3.66	7.33
2	0.111	0.19	1.08	0.13	0.25	1.13	2.26
3	0.116	n=1	1.08	n=1	3.63		
4							
5							
6.1							
6.2							
7.1							
7.2							
8	0.117	0.00	1.08	0.00	4.71	4.83	9.66
9	0.114	0.13	1.08	0.09	2.40	2.64	5.28
10	0.112	0.02	1.08	0.02	0.19	1.09	2.18
11	0.109	1.00	1.08	0.71	2.01	2.59	5.19
12	0.111	0.05	1.08	0.04	0.82	1.35	2.71
13	0.112	n=1	1.08	n=1	0.00		
14	0.115	n=1	1.08	n=1	2.88		
15	0.114	0.02	1.08	0.02	2.41	2.64	5.27
Arg D/L rpHPLC		σ std dev as CV% ⁴	$u(\hat{X})$ as RSU% ⁵	$u(\bar{x})$ as RSU% ⁶	Relative bias % ⁷	combined $u_c\%$	Expanded $U\% (k = 2)$
1							
2	0.088	0.08	2.37	0.06	12.04	12.27	24.55
3	0.105	n=1	2.37	n=1	4.79		
4							
5							
6.1							
6.2							
7.1							
7.2							
8							
9	0.114	1.06	2.37	0.75	14.75	14.99	29.98
10	0.109	2.13	2.37	1.51	9.35	9.99	19.98
11	0.097	4.65	2.37	3.29	2.39	6.61	13.23
12	0.100	6.37	2.37	4.51	0.13	8.16	16.32
13	0.095	n=1	2.37	n=1	5.09		
14	0.100	n=1	2.37	n=1	0.00		
15	0.099	4.29	2.37	3.03	1.08	5.86	11.73

⁴ = σ is the standard deviation of submitted results, expressed as a relative % i.e.; $CV\% = (\sigma/\bar{x}) \times 100$ (see Section 4).⁵ = $u(\hat{X})$ is the uncertainty of the assigned value (\hat{X}) expressed as a relative % i.e.; $RSU_{\hat{X}}\% = (u(\hat{X})/\hat{X}) \times 100$ (see Section 5)⁶ = $u(\bar{x})$ is the bias standard deviation for submitted results (\bar{x}) expressed as a relative % $RSU_{\bar{x}}\% = (u(\bar{x})/\bar{x}) \times 100$ (see Section 4).⁷ = Relative bias expressed as a % i.e.; $Bias\% = (\bar{x} - \hat{X}/\hat{X}) \times 100$

Table 6.2: Estimation of Relative Standard Uncertainty, Combined and Expanded Uncertainty Estimations for Individual Laboratories (continued).

laboratory number	mean result	Std uncertainty contributions				Combined & Expanded uncertainties	
		Precision ⁴	Bias components ^{5,6,7}			combined $u_c\%$	Expanded $U\% (k = 2)$
Ala D/L		σ std dev as CV% ⁴	$u(\hat{X})$ as RSU% ⁵	$u(\bar{x})$ as RSU% ⁶	Relative bias % ⁷		
1	0.060	14.80	4.15	4.93	3.62	16.54	33.09
2	0.058	0.13	4.15	0.09	7.00	8.14	16.28
3	0.069	n=1	4.15	n=1	10.85		
4							
5							
6.1	0.047	31.91	4.15	11.28	24.19	41.81	83.63
6.2	0.050	2.00	4.15	0.89	19.35	19.92	39.83
7.1	0.048	n=1	4.15	n=1	22.58		
7.2	0.047	n=1	4.15	n=1	24.19		
8	0.062	0.00	4.15	0.00	0.00	4.15	8.31
9	0.073	1.18	4.15	0.83	17.95	18.48	36.96
10	0.069	0.70	4.15	0.49	11.30	12.07	24.13
11	0.065	1.66	4.15	1.17	4.95	6.77	13.54
12	0.066	1.29	4.15	0.92	6.99	8.28	16.57
13	0.055	n=1	4.15	n=1	11.39		
14	0.063	n=1	4.15	n=1	1.83		
15	0.069	0.46	4.15	0.33	10.58	11.38	22.76
Ala D/L (rpHPLC)		σ std dev as CV% ⁴	$u(\hat{X})$ as RSU% ⁵	$u(\bar{x})$ as RSU% ⁶	Relative bias % ⁷	combined $u_c\%$	Expanded $U\% (k = 2)$
1	0.060	14.80	2.51	4.93	8.16	17.78	35.57
2	0.058	0.13	2.51	0.09	11.38	11.66	23.32
3	0.069	n=1	2.51	n=1	5.62		
4							
5							
6.1							
6.2							
7.1							
7.2							
8	0.062	0.00	2.51	0.00	4.71	5.34	10.69
9	0.073	1.18	2.51	0.83	12.39	12.73	25.45
10	0.069	0.70	2.51	0.49	6.05	6.61	13.21
11	0.065	1.66	2.51	1.17	0.00	3.23	6.47
12	0.066	1.29	2.51	0.92	1.95	3.55	7.11
13	0.055	n=1	2.51	n=1	15.57		
14	0.063	n=1	2.51	n=1	2.97		
15	0.069	0.46	2.51	0.33	5.37	5.95	11.91

⁴ = σ is the standard deviation of submitted results, expressed as a relative % i.e.; $CV\% = (\sigma/\bar{x}) \times 100$ (see Section 4).⁵ = $u(\hat{X})$ is the uncertainty of the assigned value (\hat{X}) expressed as a relative % i.e.; $RSU_{\hat{X}}\% = (u(\hat{X})/\hat{X}) \times 100$ (see Section 5)⁶ = $u(\bar{x})$ is the bias standard deviation for submitted results (\bar{x}) expressed as a relative % $RSU_{\bar{x}}\% = (u(\bar{x})/\bar{x}) \times 100$ (see Section 4).⁷ = Relative bias expressed as a % i.e.; $Bias\% = (\bar{x} - \hat{X}/\hat{X}) \times 100$

Table 6.2: Estimation of Relative Standard Uncertainty, Combined and Expanded Uncertainty Estimations for Individual Laboratories (continued)

laboratory number	mean result	Std uncertainty contributions				Combined & Expanded uncertainties	
		Precision ⁴	Bias components ^{5,6,7}			combined $u_c\%$	Expanded $U\% (k = 2)$
Val D/L	σ std dev as CV% ⁴	$u(\hat{X})$ as RSU% ⁵	$u(\bar{x})$ as RSU% ⁶	Relative bias % ⁷			
1	0.017	1.08	2.97	0.34	13.58	13.94	27.89
2	0.021	3.70	2.97	2.62	2.08	5.80	11.60
3	0.021	n=1	2.97	n=1	2.37		
4							
5							
6.1	0.015	6.67	2.97	2.52	25.57	26.71	53.42
6.2	0.014	7.14	2.97	3.19	30.53	31.66	63.32
7.1	0.012	n=1	2.97	n=1	40.46		
7.2	0.014	n=1	2.97	n=1	30.53		
8	0.021	10.35	2.97	7.32	1.72	13.13	26.26
9	0.022	1.40	2.97	0.99	7.75	8.48	16.95
10	0.020	5.62	2.97	3.97	0.00	7.49	14.99
11	0.019	3.01	2.97	2.13	3.61	5.95	11.90
12	0.020	10.62	2.97	7.51	0.31	13.34	26.69
13	0.016	n=1	2.97	n=1	22.39		
14	0.023	n=1	2.97	n=1	15.61		
15	0.020	6.01	2.97	4.25	1.46	8.07	16.13
Val D/L (rpHPLC)	σ std dev as CV% ⁴	$u(\hat{X})$ as RSU% ⁵	$u(\bar{x})$ as RSU% ⁶	Relative bias % ⁷	combined $u_c\%$	Expanded $U\% (k = 2)$	
1	0.017	1.08	0.64	0.34	14.82	14.88	29.76
2	0.021	3.70	0.64	2.62	0.61	4.62	9.23
3	0.021	n=1	0.64	n=1	0.90		
4							
5							
6.1							
6.2							
7.1							
7.2							
8	0.021	10.35	0.64	7.32	0.25	12.69	25.38
9	0.022	1.40	0.64	0.99	6.20	6.46	12.93
10	0.020	5.62	0.64	3.97	1.44	7.06	14.12
11	0.019	3.01	0.64	2.13	5.00	6.24	12.48
12	0.020	10.62	0.64	7.51	1.14	13.07	26.14
13	0.016	n=1	0.64	n=1	23.51		
14	0.023	n=1	0.64	n=1	13.94		
15	0.020	6.01	0.64	4.25	0.00	7.38	14.77

⁴ = σ is the standard deviation of submitted results, expressed as a relative % i.e.; $CV\% = (\sigma/\bar{x}) \times 100$ (see Section 4).⁵ = $u(\hat{X})$ is the uncertainty of the assigned value (\hat{X}) expressed as a relative % i.e.; $RSU_{\hat{X}}\% = (u(\hat{X})/\hat{X}) \times 100$ (see Section 5)⁶ = $u(\bar{x})$ is the bias standard deviation for submitted results (\bar{x}) expressed as a relative % $RSU_{\bar{x}}\% = (u(\bar{x})/\bar{x}) \times 100$ (see Section 4).⁷ = Relative bias expressed as a % i.e.; $Bias\% = (\bar{x} - \hat{X}/\hat{X}) \times 100$

Table 6.2: Estimation of Relative Standard Uncertainty, Combined and Expanded Uncertainty Estimations for Individual Laboratories (continued).

laboratory number	mean result	Std uncertainty contributions			Combined & Expanded uncertainties	
		Precision ⁴	Bias components ^{5,6,7}		combined $u_c\%$	Expanded $U\% (k = 2)$
Phe D/L		σ std dev as CV% ⁴	$u(\hat{X})$ as RSU% ⁵	$u(\bar{x})$ as RSU% ⁶	Relative bias % ⁷	
1	0.050	3.21	2.74	1.07	6.70	7.99
2	0.046	1.50	2.74	1.06	14.12	14.50
3	0.049	n=1	2.74	n=1	8.55	
4						
5						
6.1	0.051	25.49	2.74	9.01	5.10	27.65
6.2						55.30
7.1						
7.2						
8	0.051	0.00	2.74	0.00	5.10	5.79
9	0.060	0.39	2.74	0.28	10.90	11.25
10	0.058	1.44	2.74	1.02	7.71	8.37
11	0.056	0.77	2.74	0.54	4.99	5.77
12	0.056	1.82	2.74	1.29	3.87	5.24
13	0.057	n=1	2.74	n=1	6.75	
14	0.054	n=1	2.74	n=1	0.03	
15	0.056	0.37	2.74	0.26	3.85	4.75
						9.49
Phe D/L (rpHPLC)		σ std dev as CV% ⁴	$u(\hat{X})$ as RSU% ⁵	$u(\bar{x})$ as RSU% ⁶	Relative bias % ⁷	combined $u_c\%$
						Expanded $U\% (k = 2)$
1	0.050	3.21	1.66	1.07	10.16	10.83
2	0.046	1.50	1.66	1.06	17.30	17.48
3	0.049	n=1	1.66	n=1	11.94	
4						
5						
6.1						
6.2						
7.1						
7.2						
8	0.051	0.00	1.66	0.00	8.62	8.77
9	0.060	0.39	1.66	0.28	6.79	7.01
10	0.058	1.44	1.66	1.02	3.72	4.43
11	0.056	0.77	1.66	0.54	1.10	2.20
12	0.056	1.82	1.66	1.29	0.02	2.78
13	0.057	n=1	1.66	n=1	2.79	
14	0.054	n=1	1.66	n=1	3.68	
15	0.056	0.37	1.66	0.26	0.00	1.72
						3.44

⁴ = σ is the standard deviation of submitted results, expressed as a relative % i.e.; $CV\% = (\sigma/\bar{x}) \times 100$ (see Section 4).⁵ = $u(\hat{X})$ is the uncertainty of the assigned value (\hat{X}) expressed as a relative % i.e.; $RSU_{\hat{X}}\% = (u(\hat{X})/\hat{X}) \times 100$ (see Section 5)⁶ = $u(\bar{x})$ is the bias standard deviation for submitted results (\bar{x}) expressed as a relative % $RSU_{\bar{x}}\% = (u(\bar{x})/\bar{x}) \times 100$ (see Section 4).⁷ = Relative bias expressed as a % i.e.; $Bias\% = (\bar{x} - \hat{X})/\hat{X} \times 100$

Table 6.2: Estimation of Relative Standard Uncertainty, Combined and Expanded Uncertainty Estimations for Individual Laboratories (continued).

laboratory number	mean result	Std uncertainty contributions			Combined & Expanded uncertainties		
		Precision ⁴	Bias components ^{5,6,7}		combined $u_c\%$	Expanded $U\% (k = 2)$	
D-Aile/L-Ile		σ std dev as CV% ⁴	$u(\hat{X})$ as RSU% ⁵	$u(\bar{x})$ as RSU% ⁶	Relative bias % ⁷	combined $u_c\%$	Expanded $U\% (k = 2)$
1	0.018	3.38	3.59	1.07	38.39	38.72	77.44
2	0.033	2.33	3.59	1.65	10.00	11.00	22.00
3	0.039	n=1	3.59	n=1	29.12		
4	0.024	0.00	3.59	0.00	20.00	20.32	40.64
5	0.024	0.00	3.59	0.00	20.00	20.32	40.64
6.1	0.026	3.85	3.59	1.36	13.33	14.40	28.80
6.2	0.030	3.33	3.59	1.49	0.00	5.12	10.25
7.1	0.025	n=1	3.59	n=1	16.67		
7.2	0.028	n=1	3.59	n=1	6.67		
8	0.019	0.00	3.59	0.00	36.67	36.84	73.68
9	0.035	2.48	3.59	1.75	16.51	17.17	34.34
10	0.031	11.03	3.59	7.80	4.22	14.61	29.21
11	0.030	0.53	3.59	0.38	0.03	3.65	7.31
12	0.033	3.36	3.59	2.37	9.33	10.81	21.63
13	0.032	n=1	3.59	n=1	5.71		
14	0.031	n=1	3.59	n=1	2.96		
15	0.030	4.10	3.59	2.90	0.41	6.19	12.38
D-Aile/L-Ile rpHPLC		σ std dev as CV% ⁴	$u(\hat{X})$ as RSU% ⁵	$u(\bar{x})$ as RSU% ⁶	Relative bias % ⁷	combined $u_c\%$	Expanded $U\% (k = 2)$
1	0.018	3.38	2.19	1.07	40.88	41.10	82.19
2	0.033	2.33	2.19	1.65	5.55	6.61	13.22
3	0.039	n=1	2.19	n=1	23.89		
4							
5							
6.1							
6.2							
7.1							
7.2							
8	0.019	0.00	2.19	0.00	39.23	39.29	78.58
9	0.035	2.48	2.19	1.75	11.80	12.38	24.75
10	0.031	11.03	2.19	7.80	0.00	13.69	27.38
11	0.030	0.53	2.19	0.38	4.08	4.68	9.36
12	0.033	3.36	2.19	2.37	4.91	6.77	13.54
13	0.032	n=1	2.19	n=1	1.43		
14	0.031	n=1	2.19	n=1	1.21		
15	0.030	4.10	2.19	2.90	3.66	6.59	13.18

⁴ = σ is the standard deviation of submitted results, expressed as a relative % i.e.; $CV\% = (\sigma/\bar{x}) \times 100$ (see Section 4).⁵ = $u(\hat{X})$ is the uncertainty of the assigned value (\hat{X}) expressed as a relative % i.e.; $RSU_{\hat{X}}\% = (u(\hat{X})/\hat{X}) \times 100$ (see Section 5)⁶ = $u(\bar{x})$ is the bias standard deviation for submitted results (\bar{x}) expressed as a relative % $RSU_{\bar{x}}\% = (u(\bar{x})/\bar{x}) \times 100$ (see Section 4).⁷ = Relative bias expressed as a % i.e.; $Bias\% = (\bar{x} - \hat{X}/\hat{X}) \times 100$

Table 6.2: Estimation of Relative Standard Uncertainty, Combined and Expanded Uncertainty Estimations for Individual Laboratories (continued).

laboratory number	mean result	Std uncertainty contributions			Combined & Expanded uncertainties	
		Precision ⁴	Bias components ^{5,6,7}		combined $u_c\%$	Expanded $U\% (k = 2)$
Leu D/L		σ std dev as CV% ⁴	$u(\hat{X})$ as RSU% ⁵	$u(\bar{x})$ as RSU% ⁶	Relative bias % ⁷	
1	0.048	8.22	2.94	2.60	7.00	11.48
2						22.97
3						
4						
5						
6.1	0.036	5.56	2.94	1.96	30.05	30.77
6.2	0.048	2.08	2.94	0.93	6.74	7.70
7.1	0.058	n=1	2.94	n=1	12.69	
7.2	0.052	n=1	2.94	n=1	1.04	
8	0.044	0.00	2.94	0.00	14.51	14.80
9	0.052	2.07	2.94	1.47	1.35	4.11
10	0.060	29.89	2.94	21.13	17.25	40.57
11	0.053	1.13	2.94	0.80	3.21	4.57
12	0.051	0.04	2.94	0.03	1.04	3.12
13						6.23
14	0.054	n=1	2.94	n=1	4.30	
15	0.047	0.42	2.94	0.30	7.88	8.42
						16.85
Leu D/L rpHPLC		σ std dev as CV% ⁴	$u(\hat{X})$ as RSU% ⁵	$u(\bar{x})$ as RSU% ⁶	Relative bias % ⁷	combined $u_c\%$
						Expanded $U\% (k = 2)$
1	0.048	8.22	2.96	2.60	7.15	11.58
2						23.16
3						
4						
5						
6.1						
6.2						
7.1						
7.2						
8	0.044	0.00	2.96	0.00	14.64	14.94
9	0.052	2.07	2.96	1.47	1.19	4.08
10	0.060	29.89	2.96	21.13	17.07	40.49
11	0.053	1.13	2.96	0.80	3.05	4.47
12	0.051	0.04	2.96	0.03	1.19	3.19
13						6.38
14	0.054	n=1	2.96	n=1	4.14	
15	0.047	0.42	2.96	0.30	8.02	8.57
						17.13

⁴ = σ is the standard deviation of submitted results, expressed as a relative % i.e.; $CV\% = (\sigma/\bar{x}) \times 100$ (see Section 4).⁵ = $u(\hat{X})$ is the uncertainty of the assigned value (\hat{X}) expressed as a relative % i.e.; $RSU_{\hat{X}}\% = (u(\hat{X})/\hat{X}) \times 100$ (see Section 5)⁶ = $u(\bar{x})$ is the bias standard deviation for submitted results (\bar{x}) expressed as a relative % $RSU_{\bar{x}}\% = (u(\bar{x})/\bar{x}) \times 100$ (see Section 4).⁷ = Relative bias expressed as a % i.e.; $Bias\% = (\bar{x} - \hat{X}/\hat{X}) \times 100$

Table 6.2: Estimation of Relative Standard Uncertainty, Combined and Expanded Uncertainty Estimations for Individual Laboratories (continued).

laboratory number	mean result	Std uncertainty contributions			Combined & Expanded uncertainties	
		Precision ⁴	Bias components ^{5,6,7}		combined $u_c\%$	Expanded $U\% (k = 2)$
Try D/L		σ std dev as CV% ⁴	$u(\hat{X})$ as RSU% ⁵	$u(\bar{x})$ as RSU% ⁶	Relative bias % ⁷	
1						
2						
3						
4						
5						
6.1						
6.2						
7.1						
7.2						
8						
9	0.064	0.46	3.86	0.33	6.79	7.83
10	0.060	0.55	3.86	0.39	0.00	3.91
11	0.065	n=1	3.86	n=1	8.73	
12	0.060	6.74	3.86	4.77	0.01	9.11
13	0.055	n=1	3.86	n=1	8.71	18.22
14	0.056	n=1	3.86	n=1	6.88	
15	0.054	0.59	3.86	0.42	9.84	10.60
						21.19

⁴ = σ is the standard deviation of submitted results, expressed as a relative % i.e.; $CV\% = (\sigma/\bar{x}) \times 100$ (see Section 4).

⁵ = $u(\hat{X})$ is the uncertainty of the assigned value (\hat{X}) expressed as a relative % i.e.; $RSU_{\hat{X}}\% = (u(\hat{X})/\hat{X}) \times 100$ (see Section 5)

⁶ = $u(\bar{x})$ is the bias standard deviation for submitted results (\bar{x}) expressed as a relative % $RSU_{\bar{x}}\% = (u(\bar{x})/\bar{x}) \times 100$ (see Section 4).

⁷ = Relative bias expressed as a % i.e.; $Bias\% = (\bar{x} - \hat{X}/\hat{X}) \times 100$

Figure 6.2: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for **Aspartic acid / Asparagine D/L** Values in Ostrich Egg Shell (B) Test Material




Figure 6.3: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on **Aspartic acid / Asparagine D/L** Values in Ostrich Egg Shell (B) Test Material




Figure 6.4: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for **Aspartic acid / Asparagine rpHPLC D/L** Values in Ostrich Egg Shell (B) Test Material




Figure 6.5: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on **Aspartic acid / Asparagine rpHPLC D/L** Values in Ostrich Egg Shell (B) Test Material




Figure 6.6: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for **Glutamic acid / Glutamine D/L** Values in Ostrich Egg Shell (B) Test Material




Figure 6.7: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on **Glutamic acid / Glutamine D/L** Values in Ostrich Egg Shell (B) Test Material




Figure 6.8: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for **Glutamic acid /Glutamine rpHPLC D/L** Values in Ostrich Egg Shell (B) Test Material




Figure 6.9: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on **Glutamic acid / Glutamine rpHPLC D/L** Values in Ostrich Egg Shell (B) Test Material




Figure 6.10: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Serine D/L Values in Ostrich Egg Shell (B) Test Material




Figure 6.11: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Serine D/L Values in Ostrich Egg Shell (B) Test Material




Figure 6.12: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for **Arginine D/L** Values in Ostrich Egg Shell (B) Test Material




Figure 6.13: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on **Arginine D/L** Values in Ostrich Egg Shell (B) Test Material




Figure 6.14: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Alanine D/L Values in Ostrich Egg Shell (B) Test Material




Figure 6.15: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Alanine D/L Values in Ostrich Egg Shell (B) Test Material




Figure 6.16: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for **Alanine (rpHPLC)** D/L Values in Ostrich Egg Shell (B) Test Material




Figure 6.17: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on **Alanine (rpHPLC)** D/L Values in Ostrich Egg Shell (B) Test Material




Figure 6.18: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Valine D/L Values in Ostrich Egg Shell (B) Test Material




Figure 6.19: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Valine D/L Values in Ostrich Egg Shell (B) Test Material




Figure 6.20: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for **Valine (rpHPLC) D/L** Values in Ostrich Egg Shell (B) Test Material




Figure 6.21: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on **Valine (rpHPLC) D/L** Values in Ostrich Egg Shell (B) Test Material




Figure 6.22: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for **Phenylalanine D/L** Values in Ostrich Egg Shell (B) Test Material




Figure 6.23: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on **Phenylalanine D/L** Values in Ostrich Egg Shell (B) Test Material




Figure 6.24: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for **Phenylalanine (rpHPLC) D/L** Values in Ostrich Egg Shell (B) Test Material




Figure 6.25: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on **Phenylalanine (rpHPLC) D/L** Values in Ostrich Egg Shell (B) Test Material




Figure 6.26: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for **D-Alloisoleucine/L-Isoleucine** Values in Ostrich Egg Shell (B) Test Material




Figure 6.27: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on **D-Alloisoleucine/L-Isoleucine** Values in Ostrich Egg Shell (B) Test Material




Figure 6.28: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for **D-Alloisoleucine/L-Isoleucine rpHPLC** Values in Ostrich Egg Shell (B) Test Material




Figure 6.29: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on **D-Alloisoleucine/L-Isoleucine rpHPLC** Values in Ostrich Egg Shell (B) Test Material




Figure 6.30: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Leucine D/L Values in Ostrich Egg Shell (B) Test Material




Figure 6.31: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Leucine D/L Values in Ostrich Egg Shell (B) Test Material




Figure 6.32: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for **Leucine rpHPLC D/L** Values in Ostrich Egg Shell (B) Test Material




Figure 6.33: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on **Leucine rpHPLC D/L** Values in Ostrich Egg Shell (B) Test Material




Figure 6.34: Standard Uncertainty Contributions and Combined Uncertainty for each Laboratory against an Estimated Average Combined Uncertainty for Tyrosine D/L Values in Ostrich Egg Shell (B) Test Material





Figure 6.35: Effect of Expanded Uncertainty for each Laboratory at 95% Confidence on Tyrosine D/L Values in Ostrich Egg Shell (B) Test Material



Appendix 1: Analytical Methods Used by Participants

Reverse Phase HPLC/ HPLC-Ion Exchange

REFERENCES	
Please give details of any method relevant references;	
Kaufman & Manley 1998	009, 010, 011, 012, 013, 014, 015
HYDROLYSIS FOR THAA's	
Sample Weight used for analysis (mg):	
3.5 – 5 mg	003
1 – 10 mg	008, 009, 010, 011, 012, 013, 014, 015
>10 – 20 mg	001, 002, 004, 005,
Vials used for hydrolysis:	
Glass	001, 002, 003, 004, 005, 008, 009, 010, 011, 012, 013, 014, 015
Acid Used:	
7M HCl	001, 002, 003, 004, 005, 008, 009, 010, 011, 012, 013, 014, 015
Vials flushed with N ₂ :	
Yes	001, 002, 003, 004, 005, 008, 009, 010, 011, 012, 013, 014, 015
Please give details of any other treatment prior to hydrolysis:	
Comments received;	
1)20µl/mg of 7M HCl added to samples	001, 009, 010, 011, 012, 013, 014, 015
2)2ml hydrolysis vials used	009, 010, 011, 012, 013, 014, 015
3)samples weighed & transferred to microvial or 4ml vial depending on size.	002, 003, 004, 005
Oven Temperature (°C):	
100 °C	001
110 °C	009, 010, 011, 012, 013, 014, 015
Heating Time (hours):	
6 hrs	002, 003
20 hrs	001
22 hrs	004, 005, 008
24 hrs	009, 010, 011, 012, 013, 014, 015
Was sample dried prior to analysis?:	
Yes	001, 002, 003, 004, 005, 008, 009, 010, 011, 012, 013, 014, 015
Please give details of sample drying conditions:	
Under vacuum	001, 002, 003, 004, 005, 008, 009, 010, 011, 012, 013, 014, 015
Ambient / room temp	001, 002, 003, 004, 005, 008, 009, 010, 011, 012, 013, 014, 015
Dried overnight	001, 002, 003, 004, 005, 008, 009, 010, 011, 012, 013, 014, 015

THAA's REHYDRATION	
Volume of rehydration fluid added as $\mu\text{l}/\text{mg}$ of original sample	
10 $\mu\text{l}/\text{mg}$	001
20 $\mu\text{l}/\text{mg}$	002, 003, 004, 005, 008, 009, 010, 011, 012, 013, 014, 015
Internal Standard Used?:	
L-homo-Arginine	001, 002, 003, 008, 009, 010, 011, 012, 013, 014, 015
Norleucine	004, 005
Concentration of Internal std used (M):	
0.03 mM	001
0.01mM	002, 003, 008, 009, 010, 011, 012, 013, 014, 015
6.25 mM	004, 005
Source / supplier of internal standard:	
Sigma	001, 002, 003, 004, 005
Sigma Aldrich (Fluka)	008
Other constituents and their concentrations (M or mM) in rehydration fluid:	
0.01M HCl	002, 003, 004, 005, 009, 010, 011, 012, 013, 014, 015
1.5mM Sodium Azide	009, 010, 011, 012, 013, 014, 015
ANALYSIS	
Please state method used	
Reverse phase HPLC	001, 002, 003, 008, 009, 010, 011, 012, 013, 014, 015
Ion Exchange HPLC	004, 005
Instrument used	
Agilent 1100 Series	001, 008, 009, 012, 013
Agilent / Hewlet Packard 1100 Series	002, 003, 010, 011, 014, 015
Agilent 1200 Series	004, 005
Agilent 6890 GC, Flame Ionization	006, 007
Pre-column Derivatization Reagent constituents and their concentrations (M or mM):	
OPA 170 mM	001, 002, 003, 008, 009, 010, 011, 012, 013, 014, 015
IBLC 260 mM	001, 002, 003, 008, 009, 010, 011, 012, 013, 014, 015
Potassium borate buffer 1M	001, 002, 003, 008, 009, 010, 011, 012, 013, 014, 015
pH adjusted to:	
10.4	001, 002, 003, 008, 009, 010, 011, 012, 013, 014, 015
Sample injection volume (μl)	
2 μl	001, 002, 003, 009, 010, 011, 012, 013, 014, 015
4 μl	008
20 μl	004, 005

HPLC COLUMN	
Column Make/Type & Phase(i.e.; Hypersil BDS)/ Batch Number:	
Thermo/Hypersil BDS C18/0742018X Hypersil BDS Hypersil BDS /5/120/4772 Pickering Labs Sodium Cation Exchange Supelcosil LC-18-DB(rp)/6520/5-1452	001 009, 010, 011, 012, 013, 014, 015 002, 003 004, 005 008
Column Packing:	
Silica Sodium Functional group; C ₁₈ End capped (Yes)	002, 003, 008 004, 005 001, 002, 003, 008, 009, 010, 011, 012, 013, 014, 015 002, 003, 008
Column width (mm)	
3mm 5mm	001, 002, 003, 004, 005 009, 010, 011, 012, 013, 014, 015
Column length (mm)	
250mm	001, 002, 003, 004, 005, 009, 010, 011, 012, 013, 014, 015
Guard Column not used	
No	001, 002, 003, 004, 005
HPLC Column Temperature (°C):	
25 °C 30 °C	001, 009, 010, 011, 012, 013, 014, 015 002, 003, 004, 005, 008
MOBILE PHASE	
Mobile phase programme:	
Gradient	001, 002, 003, 004, 005, 009, 010, 011, 012, 013, 014, 015
Mobile phase components (please state; i.e.; sodium acetate buffer/ methanol/ acetonitrile):	
Sodium acetate Buffer (pH 6.00) Methanol Acetonitrile Sodium citrate buffer (pH 3.12) Sodium citrate buffer (pH 3.86) Sodium chloride buffer (pH 11.5)	001, 002, 003, 008, 009, 010, 011, 012, 013, 014, 015 001, 002, 003, 008, 009, 010, 011, 012, 013, 014, 015 001, 002, 003, 008, 009, 010, 011, 012, 013, 014, 015 004, 005 004, 005 004, 005
Sodium acetate Buffer (pH 6.00) Gradient: Starting % Final % time (mins) flow rate (ml/min)	
95% 76.6% 31mins 0.56ml/min 76.6% 46.2% 95min 0.60ml/min 95% 5% 83min 0.500ml/min 95% 50% 88min 0.560ml/min 95% % 95min 0.56ml/min	001a 001b 002, 003 008 009, 010, 011, 012, 013, 014, 015

MOBILE PHASE continued	
Methanol Gradient: Starting % Final % time (mins) flow rate (ml/min)	
5% 23% 31mins 0.56ml/min 23% 48.8% 95min 0.60ml/min 5% 95% 83min 0.500ml/min 5% 45% 88min 0.560ml/min 5% 50% 95min 0.56ml/min	001a 001b 002, 003 008 009, 010, 011, 012, 013, 014, 015
Acetonitrile Gradient: Starting % Final % time (mins) flow rate (ml/min)	
0% 0.4% 31mins 0.56ml/min 0.4% 5% 95min 0.60ml/min 0.4% 5% 83min 0.500ml/min 0% 5% 88min 0.560ml/min 0% 5% 95min 0.56ml/min	001a 001b 002, 003 008 009, 010, 011, 012, 013, 014, 015
Sodium citrate buffer (pH3.12) Gradient: Starting % Final % time (mins) flow rate (ml/min)	
100% 0% 99mins 0.140ml/min	004, 005
Sodium citrate buffer (pH3.86) Gradient: Starting % Final % time (mins) flow rate (ml/min)	
0% 0% 99mins 0.140ml/min	004, 005
Sodium chloride buffer (pH11.5) Gradient: Starting % Final % time (mins) flow rate (ml/min)	
0% 100% 99mins 0.140ml/min	004, 005
Post-column Derivatization Reagent constituents and their concentrations (M or mM):	
Boric Acid 0.5M OPA 0.0075M Ethanol 1% 2-mercaptoethanol 0.00075%	004,005 004,005 004,005 004,005
pH adjusted to 10.4	004,005
DETECTION	
Detector Type:	
Fluorescence	001, 002, 003, 004, 005, 008, 009, 010, 011, 012, 013, 014, 015
Excitation wavelength (nm):	
230 250 335 340	008, 009, 010, 011, 012, 013, 014, 015 002, 003 001 004, 005
Emission wavelength (nm):	
410 445 455	002, 003 001, 008, 009, 010, 011, 012, 013, 014, 015 004, 005

Gas Chromatography

REFERENCES	
Please give details of any method relevant references;	
Goodfriend 1991 with modifications	006, 007
HYDROLYSIS FOR THAA's	
Sample Weight used for analysis (mg):	
75 - 90 mg	006, 007
Vials used for hydrolysis:	
Glass	006, 007
Acid Used:	
6M HCl	006, 007
Vials flushed with N ₂ :	
Yes	006, 007
Please give details of any other treatment prior to hydrolysis:	
Comments received (006, 007); Samples weighed into hydrolysis vials without drying; fossil samples are always dried in vacuo prior to weighing for hydrolysis.	
Oven Temperature (°C):	
105 °C	006, 007
Heating Time (hours):	
22 hrs	006, 007
SAMPLE CLEAN UP / DESALTING	
Was cation exchange resin used?	
No	006, 007
Was HF used to separate amino acids from precipitate?	
Yes	006, 007
Was sample dried prior to Derivatization?:	
Yes	006, 007
Please give details of sample drying conditions:	
Under nitrogen stream	006, 007
Drying Temp; 50 °C (in heating block)	006, 007
Drying time; 1 hr	006, 007

SAMPLE CLEAN UP / DESALTING continued	
Comments received (006, 007); After HF removal of Ca, solution of AA was dried under N ₂ to remove HF, then transferred with 1N HCl to a glass vial for additional N ₂ drying and vacuum oven drying (total drying time ~2 hours at 60 deg C). This dried residue was then ready for esterification.	
ESTERIFICATION	
Esterification reagents:	
isopropanol	006, 007
Esterification conditions:	
Flushed under nitrogen Oven Temperature; 50°C Heating time; 1hr	006, 007 006, 007 006, 007
Was sample dried prior to acylation?:	
Yes	006, 007
Please give details of sample drying conditions:	
Under vacuum Under nitrogen stream Drying Temp; 55 °C Drying time; 1 hr	006, 007 006, 007 006, 007 006, 007
ACYLATION	
Acylation reagents:	
TFAA	006, 007
Acylation conditions:	
Flushed under nitrogen Room Temperature Heating time; 2hr minimum	006, 007 006, 007 006, 007
Comments received (006, 007); Isopropanol has to be removed before TFA can be added (with Methylene chloride)	
Was sample dried prior to GC analysis?	
Yes	006, 007
Please give details of sample drying conditions:	
Flushed under nitrogen Room Temperature Heating time; <5 minutes	006, 007 006, 007 006, 007
Comments received (006, 007); Derivative is in TFA/Meth Chloride – this solution was dried under N ₂ and transferred to small vials for storage and GC injection; final solution containing derivative is in cyclohexane. Derivatives are injected on GC using cyclohexane	

THAA's REHYDRATION	
Volume of rehydration fluid added as μ l	
20 – 30 μ l	006, 007
Internal Standard Used?:	
No	006, 007
ANALYSIS	
Sample injection volume (μ l)	
1 -3 μ l	006, 007
GC injection mode:	
Splitless	006, 007
GC COLUMN	
Column Type;	
Capillary	006, 007
Column Make / Batch Number:	
Alltech, Catalog #13633, Serial # 5653, purchased in 1998, in continuous use	006, 007
Column Packing:	
Chiral Phase: Chirasil-val	006, 007
Column width (mm)	
0.25mm	006, 007
Column length (mm)	
25m	006, 007
Column Temperature ($^{\circ}$ C):	
See below for program	006, 007
Mobile phase / Carrier gas	
Helium	006, 007
Mobile phase flow rate (ml/min):	
Flow variable with temperature; pressure 7.6psi	006, 007

DETECTION	
Detector Type:	
Flame ionisation	006, 007
Comments received (006, 007); NDP not used for these samples, but used in previous studies – both NPD and FID give same D/L values	
ANYTHING ELSE?	
Please use this space for any additional information you would like to record concerning method details not covered above:	
Comments received (006, 007); Summary of the preparation sequence: 1) Dissolution in stoichiometric amount of conc. HCl to bring final solution to 6N 2) Purge with N2, seal hydrolysis tube, hydrolyse for 22 hours at 105 deg. 3) After hydrolysis, HCl solution is transferred to plastic centrifuge tube and appropriate amount of HF is added to remove Ca. After centrifuging, solution is transferred to another plastic tube for N2 drydown in a heating block (~60 deg). Drydown requires about one hour. 4) Dried residue is transferred using ~0.2 ml 1N HCl to a screwcap vial. This solution is dried with N2, then further dried in a vacuum oven (1 hour, 50 deg.) prior to esterification with isopropanol. 5) Isopropanol esterification – one hour at 105 deg. 6) Isopropanol is then dried down with N2 in 50 deg heating block (~10 minutes), then methylene chloride (Dichloromethane, or DCM) and TFA are added. This complete derivative is then usually stored overnight prior to GC analysis. 7) The DCM/TFA solution is transferred to a small GC vial, dried with N2, then cyclohexane is added to ready the derivative for GC injection. The amount of cyclohexane is variable depending on the sample size, but there is no “formula” for this because the GC analysis is not quantitative. Derivatives remain in the cyclohexane solution until GC injection – in most cases, five or six chromatograms are obtained over a period of one to two weeks. Injection amounts are usually 1 μ l; if samples are small, 2 or even 3 μ l will be injected. 8) GC temperature program: inject at 60 deg, hold for one minute; 20 deg/min up to 80 deg; hold for 10 minutes; 0.85 deg/min to 135 deg, 1 minute hold; 5 deg/min to 160, 10 minutes hold; recycle. All important peaks are eluted within 100 minutes; last phases of temperature program are to clean out the column.	

Internal Quality Control

INSTRUMENT CALIBRATION	
Was the instrument calibrated prior to analysis?	
Yes, prior to analytical run Yes, within the last year No	001 008 002, 003, 004, 005, 006, 007, 009, 010, 011, 012, 013, 014, 015
If Yes, type of calibration:	
Calibration curve/std addition-single level Calibrated by Agilent Technician	001 008
If Yes, what reference materials / standards are used?	
In-house std solution(s) NB: Solution prepared from single powdered AA standards	001
Source of reference materials/standards:	
Sigma	001
RECOVERY OR INTERNAL STANDARD	
Was % recovery determined?	
No	001, 002, 003, 004, 005, 008, 009, 010, 011, 012, 013, 014, 015
If No, was an internal standard used?	
Yes, as component of rehydration fluid	001, 002, 003, 004, 005, 008, 009, 010, 011, 012, 013, 014, 015
Internal Standard Used?:	
L-homo-Arginine Norleucine No	001, 002, 003, 008, 009, 010, 011, 012, 013, 014, 015 004, 005 006, 007
Concentration of Internal std used (M):	
0.03 mM 0.01mM 6.25 mM	001 002, 003, 008, 009, 010, 011, 012, 013, 014, 015 004, 005
Source / supplier of internal standard:	
Sigma Sigma Aldrich (Fluka)	001, 002, 003, 004, 005 008

D/L RATIO CALCULATION	
Do you routinely calculate concentrations?	
Yes	001, 009, 010, 011, 012, 013, 014, 015
No	002, 003, 004, 005, 006, 007, 008
Comments received; (001) Concentration of a single enantiomer in solution (milimol/L)= (enantiomer area x Internal Standard concentration)/ Internal Standard area Concentration of a single enantiomer in the sample (picomol/mg)= [Concentration of enantiomer in solution (milimol/L) x Volume of rehydration fluid added (L) x 10-9 picomol/milimol)]/sample weight (mg) (006, 007): Only peak areas are reported under most circumstances but both are measured to check for reliability and peak distortion/overload.	
D/L values are routinely calculated using:	
Peak heights	004, 005, 006, 007
Peak areas	001, 002, 003, 006, 007, 008
Concentrations based on peak areas	009, 010, 011, 012, 013, 014, 015
QUALITY CONTROL	
Do you routinely use lab QC materials or standards.	
Yes	001, 002, 003, 004, 005, 006, 007, 008, 009, 010, 011, 012, 013, 014, 015
If Yes, are they:	
In-house std solution(s) (Matrix-matched) ILC stds (Wehmiller)	001, 002, 003, 004, 005, 009, 010, 011, 012, 013, 014, 015 002, 003, 004, 005, 006, 007, 008, 009, 010, 011, 012, 013, 014, 015
Source of QC materials:	
Sigma J.F. Wehmiller	001, 002, 003, 004, 005, 009, 010, 011, 012, 013, 014, 015 002, 003, 004, 005, 006, 007, 008, 009, 010, 011, 012, 013, 014, 015
How do you use QC materials?	
Control charts Visual inspection of chromatograms/data D/L comparison to lit Comparison in ILC's with long term mean	001, 002, 003, 004, 005 008, 009, 010, 011, 012, 013, 014, 015 008 006, 007
MEASUREMENT UNCERTAINTY	
How do you determine Measurement Uncertainty (MU) of your data	
As the standard deviation	001, 002, 003, 004, 005, 006, 007, 008, 009, 010, 011, 012, 013, 014, 015
If you do, how often do you determine the MU?	
Routinely per run Approx once a month When its needed As the SD of multiple chromatograms from each derivative.	008 002, 003, 004, 005, 001, 009, 010, 011, 012, 013, 014, 015 006, 007, 009, 010, 011, 012, 013, 014, 015

Appendix 2: Glossary of Abbreviations, Symbols, Terms & Definitions

Abbreviations

ANOVA	Analysis of Variance
CRM	Certified Reference Material
CV	Coefficient of Variation
EQC	External Quality Control
IQC	Internal Quality Control
MU	Uncertainty of Measurement / Measurement Uncertainty
PT	Proficiency test
QA	Quality Assurance
QC	Quality Control

Symbols

k	Coverage Factor
RMS_{bias}	Bias Root Mean Square
$RSD_L\%$	Relative Between Sample Standard Deviation (expressed as a percentage)
$RSU\%$	Relative Standard Uncertainty (expressed as a percentage)
$RSD\%$	Relative standard deviation (expressed as a percentage)
$RSD_r\%$	Relative Repeatability standard deviation (expressed as a percentage)
$RSD_R\%$	Relative Reproducibility standard deviation (expressed as a percentage)
s_{an}	(Homogeneity) Analytical Precision
s_{an}^2	(Homogeneity) Analytical Variance
s_{sam}	(Homogeneity) Sampling Precision
s_{sam}^2	(Homogeneity) Sampling Variance
s_{all}^2	(Homogeneity) Total Permissible Sampling Variance
s, sd or σ	Standard Deviation
S_L	Between-sample standard deviation
S_r	Repeatability Standard Deviation
S_R	Reproducibility Standard Deviation (Inter-Laboratory)
S_{RW}	Reproducibility Standard Deviation (Intra-Laboratory) or Intermediate Precision
σ_p	Target Standard Deviation
σ_h	Homogeneity Target standard deviation
$\hat{\sigma}$	Assigned Value standard deviation
$u(x)$	Standard Uncertainty

$u(\hat{X})$	Standard Uncertainty of the Assigned Value
$u(bias)$	Standard Uncertainty due to Bias
$u(\bar{x})$	Standard Uncertainty of Participant's Results
u_c	Combined (standard) Uncertainty
U	Expanded Uncertainty
x or x_i	Submitted Result or Value
\bar{x}	Measurement Result / Mean submitted result
\hat{X}	Assigned Value

Terms and Definitions

Specific references for terms that can be found in International Standards or guidance documents have been given in brackets at the end of each definition. Here, **VIM** refers to '*International vocabulary of metrology*' (JCGM 200; 2008), **GUM** refers to the '*Guide to the expression of uncertainty in Measurement*' (JCGM 100; 2008) and **ISO (1)**, refers to (ISO 5725-1, 1994) on the '*Accuracy (trueness and precision) of measurement methods and results*'. Terms shown in bold indicate further definitions that may be found in this section.

Readers are recommended to consult these documents for additional notes and comments not included here.

Accuracy

closeness of agreement between a measured result and the true value (if it could be known), or a reference value. (VIM 2.13)

NOTE 1; Accuracy is a concept that cannot be directly quantified. It does not possess a numerical value.

NOTE 2; Accuracy describes **random** and **systematic error** effects and as such is composed of both **precision** and **bias** components.

Analysis of Variance (ANOVA)

A group of statistical techniques that enable the different contributions from various sources of the observed variance in experimental data to be separated and estimated. (Currell and Dowman, 2005, Miller and Miller, 2005).

NOTE 1; A one-way ANOVA uses the F-test to compare the effect of one factor plus the experimental precision, eg; the effect of the measurement process on different samples, (between-sample variance) against the inherent experimental precision (within-sample variance).

NOTE 2; Whilst it is possible to carry out the analysis by hand more commonly statistical software packages are more convenient such as the Excel Data Analysis tools as this also carries out the F-test evaluation at the same time.

Assigned Value \hat{X}

The best estimate of the true value of the measurand.

NOTE; This may be the certified reference value of a CRM, a reference value from a reference laboratory or the consensus value from participants' results calculated as the robust mean, median or mode.

Assigned Value standard deviation ($\hat{\sigma}$)

Standard deviation of the assigned value.

NOTE; This may be the robust standard deviation, sMAD (median absolute deviation) or SEM (standard error of the mode)

Between-sample standard deviation (S_L);

The precision or dispersion between independent measurements carried out on different samples of the same material under **reproducibility conditions**.

NOTE: it includes the between-operator, between-day, between-instruments, and between-laboratory variability's, etc. and is a component of **reproducibility standard deviation**. It is determined using **ANOVA**, such that;

$$S_L = \sqrt{\frac{\text{between group mean square} - \text{within group mean square}}{n}}$$

Bias

estimate of a systematic measurement error (VIM 2.18)

$$bias = (\bar{x} - \hat{X})$$

Bias Root Mean Square (RMS_{bias})

A component of the bias standard uncertainty taking into account both the bias and bias variation.

See **Standard uncertainty due to bias ($u(bias)$)**.

Certified Reference Material (CRM);

a reference material accompanied by certified traceable measurement and uncertainty values determined using validated procedures (VIM 5.14)

Cochran's Test

A statistical test that detects extreme variances between observations by calculating the Cochran's (C) value as the ratio between the largest squared difference (D_{max}^2) to the sum of all the squared differences ($\sum D_i^2$) and comparing this against tabulated critical values. (ISO 5752-2: 1994)

$$C = \frac{D_{max}^2}{\sum D_i^2}$$

Coefficient of Variation ($CV\%$) (expressed as a percentage).

See **Relative standard deviation ($RSD\%$)**

Combined (standard) Uncertainty (u_c)

The combined standard uncertainty of a measurement result taking into account various contributions from different standard uncertainty sources. (GUM 2.3.4)

NOTE 1; There are two common rules for the combination of **standard uncertainty** values which depend on the model used for deriving the measurement value;

Eg; a). If the model involves the addition or subtraction of values,
i.e.; $y = (a + b + c \dots)$ then the combined standard uncertainty, $u_c(y)$ is given by;

$$u_c(y(a, b, c \dots)) = \sqrt{u(a)^2 + u(b)^2 + u(c)^2 + \dots}$$

Eg; b). If the model involves the product or quotient of values,
i.e.; $y = (a \times b \times c \dots)$ or $y = a/(b \times c \dots)$ then the combined standard uncertainty, $u_c(y)$ is given by;

$$u_c(y(a, b, c \dots)) = y \sqrt{\left(\frac{u(a)}{a}\right)^2 + \left(\frac{u(b)}{b}\right)^2 + \left(\frac{u(c)}{c}\right)^2 + \dots}$$

NOTE 2; For proficiency testing the format given in the first example has been used, thus;

$$u_c = \sqrt{S_{Rw}^2 + u(\bar{x})^2 + u(\hat{X})^2 + (bias)^2}$$

Where; $\sqrt{S_{Rw}^2}$ = uncertainty due to precision, and
 $\sqrt{u(\bar{x})^2 + u(\hat{X})^2 + (bias)^2}$ = $u(bias)$ i.e.; the **uncertainty due to bias**.

Coverage Factor (*k*)

Factor used to multiply the combined uncertainty by in order to derive the Expanded uncertainty value.

NOTE; For large data sets where the distribution approximates to normality the value of *k* to use is taken from the level of confidence required in the measurement result. Most often a 95% or 2 standard deviation level of confidence is required for the reporting of measurement results, thus *k*=2.

For smaller data sets where the distribution of measurement results is better described by a t-distribution, the equivalent t-value is used as the multiplier, thus $k=t_{(0.5,df)}$.

Error

measured quantity value minus a reference value or true value (VIM 2.16)

NOTE 1; To some extent the concept of error is a theoretical one as it is not possible to be sure of a measurand's true value, only a best estimation of it from measurement determinations. If a reference value is to be used then it is more accurate to determine the precision and bias as estimates of random and systematic error contributions which can be quantified.

Expanded Uncertainty (*U*)

A quantity defined by a specified interval (i.e.; 2 standard deviations) or confidence level (i.e.; 95% confidence) about the measurement result and describes the dispersion where a large number of repeated **measurement results** would be expected to lie.

$$U = u_c \times k \quad \text{where } k = \text{the coverage factor, and} \\ u_c = \text{the combined uncertainty}$$

Experimental standard deviation of the mean.

See **Standard Uncertainty (*u(x)*)**

External Quality Control (EQC)

See **Quality Control (QC)**.

F₁ and F₂

Are constants used to test the hypothesis that there is no significant evidence that the sampling standard deviation exceeds the allowable fraction of the target standard deviation and that the test for sufficient homogeneity has been passed (Fearn, T. and Thompson, M., 2001).

$$s_{sam}^2 = F_1 s_{all}^2 + F_2 s_{an}^2$$

Values for F₁ and F₂ may be derived from statistical tables;

$$F_1 = \frac{\chi^2_{(m-1,0.95)}}{m-1} \quad \text{where } m = \text{the number of samples measured in duplicate}$$

$$F_2 = \frac{F_{(m-1,m,0.95)} - 1}{2}$$

NOTE; The (Fisher) F-Test is a test for significant differences between the variances of two data sets and compares random error effects. The F-test may also be used within other tests such as ANOVA, (Currell, G., & Dowman, A., 2005, Miller, J.N. & Miller, J.C., 2005)

Thus; F-statistic
$$F = \frac{s_a^2}{s_b^2} \text{ or } = \frac{MS_{between}}{MS_{within}}$$

(Homogeneity) Analytical Precision (s_{an})

The homogeneity within-sample standard deviation for the replicate values (i.e.; a and b) used in the test for sufficient homogeneity of the test materials. Calculated from the ANOVA within group mean square;

$$s_{an} = \sqrt{MS_w}$$

(Homogeneity) Analytical Variance (s_{an}^2)

The square of the analytical precision. Calculated from the ANOVA within group mean square;

$$s_{an}^2 = MS_w$$

(Homogeneity) Sampling Precision (s_{sam})

The homogeneity between-sample standard deviation for the samples (i.e.; 1, 2...10) used in the test for sufficient homogeneity of the test materials. Calculated from the ANOVA between and within group mean square values;

$$s_{sam} = \sqrt{\frac{MS_b - MS_w}{2}}$$

(Homogeneity) Sampling Variance (s_{sam}^2)

The square of the sampling precision. Calculated from the ANOVA between and within group mean square values;

$$s_{sam}^2 = \frac{MS_b - MS_w}{2}$$

Homogeneity Target standard deviation (σ_h).

In the absence of an external value for target standard deviation (σ_p), a target value sufficient homogeneity (σ_h) can be determined using fitness-for-purpose criteria.

(Homogeneity) Total Permissible Sampling Variance (s_{all}^2)

The total allowable between-sample variance that must not be exceeded by the sampling variance in order for the test materials to be considered homogeneous. s_{all}^2 is derived from the homogeneity target standard deviation (either σ_p or σ_h).

$$s_{all}^2 = (0.3 \times \sigma_p)^2$$

Intermediate conditions

Independent measurement results obtained for identical test items using the same measurement procedure under a specified set of conditions within the same laboratory that include, different operators, different operating conditions, different locations over any given period of time, (VIM 2.22). See **Reproducibility Standard Deviation (Intra-Laboratory) or Intermediate Precision (S_{RW})**

Internal Quality Control (IQC)

See **Quality Control (QC)**

Measurement Result / Mean submitted result (\bar{x})

The average of an individual participant's replicate measurement results for the same analyte in the proficiency test.

Precision

closeness of agreement between repeated measurement results on the same material under specified conditions (VIM 2.15)

NOTE 1; Precision can be quantified and usually expressed as a measure of imprecision such as standard deviation, variance, relative std dev or CV and is a measure of random error.

NOTE 2; Specific measurement conditions can be repeatability, intermediate or reproducibility conditions.

Proficiency test (PT);

An **external quality control (EQC)** procedure through which the **accuracy** of a laboratory's measurement result can be objectively evaluated. Performance is assessed by providing a comparison of **trueness** with other participating laboratories

NOTE: **Trueness** is determined through the evaluation of laboratory **bias** against a reference value. This may be presented as **z-scores** or other assessment of **bias**.

Quality Assurance (QA);

Documented procedures that describe a quality management system designed to control activities and maintain a quality output.

Quality Control (QC);

Specific activities that are carried out in order to implement the procedures documented under the **Quality Assurance** programme.

NOTE; This may be in the form of **Internal Quality control (IQC)** that are carried out internally by the organization such as method validation, calibration, control charts, etc, or **External Quality Control (EQC)** coordinated by an external organization such as interlaboratory comparisons eg; proficiency tests or collaborative trials.

Random error

component of measurement error that in replicate measurements varies unpredictably (VIM 2.19)

NOTE 1; A random error value is determined as the precision that would result from a number of replicate measurements of the same measurand, expressed as a distribution.

Relative Bias % (expressed as a percentage)

Bias divided by the assigned value (x 100)

$$\text{relative bias \%} = \frac{(\bar{x} - \hat{X})}{\hat{X}} \times 100$$

Relative Between Sample Standard Deviation ($RSD_L\%$), (expressed as a percentage)

The **between-sample standard deviation** divided by the (average) measurement result (x 100)

$$RSD_L\% = \left(\frac{s_L}{\bar{x}} \right) \times 100$$

Relative Standard Uncertainty ($RSU\%$), (expressed as a percentage)

The **standard uncertainty** divided by the (average) measurement result (x 100)

$$RSU\% = \left(\frac{u(\bar{x})}{\bar{x}} \right) \times 100$$

Relative standard deviation ($RSD\%$) or Coefficient of Variation ($CV\%$) (expressed as a percentage)

The **standard deviation** divided by the (average) measurement result (x 100)

$$RSD\% \text{ or } CV\% = \left(\frac{s}{\bar{x}} \right) \times 100$$

Relative Repeatability standard deviation ($RSD_r\%$), (expressed as a percentage)

The **repeatability standard deviation** divided by the (average) measurement result (x 100)

$$RSD_r\% = (s_r/\bar{x}) \times 100$$

Relative Reproducibility standard deviation ($RSD_R\%$), expressed as a percentage

The **Reproducibility standard deviation** divided by the (average) measurement result (x 100)

$$RSD_R\% = (s_R/\bar{x}) \times 100$$

Repeatability conditions ;

Independent measurement results are obtained for identical test items under a specified set of conditions that include the same measurement procedure, same measurement system or laboratory, same operators, same operating conditions, same location and in as short a time as period as possible, (VIM 2.20, ISO (1) 3.14). See **Repeatability Standard Deviation (s_r)**

Repeatability Standard Deviation (s_r)

The dispersion or precision of replicate measurement values carried out under repeatability conditions (ISO (1) 3.15)

NOTE; Often calculated using **ANOVA** from the within group mean square (MS), such that;

$$s_r = \sqrt{\text{within group mean square}}$$

Eg; a). Within-sample (or instrumental/analytical) repeatability standard deviation is the dispersion of replicate instrumental measurements carried out on the same sample in the same analytical run, eg; an individual laboratory's replicate PT results.

b). Intra-laboratory (or method + analytical) repeatability standard deviation is the dispersion of independent measurements carried out by a single laboratory on different samples of the same material, under repeatability conditions, eg. From Intra-laboratory method validation data or homogeneity analytical precision data (s_{an}).

c). Inter-laboratory repeatability (laboratory+method+analytical) standard deviation is the dispersion of independent measurements carried out by more than one laboratory on different samples of the same material, under repeatability conditions, eg, collaborative trial precision data.

Reproducibility Conditions;

Independent measurement results obtained for identical test items using the same measurement procedure under a specified set of conditions that include, different measurement systems and laboratories, different operators, different operating conditions, different locations over any given period of time, (VIM 2.24, ISO (1) 3.18). See **Reproducibility Standard Deviation (Inter-Laboratory) (s_R)**

Reproducibility Standard Deviation (Inter-Laboratory) (s_R)

The overall dispersion or precision of independent measurement values carried out on different samples of the same material by different laboratories, under **reproducibility conditions** and incorporates both within (repeatability) and between-sample precision estimates (ISO (1) 3.19)

Thus; $s_R = \sqrt{s_r^2 + s_L^2}$

Eg; a). The Inter-laboratory reproducibility standard deviation (s_R) obtained from a collaborative trial represents the maximum dispersion for the measurement procedure carried out across laboratories and provides an estimate of best practice for the measurement procedure for a specified matrix / analyte/ concentration. Providing a laboratory's own repeatability is in agreement with the inter-laboratory repeatability precision estimate, then the laboratory can claim the Reproducibility

standard deviation from a collaborative trial as their own **standard uncertainty** estimate.

Reproducibility Standard Deviation (Intra-Laboratory) or Intermediate Precision (S_{RW})

The overall dispersion or precision of independent measurement values carried out on different samples of the same material by the same laboratory, under **reproducibility conditions** and incorporates both within (repeatability) and between-sample precision estimates (VIM 2.23)

Thus; $S_{RW} = \sqrt{s_r^2 + s_L^2}$

Eg; Intra-laboratory reproducibility standard deviation (S_{RW}) represents the maximum dispersion for the measurement procedure carried out by an individual laboratory and is often used in method validation as the method precision for a particular matrix / analyte /concentration and used as the **standard uncertainty**.

Standard Deviation (s , sd or σ)

A term used to describe the dispersion or spread of measurement values and has the same units as the measurement value.

NOTE; by convention the symbol used for standard deviation depends on whether it is describing sample statistics or population parameters. Thus;

$$\text{Sample statistics; } s = \sigma_{n-1} = \sqrt{\frac{\sum_{i=1}^n (x_i - \bar{x})^2}{n-1}}$$

$$\text{Population parameters; } \sigma = \sqrt{\frac{\sum_{i=1}^n (x_i - \mu)^2}{n}}$$

Where x_i = individual measurement values

\bar{x} = average measurement value for the sample

μ = population mean

n = number of measurement values or population size

Standard Error of the Mean.

See **Standard Uncertainty ($u(x)$)**

Standard Uncertainty ($u(x)$)

The uncertainty of a measurement result expressed as a standard deviation, (GUM 2.3.1)

NOTE; When determined from a series of repeated measurements this can also be found referred to in texts as the experimental standard deviation or standard error of the mean.

Thus; $u(x) = s / \sqrt{n}$

Standard Uncertainty of the Assigned Value ($u(\hat{X})$)

The uncertainty of the **Assigned Value**, expressed as a standard deviation, (GUM 2.3.1).

$u(\hat{X}) = \hat{\sigma} / \sqrt{m}$ where $\hat{\sigma}$ = the **assigned value** std dev
and m = the number of participants' measurement results

NOTE; $u(\hat{X})$ is also a component of the **standard uncertainty due to bias $u(bias)$** .

Standard Uncertainty due to Bias ($u(bias)$).

The uncertainty of the bias component of a participant's measurement result, expressed as a standard deviation, (GUM 2.3.1).

NOTE 1; An individual laboratory's standard uncertainty due to bias for a single proficiency test, is given as;

$$u(bias) = \sqrt{(bias)^2 + u(\bar{x})^2 + u(\hat{X})^2}$$

NOTE 2; An individual laboratory's standard uncertainty due to bias over multiple proficiency tests, is given as;

$$u(\text{bias}) = \sqrt{RMS_{\text{bias}}^2 + u(\hat{X})^2}$$

where; RMS_{bias} = the **bias root mean square** and given as;

$$RMS_{\text{bias}} = \sqrt{\frac{\sum(\text{bias}_i)^2}{m}}$$

and $u(\hat{X})$ = the average standard uncertainty of the assigned value;

$$u(\hat{X}) = \frac{\sum \hat{\sigma}_i}{\sqrt{\sum n_i}}$$

m = the number of proficiency tests or number of bias values, and

n = the number of participants' measurement results in each PT.

NOTE 3; It often helps to carry out these calculations as the relative percentage values.

Standard Uncertainty of Participant's Results ($u(\bar{x})$)

The uncertainty of a participant's submitted replicate results, expressed as a standard deviation, (GUM 2.3.1).

$$u(\bar{x}) = \frac{s_{\bar{x}}}{\sqrt{n}} \text{ where } s_{\bar{x}} = \text{the std dev of replicate values}$$

and n = the number of replicate values submitted

NOTE; $u(\bar{x})$ is also a component of the **standard uncertainty due to bias $u(\text{bias})$** .

Submitted Result or Value (x or x_i)

An individual participant's submitted measurement result for the proficiency test.

Systematic Error

component of measurement error that in replicate measurements remains constant or varies predictably (VIM 2.17)

NOTE 1; A systematic error value is determined as the bias, i.e.; the difference between a measured result and the true or reference value. Measurement results should always be corrected where significant bias is detected.

Target Standard Deviation (σ_p)

The target value for standard deviation for the proficiency test used to calculate z-scores and assess homogeneity data.

NOTE; often determined independently from data external to the proficiency test, such as the reproducibility standard deviation ($RSD_R\%$) from a collaborative trial or using a predictive model such as the Horwitz function when appropriate of fitness-for purpose criteria. The target std dev is usually matrix / analyte specific.

$$\text{Eg; a) From a collaborative trial; } \sigma_p = \frac{RSD_R}{100} \times c$$

where RSD_R = Relative Standard Deviation of Reproducibility from collaborative trial data, expressed as %

and c = concentration, i.e. the assigned value, \hat{X} , expressed in relevant units.

Eg; b) Using the Horwitz equation; $\sigma_p = 0.02c^{0.8495}$

Or modified form; for concentrations less than 120ppb (1.2×10^{-7}); $\sigma_p = 0.22c$
and for concentrations greater than 13.8% (0.138); $\sigma_p = 0.01c^{0.5}$

Where the concentration (c) is expressed as a mass fraction as shown in () above.

Trueness

closeness of agreement between the average of a large number of replicate measurement results and the true value (if it could be known) or a reference value (VIM 2.14)

NOTE 1; Trueness is a concept that cannot be directly quantified. It does not possess a numerical value.

NOTE 2; Trueness is usually expressed as bias and a measure of systematic error.

t-value

2-tailed t-value is used as a correction factor in the determination of confidence intervals for small values of n . Derived from the t-distribution for sample data sets and described using $t(\bar{x}, s)$, compared to the normal distribution for populations described as $N(\mu, \sigma)$. Values for t may be obtained from statistical tables. (Currell and Dowman, 2005, Miller and Miller, 2005).

Such that, for a 95% confidence interval;

$$CI = \bar{x} \pm \left[t_{(2,0.05,df)} \times \frac{\sigma}{\sqrt{n}} \right]$$

NOTE; The (student's) t-Test is a test for significant differences between the mean of two data sets and compares systematic error effects.

Thus; t-statistic
$$t = \frac{(x - \mu)}{s/\sqrt{n}}$$

Uncertainty of Measurement / Measurement Uncertainty (MU)

A parameter associated with a measurement result (taken as the best estimate of the true value) and characterizes the dispersion of values that could be attributed to the measurement result, taking into account both random and systematic error contributions from all possible sources and represents the degree of doubt associated with the measurement result (GUM 2.2).

Welch-Satterthwaite formula

Formula used for deriving the effective degrees of freedom for the calculation of Expanded uncertainty, when various standard uncertainties are combined with differing degrees of freedom.

$$v_{eff} = u_c^4(y) / \sum \frac{u_i^4(y)}{v_i}$$

Where v_{eff} = the effective degrees of freedom,
 v_i = degrees of freedom of individual uncertainty components,
 u_c = combined standard uncertainty
 u_i = individual uncertainty components.

z-Score

A standardized measure of laboratory bias derived from the assigned value and target standard deviation, enabling a comparison of performance between laboratories. Satisfactory performance is considered if a $|z| \leq 2$.

$$z = \frac{(x - \hat{X})}{\sigma_p}$$

Appendix 3: Tables of Critical Values

Student *t*-distribution

df	95%	99%	df	95%	99%
1	12.7100	63.6600	26	2.0555	2.7787
2	4.3027	9.9250	27	2.0518	2.7707
3	3.1824	5.8408	28	2.0484	2.7633
4	2.7765	4.6041	29	2.0452	2.7564
5	2.5706	4.0321	30	2.0423	2.7500
6	2.4469	3.7074	31	2.0395	2.7440
7	2.3646	3.4995	32	2.0369	2.7385
8	2.3060	3.3554	33	2.0345	2.7333
9	2.2622	3.2498	34	2.0322	2.7284
10	2.2281	3.1693	35	2.0301	2.7238
11	2.2010	3.1058	36	2.0281	2.7195
12	2.1788	3.0545	37	2.0262	2.7154
13	2.1604	3.0123	38	2.0244	2.7116
14	2.1448	2.9768	39	2.0227	2.7079
15	2.1315	2.9467	40	2.0211	2.7045
16	2.1199	2.9208	41	2.0195	2.7012
17	2.1098	2.8982	42	2.0181	2.6981
18	2.1009	2.8784	43	2.0167	2.6951
19	2.0930	2.8609	44	2.0154	2.6923
20	2.0860	2.8453	45	2.0141	2.6896
21	2.0796	2.8314	46	2.0129	2.6870
22	2.0739	2.8188	47	2.0117	2.6846
23	2.0687	2.8073	48	2.0106	2.6822
24	2.0639	2.7970	49	2.0096	2.6800
25	2.0595	2.7874	50	2.0086	2.6778

Factors F_1 and F_2 (95% significance level)

m	20	19	18	17	16	15	14	13	12	11	10	9	8	7
F_1	1.59	1.60	1.62	1.64	1.67	1.69	1.72	1.75	1.79	1.83	1.88	1.94	2.01	2.10
F_2	0.57	0.59	0.62	0.64	0.68	0.71	0.75	0.80	0.86	0.93	1.01	1.11	1.25	1.43

(Fearn and Thompson, 2001)

Cochran's Critical values (95% significance level)

No of Samples (m)	No of sample replicates (n)	
	2	3
2	99.9	97.5
3	96.7	87.1
4	90.7	76.8
5	84.1	68.4
6	78.1	61.6
7	72.7	56.1
8	68.0	51.6
9	63.9	47.8
10	60.2	44.5
11	57	41.7
12	54.1	39.2
13	51.5	37.1
14	49.2	35.2
15	47.1	33.5
16	45.2	31.9
17	43.4	30.5
18	41.8	29.3
19	40.3	28.1
20	38.9	27.1

(ISO 5725-2, 1994)

Appendix 4: References

AOAC (2000) AOAC Official Methods Program Manual Part 12. Appendix D: Guidelines for Collaborative Study Procedures to Validate Characteristics of a Method of Analysis. Available from; <http://www.aoac.org/vmeth/omamanual/omamanual.htm>. AOAC International.

BARWICK, V. J. & ELLISON, S. L. R. (2000) The evaluation of measurement uncertainty from method validation studies. *Accreditation and Quality Assurance: Journal for Quality, Comparability and Reliability in Chemical Measurement*, 5, 47-53.

CURRELL, G. & DOWMAN, A. (2005) *Essential Mathematics and Statistics for Science*, Chichester, John Wiley & Sons Ltd.

ELLISON, S. (2002a) Kernal.xls, version1.0e. Kernel density estimation based on RSC AMC Technical Brief No 4. Available to download from; <http://www.rsc.org/Membership/Networking/InterestGroups/Analytical/AMC/Software/index.asp>. RSC.

ELLISON, S. (2002b) Robstat.xls version 1.0. Robust Statistics Tool Kit based on RSC AMC Technical Brief No 6. Available to download from: <http://www.rsc.org/Membership/Networking/InterestGroups/Analytical/AMC/Software/index.asp>.

EURACHEM / CITAC (2000) Guide CG 4: Quantifying Uncertainty in Analytical Measurements. 2 ed., Available from; <http://www.citac.cc/QUAM2000-1.pdf>.

EUROLAB (2006) Technical Report No. 1/2006. Guide to the evaluation of measurement uncertainty for Quantitative test results. Available from; http://www.eurolab.org/docs/technical%20report/EL_11_01_06_387%20Technical%20report%20-%20Guide_Measurement_uncertainty.pdf.

EUROLAB (2007) Technical Report No. 1/2007. Measurement uncertainty revisited: Alternative approaches to uncertainty evaluation. Available from; http://www.eurolab.org/pub/i_pub.html.

FEARN, T. & THOMPSON, M. (2001) A new test for 'sufficient homogeneity'. *The Analyst*, 126, 1414-1417.

HORWITZ, W. (1982) Evaluation of analytical methods used for regulation of foods and drugs. *Analytical Chemistry*, 54, 67A-76A.

HORWITZ, W. (1995) IUPAC Protocol for the design, conduct and interpretation of method-performance studies.

HORWITZ, W., KAMPS, L. R. & BOUYER, K. W. (1980) Quality assurance in the analysis of foods and trace constituents. *J.AOAC*, 63, 1344-1354.

ISO 5725-1 (1994) Accuracy (trueness and precision) of measurement methods and results. Part 1: General principles and definitions., International Standards Organisation.

ISO 5725-2 (1994) Accuracy (trueness and precision) of measurement methods and results. Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method., International Standards Organisation.

ISO 5725 (1994) Accuracy (trueness and precision) of measurement methods and results - Part 2; Basic method for the determination of repeatability and reproducibility of a standard measurement method., International Standards Organisation.

ISO 13528 (2005) Statistical Methods for use in Proficiency Testing by Inter-Laboratory Comparisons. International Standards Organisation.

ISO 21748 (2010) Guidance for the use of repeatability, reproducibility and trueness estimates in measurement uncertainty estimation. International Standards Organisation.

ISO / IEC 17025 (2005) General requirements for the competence of testing and calibration laboratories. International Standards Organisation.

JCGM 100: (2008) Evaluation of measurement data - Guide to the expression of uncertainty in measurement (GUM). 1 ed., Available from;
http://www.bipm.org/utils/common/documents/jcgm/JCGM_100_2008_E.pdf.

JCGM 200: (2008) International Vocabulary of Metrology - Basic and general concepts and associated terms (VIM). Available from;
<http://www.bipm.org/en/publications/guides/vim.html>

KAUFMAN, D. S. & MANLEY W.F. (1998) A New Procedure for determining DL amino acid ratios in fossils using reverse phase liquid chromatography. *Quaternary Geochronology*, 17, 987-1000.

KOSNIK, M. A., KAUFMAN, D. S. & HUA, Q. (2008) Identifying outliers and assessing the accuracy of amino acid racemization measurements for geochronology: I. Age calibration curves. *Quaternary Geochronology*, 3, 308-327.

LOWTHIAN, P. J. & THOMPSON, M. (2002) Bump-hunting for the proficiency tester - searching for multimodality. *The Analyst*, 127, 1359-1364.

MAGNUSSON, B., NAYKKI, T., HOVIND, H. & KRYSELL, M. (2004) NORDTEST Report TR 537. Handbook for calculation of measurement uncertainty in Environmental Laboratories. Available from;
<http://www.nordicinnovation.net/nordtestfiler/tec537.pdf>. 2 ed.

MCCOY, W. D. (1987) The Precision of Amino Acid Geochronology and Paleothermometry. *Quaternary Science Reviews*, 6, 43-54.

MILLER, J. N. & MILLER, J. C. (2005) *Statistics and Chemometrics for Analytical Chemistry*, Harlow, England., Pearson Education Ltd.

RSC ANALYTICAL METHODS COMMITTEE (1989) AMC Technical Briefs; Robust Statistics-how not to reject outliers: Part 1, Basic concepts. *The Analyst*, 114, 1693-1697.

RSC ANALYTICAL METHODS COMMITTEE (1995) Uncertainty of measurement: implication of its use in analytical science. *The Analyst*, 120, 2303-2308.

RSC ANALYTICAL METHODS COMMITTEE (2001) AMC Technical Briefs No 6; Robust Statistics: a method of coping with outliers. Available from;
<http://www.rsc.org/Membership/Networking/InterestGroups/Analytical/AMC/TechnicalBriefs.asp>.

RSC ANALYTICAL METHODS COMMITTEE (2004) AMC Technical Briefs No 17; The Amazing Horwitz function. Available from;
<http://www.rsc.org/Membership/Networking/InterestGroups/Analytical/AMC/TechnicalBriefs.asp>.

RSC ANALYTICAL METHODS COMMITTEE (2006) AMC Technical Briefs No 4; Representing data distributions with kernel density estimates. Available from;
<http://www.rsc.org/Membership/Networking/InterestGroups/Analytical/AMC/TechnicalBriefs.asp>.

THOMPSON, M., ELLISON, S. L. R. & WOOD, R. (2006) The International Harmonized Protocol for the Proficiency Testing of Analytical Chemistry Laboratories. *Pure and Applied Chemistry*, 78, 145-196. Available from;<http://www.iupac.org/publications/pac/2006/pdf/7801x0145.pdf>.

WEHMILLER, J. F. (1984) Interlaboratory Comparison of Amino Acid Enantiomeric Ratios in Fossil Pleistocene Mollusks. *Quaternary Research*, 22, 109-120.

WEHMILLER, J. F. (1992) Aminostratigraphy of Southern California Quaternary Marine Terraces. IN FLETCHER, C. H. I. & WEHMILLER, J. F. (Eds.) *Quaternary Coasts of the United States: Marine and Lacustrine Systems*. Tusla, SEPM (Society for Sedimentary Geology). Special Edition No 48, p 317-321.

WEHMILLER, J. F. (2010) AAR Interlaboratory comparison of fossil hydrolysates (unpublished).

WEHMILLER, J. F. & MILLER, G. H. (2000) Aminostratigraphic Dating Methods in Quaternary Geology. IN NOLLER, J. S., SOWERS, J. M., COLMAN, S. M. & PIERCE, K. L. (Eds.) *Quaternary Geochronology: methods and Applications*. Washington DC, American Geophysical Union, Reference Shelf Series 4.

Contact Details;

Please direct communications regarding this Report to;

Jo Powell

BioArCh, Dept of Archaeology,
University of York,
Biology S Block,
Wentworth Way,
YORK YO10 5DD. UK

tel; +44(0)1904 328806

email: jp588@york.ac.uk